

NASA TECH BRIEF

Marshall Space Flight Center



NASA Tech Briefs announce new technology derived from the U.S. space program. They are issued to encourage commercial application. Tech Briefs are available on a subscription basis from the National Technical Information Service, Springfield, Virginia 22151. Requests for individual copies or questions relating to the Tech Brief program may be directed to the Technology Utilization Office, NASA, Code KT, Washington, D.C. 20546.

Helium Leak Measurements Using CO₂ as a Carrier

The problem:

Present methods to accurately detect the presence and level of gas leakage often incur long delays in detection.

The solution:

Leak detection is accomplished by using a helium mass spectrometer leak-detector and, by combining several known procedures, omits those defects in present techniques that result in delay of detection.

How it's done:

The improved leak-detection technique combines carbon dioxide purging, cryogenic separating helium accumulation, and the use of carbon dioxide as a carrier gas.

Helium leak-detector sensitivity using the specimen evacuation technique is better than 1×10^{-9} std. cm³/sec. The difficulty in locating leaks using this technique occurs when the specimen is a large volume-low conductance system such as the plumbing of a typical cryogenic shroud used in vacuum chambers. The controlling factor for locating leaks in such a system is the system time-constant rather than the leak-detector sensitivity.

The specimen pressurization technique, i.e., sniffing technique, is used as an attempt to circumvent the time constant effect. This method reduces the effective leak-detector sensitivity to 1×10^{-6} std. cm³/sec. at best. When using the sniffing technique, the controlling factor becomes the effective sensitivity rather than the system time-constant.

This study deals with the development of a technique that retains the basic leak-detector sensitivity while circumventing the specimen time-constant. The technique uses CO₂ as a carrier gas. The leak-detector liquid nitro-

gen trap is used as a cryogenic pump for the carrier gas. The permissible throughput of the leak-detector is a nominal 5×10^{-4} std. cm³/sec. for air. The leak-detector throughput using CO₂ may be maintained at 10^{-1} std. cm³/sec. without compromising the leak-detector ionization chamber vacuum level.

A continuous CO₂ purge is maintained through the specimen. Helium is introduced into the specimen at a rate of 2.7×10^{-8} std. cm³/sec. The leak-detector response is recorded and compared to the response when no CO₂ purge is used. The helium leak is valved off, and the "cleanup" rates with and without the CO₂ purge are recorded. Using the CO₂ purge, the response rate shows a significant increase without loss in the effective sensitivity.

Notes:

1. Information concerning this innovation may be of interest to manufacturers and users of pressure vessels, cryogenic hardware, or vacuum hardware.
2. Requests for further information may be directed to:
Technology Utilization Officer
Marshall Space Flight Center
Code A&TS-TU
Huntsville, Alabama 35812
Reference: B72-10354

Patent status:

No patent action is contemplated by NASA.

Source: B. C. Moore and R. G. Camarillo of
McDonnell Douglas Astronautics Company
under contract to
Marshall Space Flight Center
(MFS-21742)

Category 03