

Case file copy

N71-26607

NASA CR-111890

NASA CR-111890

THE PERKIN-ELMER CORPORATION
AEROSPACE DIVISION
2855 Metropolitan Place, Pomona, California 91767

FINAL REPORT
UNDERSEA ATMOSPHERIC ANALYZER
VOLUME 6 OF 6
FOR
COMBINED STUDY PROGRAM

By Burd F. Bicksler

March 1971

Perkin-Elmer SPO 30006
NASA Contract Number NAS1-9469

Prepared for
NATIONAL AERONAUTICS AND SPACE ADMINISTRATION
LANGLEY RESEARCH CENTER
Langley Station
Hampton, Virginia 23365

THE PERKIN-ELMER CORPORATION
AEROSPACE DIVISION
2855 Metropolitan Place, Pomona, California 91767

FINAL REPORT
UNDERSEA ATMOSPHERIC ANALYZER
VOLUME 6 OF 6
FOR
COMBINED STUDY PROGRAM

By Burd F. Bicksler

March 1971

Prepared By B.F. Bicksler Date 5/13/71
B.F. Bicksler, Project Manager

Approved By M.R. Ruecker Date 13 May '71
M.R. Ruecker, Manager Space Physics

Approved By W.C. Qua Date 14 May 71
W.C. Qua, Program Manager

Perkin-Elmer SPO 30006
NASA Contract Number NAS1-9469

Prepared for
NATIONAL AERONAUTICS AND SPACE ADMINISTRATION
LANGLEY RESEARCH CENTER
Langley Station
Hampton, Virginia 23365

ABSTRACT

A program was conducted in which an instrument system concept was studied to optimize the application of a mass spectrometer as a sensor for monitoring the primary atmospheric constituents, as well as atmospheric contaminants, on board a manned spacecraft. The program was divided into six individual studies representing the primary system parts complementing the spectrometer: a Carbon Monoxide Accumulator Cell (Volume 1), an Ion Pump (Volume 2), an Ion Pump Power Supply (Volume 3), an Inlet Leak (Volume 4), an Ion Source (Volume 5), and an Undersea Atmospheric Analyzer (Volume 6). The principal goal of the combined study program was the achievement of an instrument concept of minimum power, weight and size without compromising the minimum detection limits of the instrument.

TABLE OF CONTENTS

| | Page |
|---|------|
| ABSTRACT | ii |
| SUMMARY | (1) |
| INTRODUCTION | (1) |
| PERFORMANCE SUMMARY | 43 |
| SAMPLE TRANSPORT PUMP FAILURE | 46 |
| CONCLUSIONS | 46 |
| APPENDIX 1. RESULTS OF FIVE-DAY SEA TRIAL ON THE UNDERSEA ATMOSPHERIC SENSOR | 49 |
| APPENDIX 2. EVALUATION OF A FOUR GAS MASS SPECTROMETER | 61 |
| APPENDIX 3. OPERATION AND MAINTENANCE MANUAL | 71 |

LIST OF ILLUSTRATIONS

| | Page |
|--|---------|
| 1. Undersea Atmospheric Sensor System Calibration Before and After Sea Deployment | 89 |
| 2. Comparison of Oxygen Data Outputs (a-e) | 90-94 |
| 3. Undersea Atmospheric Sensor System Output During (a-e) Sea Deployment | 95-99 |
| 4. Comparison of Carbon Dioxide Outputs During (a-e) Sea Deployment | 100-104 |
| 5. Undersea Atmospheric Sensor System Output During (a-e) Sea Deployment | 105-109 |
| 6. Two Carbon Vanes from Gast Pump Showing Abrasion During Operation | 110 |
| 7. Carbon Vanes from Gast Pump Showing Failure and Abrasion | 111 |
| 8. Five-Day Sea Trial Data | 112 |
| 9. Basic Two Gas ATM Sensor | 113 |
| 10. Principles of Mass Spectrometer Operation | 114 |
| 11. Two Gas Atmosphere Analyzer Assembly | 115 |
| 12. Atmospheric Sensor System Block Diagram | 116 |
| 13. Sample and Calibration Inlet System Schematic Flow Diagram | 117 |

LIST OF ILLUSTRATIONS (Cont)

| | Page |
|---|---------|
| 14. Atmospheric Sensor Front Panel | 118 |
| 15. 90-Day Space Station Simulator Atmospheric Sensor System | 119 |
| 16. Atmospheric Sensor - Upper Bay | 120 |
| 17. Atmospheric Sensor - Lower Bay | 121 |
| 18. Comparison of the Summation of Partial Pressures to the Total Pressure during the Five Day Test | 122 |
| 19. Oxygen and Nitrogen Partial Pressures During the 90-Day Test | 123 |
| 20. Carbon Dioxide and Water Vapor Partial Pressures During the 90-Day Test | 124 |
| 21. Comparison of the Summation of Partial Pressures to the Total Pressure During the 90-Day Test | 125 |
| 22. Flight Qualified Mass Spectrometer Atmospheric Sensor System for Atmospheric and Respiratory Monitoring | 126 |
| 23. Undersea Atmospheric Sensor (UAS) | 127 |
| 24. Block Diagram - Undersea Atmosphere Sensor (UAS) | 128 |
| 25. UAS (With Panel Covers Installed) | 129 |
| 26. Identification of System Controls (a-c) | 130-132 |
| 27. Pump Current (20 l/s Ultek Ion Pump) | 133 |

SUMMARY

This report describes an Undersea Atmospheric Sensor System which has been designed to monitor and display the atmospheric constituents of nitrogen, oxygen, carbon dioxide, water vapor, and argon during at-sea operational deployment of a test vehicle. Data recorded during a five-day sea-trial and during a subsequent 62-day (continuous) sea deployment has indicated unusually stable long-term operation. In addition, data recorded by crew members of the U.S.S. Hammerhead is presented which compares the Perkin-Elmer Undersea Atmospheric Sensor System with other analyzer systems aboard ship - the Mark V and a portable Model D-2 Oxygen Analyzer.

INTRODUCTION

Under the general scope of contract NAS1-9469, the Perkin-Elmer Corporation, Aerospace Division, performed the effort necessary to modify a Government-furnished two gas atmospheric sensor system for use as an Undersea Atmospheric Sensor System. The intent of the experiment was to monitor the atmospheric constituents of nitrogen, oxygen, carbon dioxide, water vapor and argon during operational deployment of a test vehicle. The operation of the Undersea Atmospheric Sensor was to be essentially automatic providing visual display of the atmospheric constituent partial pressures. Trouble free, long term, reliable operation and ease of maintenance were the key system requirements.

During the period 25 September through 3 October, the Undersea Atmospheric Sensor operated continuously under in-use environment conditions, providing continuous output data and maintaining calibration. The results of the five-day sea trial were published in a Perkin-Elmer disclosure "Results of Five-Day Sea Trial on the Undersea Atmospheric Sensor - Uptite" dated October 1970 and submitted to NASA Langley Research Center, NASA contract NAS1-9469, Perkin-Elmer SPO 30006. (Reference: Appendix 1.)

Between 2 October and 6 October Perkin-Elmer personnel performed additional field calibration and final system adjustments in preparation for a long-term, at-sea deployment of the test vehicle with the Undersea Atmospheric Sensor on board. After final calibrations and adjustments, the instrument was operating within 1.0 torr of theoretical values on all channels and within 1.0 torr of the total barometric pressure. The calibration gas sample utilized prior to and at the conclusion of the five-day sea trial was the same sample that was utilized for all subsequent calibrations. The calibration gas sample consisted of 74.8% nitrogen, 20.11% oxygen, 2.13% carbon dioxide, and 3.0% argon. Water vapor calibrations were achieved utilizing relative humidity and temperature at the sample point.

From 1001, 6 October through 0900 6 December, a total of 62 continuous days, the Undersea Atmospheric Sensor System performed without functional interruption or calibration, with the exception of 8 instances, when line power was interrupted intentionally for various drills. During these line power interruptions the Undersea Atmospheric Sensor was inoperative for a total of 110 minutes. Each time line power was restored the system automatically restarted and restabilized.

Between 0900 hours and 1100 hours on 6 December the sample transport pump, a commercially procured auxillary pump utilized to move the air sample to point of analysis, ceased functioning. The non-operating condition of this pump precluded a new air sample from being transported to the analyzer. At-sea repair of the transport pump was not attempted due to impending arrival of the U.S.S. Hammerhead at port.

On 15 December a Perkin-Elmer representative performed replacement of the sample transport pump and calibration of the system. Comparing the Undersea Atmospheric Sensor System output after the sample transport pump had been replaced and the system had stabilized showed that: 1) the nitrogen output was within 4 torr of a standard laboratory air sample; 2) oxygen was within 1.1 torr; 3) argon was within 0.11 torr; and 4) carbon dioxide was within 0.06 torr. The pressure transducer output, utilized as a pressure reference, was within 2.0 torr of the local barometric pressure.

Final system calibration on the calibration gas sample showed that the nitrogen channel was within 6.0 torr of the theoretical 587 torr value and all other channels were within 1.0 torr of their theoretical values.

The Undersea Atmospheric Sensor System was initially calibrated for long term sea deployment such that it was within 0.5% on all channels except argon, a low-level channel, was within 3.0% of the theoretical value. Upon completion of the long term sea deployment of the system, calibration showed that the instrument was within 1% on all channels except for argon, which was within 2.5% of the theoretical value.

Table 1a, 1b, and 1c are copies of the engineering data recorded during the long term sea deployment by members of the U.S.S. Hammerhead's crew. This data was taken periodically for purposes of evaluating the systems internal performance by Perkin-Elmer. Table 1d, 1e, and 1f are copies of engineering data taken by Perkin-Elmer personnel during system calibration at the conclusion of long term sea deployment. The last three lines of data are final system calibration originally recorded in Perkin-Elmer Laboratory Log Book #2599, Page 122.

Table 2a through 2hh are copies of comparison data between the Undersea Atmospheric Sensor System, the Mark V and a portable Model D-2 Oxygen Analyzer. This data was also recorded by U.S.S. Hammerhead crew members.

TABLE 1a.- DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

**PERKIN-ELMER
AEROSPACE DIVISION**

MASS SPECTROMETER SENSOR DATA FORM #2

SHEET 1 of 3

Date: 7-50

| DATE | TIME | N ₂ | O ₂ | CO ₂ | H ₂ O | A | Z _{PP} | T.P. ₃ | T.P. ₄ | T.P. ₅ | T.P. ₆ | T.P. ₇ | REMARKS |
|-------|------|----------------|----------------|-----------------|------------------|-----|-----------------|-------------------|-------------------|-------------------|-------------------|-------------------|------------------------------|
| 10/6 | 1120 | 593 | 159.9 | 1.11 | 8.9 | 7.5 | | 131.6 | 15.8 | 136.9 | -17.4 | 38.5 | Bd Sr # 2 |
| 10/14 | 1305 | 606 | 150.4 | 6.59 | 11.2 | 7.6 | | 131.6 | 16.7 | 136.9 | -18.0 | 38.5 | Bd Sr # 2 |
| 10/14 | 1342 | 609 | 150.4 | 6.76 | 11.0 | 7.6 | | 131.6 | 16.5 | 136.9 | -19.8 | 38.4 | Bd Sr # 2 |
| 10/14 | 1355 | 606 | 150.5 | 6.84 | 8.1 | 7.5 | | 131.6 | 16.7 | 136.9 | -18.0 | 38.4 | |
| 10/15 | 1630 | 598 | 159.8 | .81 | 10.6 | 7.3 | 766.5 | 131.6 | 16.7 | 136.9 | -18 | 38.4 | Bd Sr # 2 |
| 10/17 | 1425 | 581 | 147.2 | 7.07 | 10.2 | 7.2 | 757.7 | 131.6 | 16.7 | 136.9 | -17.9 | 38.5 | " " |
| 10/19 | 2105 | 681 | 153.1 | 7.57 | 10.4 | 7.6 | 859.7 | 131.6 | 16.4 | 136.4 | -17.6 | 38.5 | " " |
| 10/24 | 1848 | 600 | 145.2 | 6.62 | 10.1 | 8.4 | 770.3 | 131.6 | 16.6 | 136.8 | -17.6 | 38.6 | " " |
| 11/25 | 0100 | 597 | 144.0 | 8.40 | 10.6 | 8.3 | | 131.6 | 16.7 | 136.8 | -17.9 | 38.7 | |
| 11/5 | 0015 | 587 | 153.3 | 8.61 | 8.7 | 7.7 | 766.8 | 131.6 | 16.8 | 136.8 | -18.0 | 39.1 | |
| 11/17 | 1400 | 569 | 151.5 | 1.50 | 7.6 | 7.0 | 736.6 | 131.6 | 16.8 | 136.5 | -18.0 | 39.2 | 1415 started Gas sample |
| 11/17 | 1418 | 599 | 160.5 | 1.92 | 8.0 | | | | | | | | 1438 secured Gas |
| 11/17 | 1425 | 598 | 160.5 | 1.91 | 8.6 | | | | | | | | |
| 11/17 | 1430 | 594 | 160.4 | 1.84 | 8.9 | | | | | | | | |
| 11/17 | 1435 | 594 | 160.4 | 1.81 | 9.0 | | | | | | | | |
| 11/18 | 0115 | 601 | 158.8 | 6.10 | 9.8 | 7.6 | 783.3 | 131.6 | 16.9 | 136.8 | -18.1 | 39.2 | Looks stable at this time |
| 12/8 | 0115 | 588 | 157.1 | 7.84 | 9.8 | 9.4 | 770.1 | 131.6 | 16.5 | 136.8 | -17.8 | 39.3 | |
| 12/6 | 0215 | 601 | 152.8 | 5.84 | 10.9 | 7.6 | | 131.6 | 16.9 | 136.8 | -18.1 | 39.1 | |
| 12/11 | 2030 | 594 | 158.7 | 0.87 | 8.5 | 7.4 | | 131.6 | 16.5 | 136.9 | -17.8 | 39.1 | 1415 TRANSPORT PUMP |

11-352

TABLE 1b.- DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

**PERKIN-ELMER
AEROSPACE DIVISION**

MASS SPECTROMETER SENSOR DATA FORM #2

SHEET 2 OF 3

| DATE | TIME | T.P. 8 | T.P. 9 | T.P. 10 | T.P. 11 | EHSN DR | ANODE CUR | PRESS | VALVE TEMP | ACCELR CUR | T.P. 17 | T.P. 18 | REMARKS |
|-------|------|--------|--------|---------|---------|---------|-----------|-------|------------|------------|---------|---------|---|
| 10/6 | 1180 | 36.1 | 20.0 | 18.6 | -19.7 | 13.0 | 23.6 | 77.1 | 19.2 | 53.8 | | | Bd St #2 |
| 10/14 | 1300 | 37.6 | 20.8 | 19.4 | -20.5 | 13.3 | 35.1 | 78.3 | 19.2 | 91.5 | | | Bd St #2 |
| 10/14 | 1315 | 37.4 | 20.7 | 19.3 | -20.4 | 13.3 | 33.8 | 78.4 | 19.2 | 87.8 | | | ION CURRENT LOW ADJUSTED LEAK OFFICE TO 90 |
| 10/14 | 1355 | 37.7 | 20.9 | 19.4 | -20.6 | 12.9 | 21.0 | 78.4 | 19.2 | 46.6 | | | ION CURRENT LOW Bd St #2 |
| 10/15 | 1630 | 37.4 | 20.9 | 19.5 | -20.6 | 13.0 | 19.9 | 77.0 | 19.2 | 44.2 | | | ION CURRENT LOW Bd St #2 |
| 10/17 | 1425 | 37.7 | 20.8 | 19.4 | -20.5 | 13.0 | 20.7 | 77.2 | 19.1 | 46.5 | | | not #2 |
| 10/19 | 2405 | 37.2 | 20.5 | 19.1 | -20.5 | 13.0 | 22.1 | 77.9 | 19.1 | 49.9 | | | |
| 10/24 | 1848 | 37.5 | 20.6 | 19.3 | -20.3 | 13.9 | 29.5 | 76.8 | 19.0 | 50.8 | | | |
| 10/25 | 0100 | 37.5 | 20.6 | 19.4 | -20.4 | 12.9 | 23.4 | 76.2 | 19.0 | 53.4 | | | |
| 11/5 | 10.5 | 37.6 | 20.5 | 19.4 | -20.2 | 13.1 | 28.6 | 76.3 | 18.9 | 70.4 | | | |
| 11/17 | 1400 | 37.4 | 20.4 | 19.4 | -20.1 | 13.3 | 35.1 | 77.1 | 19.0 | 92.3 | | | |
| 11/18 | 1315 | 37.6 | 20.5 | 19.5 | -20.2 | 13.1 | 25.0 | 77.7 | 18.9 | 57.4 | | | |
| 11/27 | 0115 | 37.0 | 20.2 | 19.2 | -19.9 | 13.3 | 28.9 | 76.4 | 19.0 | 72.4 | | | |
| 12/6 | 2215 | 37.4 | 20.8 | 19.7 | -20.5 | 13.4 | 17.0 | 77.4 | 19.1 | 44.5 | | | |
| 12/11 | 2030 | 37.2 | 20.9 | 19.5 | -20.2 | 13.6 | 28.4 | 76.4 | 19.0 | 50.0 | | | |

F-353

TABLE 1c.- DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

PERKIN-ELMER
AEROSPACE DIVISION

MASS SPECTROMETER SENSOR DATA FORM #2

SHEET 3 OF 3

| DATE | TIME | T.P. 19 | T.P. 20 | T.P. 21 | T.P. 22 | T.P. 23 | SAMPLE FLOW | ION PUMP CURRENT | T/C PRESS | REMARKS | DATA READ BY |
|-------|------|---------|---------|---------|---------|---------|-------------|------------------|-----------|-------------------------------|--------------|
| 10/16 | 1120 | | | | | | .015 | 80 | | 14% 7402 | |
| 10/16 | 1125 | | | | | | .015 | 80 | | 3105 | j |
| 10/14 | 1347 | | | | | | .015 | 59 | | | |
| 10/14 | 1355 | | | | | | .015 | 98 | | | |
| 10/15 | 1630 | | | | | | .015 | 102 | | Press 30.84 74°F (771.6) | j |
| 10/17 | 1425 | | | | | | .015 | 103 | | Press 30.84 76°F (783.3) | j |
| 10/19 | 2105 | | | | | | .015 | 98 | | Press 30.63 71°F (778) | j |
| 10/14 | 1848 | | | | | | .015 | 100 | | 29.82 @ 1200 (755.4) | |
| 10/25 | 0100 | | | | | | .015 | 95 | | 30.16 @ 0200 | |
| 11/5 | 1015 | | | | | | .015 | 86 | | 89.99 (760.5) | |
| 11/17 | 1400 | | | | | | .015 | 91 | | | |
| 11/18 | 1315 | | | | | | .015 | 104 | | Press 30.55 (776) | |
| 11/19 | 0115 | | | | | | .015 | 95 | | 30.01 (762) | |
| 12/6 | 0215 | | | | | | 0 | 155 | | Fuse F2 Blown Replaced & 100g | |
| 12/11 | 9030 | | | | | | 0 | 100 | | blowing 30.00 | |

F-35

TABLE 2a.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | | | | | PORTABLE | | | | ATM CONTROL | | | | |
|--------|------|-----------------|-----------------|----------------|------------------|----------------|-----|-----------------|----------------|------|-----|----------------|-----------------|-------|----------------|----------------|-------|-------------|-----|-------------|
| | | CO ₂ | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | FIL | FIA | O ₂ | LOC | GAUG | S ₁ | S ₂ | BLEED | VENTILATION | | |
| | | mm | mm | mm | mm | mm | ppm | % | % | ppm | ppm | mm | CO ₂ | in | in | in | CFH | COM | SEC | |
| 13 OCT | 0830 | | | | | 167 | 6 | -8 | .1 | 25 | 1 | 155 | -8 | | | | | | | |
| 13 OCT | 0000 | | | | | 167 | 12 | -95 | .1 | 22 | 1 | 155 | -8 | | | | | | | KEY PUNCHED |
| 13 OCT | 0400 | | | | | 167 | 7 | .99 | .05 | 18 | 3 | 155 | | | | | | | | DEC-23-1970 |
| 13 OCT | 0700 | | | | | 166 | 4 | -92 | .02 | 20 | 1 | | | 31.2 | | | | | | |
| 13 OCT | 0810 | | | | | 165 | 8 | -92 | .1 | 18.5 | 9 | | | 29.2 | | | | | | |
| 13 OCT | 1500 | | | | | 166 | 5 | -85 | -1 | 12 | 4 | | | 30.2 | | | | | | |
| 13 OCT | 1600 | | | | | 167 | 3 | -1.0 | -1 | 39 | 8 | | | 31.00 | | | | | | |
| 13 OCT | 1700 | | | | | 167 | 3 | -1.0 | -1 | 20 | 8 | | | 31.05 | | | | | | |
| 13 OCT | 1800 | 152.8 | 8.74 | 616 | 11.5 | | | | | | | 146 | | | | | | | | |
| 13 OCT | 2000 | 143.6 | 9.89 | 585 | 11.8 | 159 | 4 | -1.3 | -1 | 25 | 4 | 138 | | 30.05 | | | | | | |
| 13 OCT | 2200 | 156.2 | 5.14 | 603 | 11.2 | | | | | | | 148 | | | | | | | | |
| 14 OCT | 0000 | 151.2 | 8.35 | 614 | 11.0 | 171 | 3 | -1.1 | -1 | 20 | 4 | 155 | | 31.52 | | | | | | |
| 14 OCT | 0200 | 153.3 | 10.16 | 620 | 10.9 | | | | | | | 146 | | | | | | | | |
| 14 OCT | 0400 | 150.5 | 10.57 | 611 | 10.7 | 168 | 3 | -1.42 | -1 | 10 | 8 | 146 | | 31.64 | | | | | | |
| 14 OCT | 0600 | 145.6 | 10.77 | 578 | 11.1 | | | | | | | 145 | | | | | | | | |
| 14 OCT | 0800 | 145.3 | 9.88 | 602 | 11.1 | 168 | 8 | -7.2 | -1 | 18.5 | 9 | 140 | | 27.42 | | | | | | |

TABLE 2b. - COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | MK X | | | | PORTABLE | | | | ATM CONTROL | | | |
|-------|------|----------------------|-----------------------|----------------------|------------------------|----------------------|-----------|----------------------|---------------------|------------|----------------------|----------|------------|----------------------|----------------------|--------------|--------------------|-----|--|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | FIR ppm | O ₂ mm | LOC | BARO in | CO ₂ % | CO ₂ % | BLEED CFH | VENTILATION CCM | SEC | |
| 14OCT | 1000 | 152.4 | 5.36 | 597 | 11.1 | | | | | | 146 | | | | | | | | |
| 14OCT | 1200 | 151 | 6.43 | 604 | 11.1 | 165 | 5 | .85 | .1 | 9 | 145 | | 31.05 | | | | | | |
| 14OCT | 1400 | 150.5 | 6.85 | 606 | 8.5 | | | | | | 145 | | | | | | | | |
| 14OCT | 1600 | 142.9 | 6.78 | 582 | 11.5 | 158 | 5 | .97 | .1 | 9.5 | 143 | | 30.01 | | | | | | |
| 14OCT | 1800 | 143.6 | 7.94 | 590 | 11.2 | 157 | 3 | .97 | .1 | 12 | 138 | | 29.86 | | | | | | |
| 14OCT | 2000 | 144.4 | 7.71 | 600 | 11.7 | 157 | 5 | 1.0 | .45 | 20 | 139 | | 30.38 | | | | | | |
| 14OCT | 2200 | 150 | 4.5 | 580 | 10.7 | | | | | | 145 | | | | | | | | |
| 15OCT | 0000 | 146 | 5.43 | 577 | 11.5 | 161 | 3 | .7 | .19 | 37 | 147 | | 29.42 | | | | | | |
| 15OCT | 0200 | | | | | | | | | | | | | | | | | | |
| 15OCT | 0400 | 144.4 | 8.25 | 584 | 10.1 | 160 | 6 | 1.15 | .1 | 24 | 146 | | 29.47 | | | | | | |
| 15OCT | 0600 | 144.2 | 8.49 | 588 | 10.2 | | | | | | 144 | | | | | | | | |
| 15OCT | 0800 | 145.3 | 8.00 | 595 | 10.3 | 160 | 6 | 1.8 | .1 | 28 | 144 | | 31.08 | | | | | | |
| 15OCT | 1000 | 148.6 | 4.75 | 593 | 11.1 | | | | | | 157 | | | | | | | | |
| 15OCT | 1200 | 158.1 | 1.84 | 586 | 10.7 | 165 | 6 | 1.1 | .1 | 110 | 160 | | 30.36 | | | | | | |
| 15OCT | 1400 | 158.9 | .867 | 586 | 10.8 | | | | | | 162 | | | | | | | | |
| 15OCT | 1600 | 159.1 | .757 | 586 | 10.5 | 165 | 3 | 1.1 | .1 | 100 | 162 | | 30.35 | | | | | | |

TABLE 2c.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | MK V | | | | PORTABLE | | | ATM CONTROL | | |
|--------|-------------------------|-----------------|------|----------------|------------------|----------------|-----|-----------------|----------------|-----|----------------|----------|------|----------------|-------------|-------------|-----|
| | | CO ₂ | CO | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | FIL | O ₂ | LOC | BARO | O ₂ | BLEED | VENTILATION | |
| | | mm | mm | mm | mm | mm | ppm | % | % | ppm | mm | | in | mm | CFH | COM | SEC |
| 15 OCT | 1800 1800 | 152.1 | 2.34 | 569 | 10.8 | | | | | 153 | | 29.46 | | | | | |
| 15 OCT | 2000 | 151.6 | 4.11 | 575 | 10.3 | 168 | .5 | -1 | 94 | 153 | | 29.70 | | | | | |
| 15 OCT | 2200 | 151 | 5.70 | 580 | 9.9 | | | | | 153 | | | | | | | |
| 16 OCT | 0000 | 157 | 3.63 | 598 | 10.3 | 172 | .5 | -1 | 11 | 161 | | 31.00 | | | | | |
| 16 OCT | 0200 | 152.7 | 5.12 | 598 | 9.9 | | | | | 160 | | | | | | | |
| 16 OCT | 0400 | 156.4 | 6.49 | 608 | 9.7 | 172 | .85 | .02 | 11 | 160 | | 31.52 | | | | | |
| 16 OCT | 0600 | 153.4 | 7.99 | 600 | 9.8 | | | | | 158 | | | | | | | |
| 16 OCT | 0800 | 146.6 | 7.79 | 597 | 10.5 | 162 | 1.1 | -1 | 38 | 149 | | 29.88 | | | | | |
| 16 OCT | 1000 | 146.2 | 8.53 | 587 | 11.0 | | | | | 147 | | | | | | | |
| 16 OCT | 1200 | 156.5 | 9.68 | 587 | 11.2 | 0 | 0 | .5 | 3 | 158 | | 30.43 | | | | | |
| 16 OCT | 1400 | 157.2 | 3.31 | 576 | 10.4 | | | | | 158 | | | | | | | |
| 16 OCT | 1600 | 156.2 | 5.63 | 606 | 10.0 | 173 | -8 | .15 | 20 | 158 | | 30.97 | | | | | |
| 16 OCT | 1800 | 156.5 | 7.19 | 605 | 9.8 | | | | | 156 | | | | | | | |
| 16 OCT | 2000 | 153.8 | 7.15 | 605 | 10.6 | 172 | 1.2 | .14 | 11 | 158 | | 31.90 | | | | | |
| 16 OCT | 2200 | 144.9 | 6.86 | 576 | 10.9 | | | | | 146 | | | | | | | |
| 17 OCT | 0000 | 147 | 5.86 | 593 | 10.5 | 160 | .8 | .1 | 43 | 160 | | 30.29 | | | | | |

TABLE 2d.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | | MK IV | | | | | PORTABLE | | | | ATM CONTROL | | |
|--------|------|-----------------|----------------|------------------|----------------|-----|-----------------|----------------|-----------------|-----------------|----------------|-----|------|----------------|----------------|----------------|-------|-------------|-----|--|
| | | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | FI ₂ | FI ₂ | O ₂ | LOC | BARO | O ₂ | O ₂ | O ₂ | BLEED | VENTILATION | | |
| | | mm | mm | mm | mm | ppm | % | % | ppm | ppm | mm | | in | mm | mm | mm | CFH | COM | SEC | |
| 17 Oct | 0200 | 160 | 600 | 10.1 | | | | | | | | | | 157.8 | TR | | | | | |
| | 0400 | 160 | 607 | 9.8 | 173 | 8 | .15 | .1 | 38 | 11 | | | | 157 | TR | | | | | |
| | 0600 | 151 | 583 | 9.7 | | | | | | | | | | 148.7 | UL OPS | | | | | |
| | 0800 | 151 | 590 | 9.6 | 166 | 12 | .85 | .1 | 22 | 8 | | | | 148.5 | UL OPS | | | | | |
| | 1000 | 152 | 593 | 9.7 | | | | | | | | | | 147.7 | UL OPS | | | | | |
| | 1200 | 155 | 597 | 9.9 | 173 | 3 | .72 | .13 | 37 | 5 | | | | 154.3 | TR | | | | | |
| | 1400 | 151 | 585 | 10.0 | | | | | | | | | | 148.6 | TR | | | | | |
| | 1600 | 146 | 586 | 10.1 | 165 | 5 | 1.18 | .1 | 11 | 3 | | | | 147.3 | TR | | | | | |
| | 1800 | 145 | 592 | 10.1 | | | | | | | | | | 147.2 | TR | | | | | |
| | 2000 | 151 | 595 | 10.8 | 175 | 0 | .38 | .1 | 39 | 13 | | | | 157.5 | TR | | | | | |
| | 2200 | 151 | 603 | 9.8 | | | | | | | | | | 154.8 | TR | | | | | |
| 18 Oct | 0000 | 158 | 615 | 10.3 | 173 | 10 | 1.2 | .1 | 28 | 4 | | | | 155.2 | UL OPS | | | | | |
| | 0100 | 161 | 607 | 9.9 | 175 | 3 | .35 | .1 | 43 | 3 | | | | 160.6 | UL OPS | | | | | |
| | 0200 | 154 | 587 | 10.2 | | | | | | | | | | 153.2 | UL OPS | | | | | |
| | 0400 | 154 | 595 | 10.0 | 171 | 12 | 1.0 | .1 | 31 | 4 | | | | 152.6 | UL OPS | | | | | |
| | 0600 | 151 | 598 | 10.1 | | | | | | | | | | 152.2 | TR | | | | | |

TABLE 2e.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | | | | | MK V | | | | PORTABLE | | | | ATM CONTROL | | | | | | |
|--------|--------|----------------|-------|-----------------|------|----------------|-----|------------------|------|----------------|-----|------|---|-----------------|--------|----------------|-------|------|----|----------------|----|-------|-----|-------------|-----|--|
| | | O ₂ | | CO ₂ | | N ₂ | | H ₂ O | | O ₂ | | CO | | CO ₂ | | H ₂ | | BARO | | O ₂ | | BLEED | | VENTILATION | | |
| | | mm | mm | mm | mm | mm | mm | mm | mm | mm | ppm | ppm | % | % | mm | mm | mm | mm | in | in | mm | mm | CFH | COM | SEC | |
| 18 Oct | 0800 | 151 | 7.51 | 606 | 10.1 | | | | | | | | | 152.3 | TR | 31.06 | | | | | | | | | | |
| | 1000 | 151 | 7.37 | 613 | 10.0 | | | | | | | | | 152.7 | TR | 31.34 | | | | | | | | | | |
| | 1200 | 151 | 2.83 | 572 | 9.8 | | | | | | | | | 150.9 | TR | 29.45 | | | | | | | | | | |
| | 1400 | 150 | 5.43 | 587 | 10.5 | | | | | | | | | 150.6 | TR | 29.92 | | | | | | | | | | |
| | 1600 | 151 | 6.91 | 600 | 10.1 | 170 | 6 | 1.07 | .1 | 17 | 8 | | | 152.6 | TR | 30.55 | | | | | | | | | | |
| | 1800 | 151 | 7.68 | 606 | 9.7 | | | | | | | | | 151.9 | UL OPS | 30.76 | | | | | | | | | | |
| | 1925 | 154.7 | 4.22 | 596 | 10.0 | | | | | | | | | 156 | UL OPS | | | | | | | | | | | |
| | 2000 | 158.4 | 2.84 | 600 | 10.2 | 175 | 0 | .45 | .1 | 12 | 1 | | | 158 | UL OPS | 30.56 | | | | | | | | | | |
| | 2200 | 152.8 | 6.06 | 595 | 9.9 | | | | | | | | | 154 | UL OPS | | | | | | | | | | | |
| | 19 Oct | 0000 | 148.6 | 7.27 | 586 | 10.3 | 165 | 8 | 1.08 | .1 | 9 | | | | 150 | TR | 29.85 | | | | | | | | | |
| 0200 | | 158.9 | 3.00 | 599 | 10.0 | | | | | | | | | 160 | TR | 30.65 | | | | | | | | | | |
| 0400 | | 156.6 | 6.31 | 607 | 10.1 | 175 | 5.5 | .99 | .1 | 26 | 1 | | | 158 | TR | 31.01 | | | | | | | | | | |
| 0600 | | 156.0 | 7.98 | 613 | 10.1 | | | | | | | | | 158 | TR | | | | | | | | | | | |
| 0800 | | 155.6 | 9.36 | 618 | 10.1 | 175 | 5.9 | 1.5 | .1 | 20 | 3 | | | 156 | TR | 31.45 | | | | | | | | | | |
| 1000 | | 149.0 | 7.84 | 599 | 10.7 | | | | | | | | | 155 | TR | | | | | | | | | | | |
| 1200 | | 156.5 | 4.97 | 603 | 10.4 | 175 | 3 | 1.2 | .1 | 20 | 3 | | | 157 | UL OPS | 30.78 | | | | | | | | | | |

TABLE 2f. - COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | | | | | MK V | | | | PORTABLE | | | | ATM CONTROL | | | |
|--------|------|----------------|------|-----------------|------|----------------|----|------------------|-----|----------------|-----|-----------------|----------------|---------|----------------|----------|------|----------------|----------------|-------------|-----|-------------|--|
| | | O ₂ | | CO ₂ | | N ₂ | | H ₂ O | | O ₂ | CO | CO ₂ | H ₂ | FIR | O ₂ | LOC | BARO | O ₂ | O ₂ | BLEED | | VENTILATION | |
| | | mm | mm | mm | mm | mm | mm | mm | mm | mm | ppm | % | % | ppm/ppm | mm | | in | | | CFH | COM | SEC | |
| 19 Oct | 1400 | 150.6 | 5.97 | 589 | 10.5 | | | | | | | | | 152 | UL OPS | | | | | | | | |
| | 1600 | 158 | 1.68 | 590 | 12.3 | | | 180 | 0 | .25 | .1 | 17 | 8 | 157 | UL OPS | 31.00 | | | | | | | |
| | 1800 | 154.4 | 4.62 | 592 | 10.5 | | | | | | | | | 156 | TR | 30.25 | | | | | | | |
| | 2000 | 153.3 | 7.08 | 600 | 10.2 | | | 190 | 9 | 1.1 | .1 | 8 | .5 | 152 | TR | 30.58 | | | | | | | |
| | 2230 | 154.2 | 5.74 | 597 | 10.1 | | | | | | | | | 155 | TR | 30.42 | | | | | | | |
| 20 Oct | 0000 | 147.8 | 5.49 | 569 | 10.4 | | | 145 | 9 | .89 | .15 | 8 | .8 | 141 | TR | 29.60 | | | | | | | |
| | 0200 | 147.1 | 7.03 | 576 | 10.1 | | | | | | | | | 142 | TR | | | | | | | | |
| | 0400 | 146.9 | 8.54 | 582 | 9.8 | | | 150 | 3 | 1.4 | .05 | 40 | 3 | 150 | TR | 29.62 | | | | | | | |
| | 0600 | 146.5 | 8.83 | 588 | 9.9 | | | | | | | | | 147 | UL OPS | | | | | | | | |
| | 0800 | 145.7 | 8.40 | 592 | 9.8 | | | 150 | 6 | 1.3 | .1 | 15 | 3 | 147 | UL OPS | 29.91 | | | | | | | |
| | 1025 | 146.2 | 6.35 | 598 | 10.5 | | | | | | | | | 153 | UL OPS | | | | | | | | |
| | 1200 | 154.4 | 4.06 | 598 | 09.9 | | | 155 | 5.5 | .6 | .12 | 9 | 13 | 156 | TR | 30.40 | | | | | | | |
| | 1400 | 152.8 | 6.62 | 605 | 09.7 | | | | | | | | | 154 | TR | 30.69 | | | | | | | |
| | 1650 | 152.6 | 4.31 | 587 | 10.1 | | | 145 | 3 | .62 | .02 | 13 | 8 | 154 | TR | 29.91 | | | | | | | |
| | 1800 | 159.1 | 2.55 | 599 | 9.9 | | | | | | | | | 147 | TR | 30.58 | | | | | | | |

TABLE 2g.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | MK V | | | | PORTABLE | | | ATM CONTROL | | | |
|--------|------|-----------------|-----------------|----------------|------------------|----------------|-----|-----------------|----------------|-----|-----------------|----------------|-------|------|----------------|-------|-------------|------|
| | | CO ₂ | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | FI | FI ₂ | O ₂ | LOC | BARO | O ₂ | BLEED | VENTILATION | |
| | | mm | mm | mm | mm | mm | ppm | % | % | ppm | ppm | mm | | in | | CFH | COM | SEC |
| 20 OCT | 2000 | 160.6 | 4.40 | 61.3 | 71.8 | 165 | 8 | .8 | 40 | 11 | 155 | TUNNEL | 31.26 | X | ∅ | | | |
| 20 OCT | 2200 | 161.4 | 6.39 | 62.4 | 9.6 | | | | | | 164 | UL OPS | 31.11 | X | ∅ | | | |
| 20 OCT | 0100 | 150.2 | 7.97 | 59.0 | 7.8 | 158 | 12 | 1.3 | 23 | 4 | 152 | UL OPS | 30.0 | | ∅ | 0100 | 0130 | |
| 21 OCT | 0200 | 155.7 | 3.34 | 58.9 | 9.4 | | | | | | 154 | UL OPS | 30.03 | | ∅ | | | |
| 21 OCT | 0400 | 156.2 | 6.9 | 60.7 | 9.8 | 155 | 8 | 1.1 | 22 | 3 | 158 | UL OPS | 30.88 | X | ∅ | | | |
| 21 OCT | 0600 | 153.8 | 8.13 | 60.4 | 9.7 | 155 | 8 | 1.34 | 23 | 3 | 155 | UL OPS | 30.58 | X | ∅ | | | |
| 21 OCT | 0800 | 147.6 | 7.65 | 58.8 | 9.8 | | | | | | 150 | | 29.91 | X | X | X | | |
| 21 OCT | 1000 | 147.6 | 8.71 | 59.4 | 9.8 | 148 | 12 | 1.48 | 38 | 3 | 148 | UL OPS | 30.11 | X | X | X | | |
| 21 OCT | 1200 | 145 | 8.49 | 59.0 | 9.6 | | | | | | 146 | UL OPS | 29.6 | | ∅ | 1200 | 1200 | |
| 21 OCT | 1400 | 151 | 7.8 | | | 151 | 7.8 | | | | | | | | | | | |
| 21 OCT | 1400 | 147.1 | 4.57 | 56.9 | 7.8 | 157 | 3 | .78 | 12 | 5 | 149 | UL OPS | 28.76 | X | | 5000 | 1310 | 1435 |
| 21 OCT | 1600 | 156.1 | 3.01 | 57.0 | 9.3 | 157 | 2.5 | .53 | 7 | 1 | 158 | UL OPS | 30.07 | X | | | | |
| 21 OCT | 1600 | | | | | | | | | | | | | | | | | |
| 21 OCT | 1800 | 154.5 | 6.58 | 60.2 | 9.6 | 157 | | | | | 157 | UL OPS | 30.6 | X | ∅ | | | |
| 21 OCT | 2000 | 152.7 | 7.15 | 59.9 | 9.6 | 190 | 6 | 1.2 | 10 | 1 | 149 | UL OPS | 30.6 | X | ∅ | | | |
| 21 OCT | 2200 | 148.1 | 5.96 | 58.7 | 9.9 | | | | | | 148 | UL OPS | 30.4 | X | ∅ | 0100 | 0130 | |

TABLE 2h.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | MK V | | | | PORTABLE | | | ATM CONTROL | | | |
|--------|------|-----------------|-----------------|----------------|------------------|----------------|-----|-----------------|----------------|------|-----|----------------|-----|-------|----------------|-------|-------------|------------|
| | | CO ₂ | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | Fill | Flt | O ₂ | LOC | BARO | O ₂ | BLEED | VENTILATION | |
| | | mm | mm | mm | mm | mm | ppm | % | % | ppm | ppm | mm | mm | in | mm | CFH | COM | SEC |
| 22 Oct | 0000 | 1545 | 320 | 588 | 9.5 | 157 | 3 | .43 | .1 | 38 | 1 | 156 | OPS | 29.96 | X | φ | | |
| 22 Oct | 0200 | 1530 | 601 | 597 | 9.5 | | | | | | | | OPS | 30.30 | X | φ | | |
| 22 Oct | 0400 | 1532 | 637 | 605 | 9.6 | 155 | 10 | .99 | .14 | 10 | 4 | 155 | OPS | 30.60 | X | φ | | |
| 22 Oct | 0600 | 1442 | 503 | 577 | 9.8 | | | | | | | 139 | OPS | 29.15 | X | φ | | |
| 22 Oct | 0500 | 1443 | 476 | 585 | 9.9 | 143 | 11 | .72 | .1 | 6 | 6 | 139 | OPS | 29.47 | X | φ | | |
| " | 1800 | 1463 | 485 | 603 | 9.8 | | | | | | | 140 | OPS | 20.26 | X | φ | | |
| 22 Oct | 1200 | 1478 | 500 | 606 | 9.9 | 152 | 17 | .8 | .1 | 11 | 5 | 143 | OPS | 30.4 | X | φ | | SUBTRACTED |
| 20 Oct | 1400 | 1581 | 1.59 | 587 | 9.5 | | | | | | | | | | X | φ | | |
| 20 Oct | 1600 | 1570 | 3.52 | 600 | 9.4 | 162 | 10 | .8 | .1 | 24 | 8 | 150 | OPS | 30.1 | X | φ | | SUBTRACTED |
| 22 Oct | 1800 | 1567 | 5.18 | 607 | 9.5 | | | | | | | 150 | OPS | 30.66 | X | φ | | |
| 22 Oct | 2000 | 1557 | 5.64 | 611 | 9.6 | 156 | 11 | .92 | .1 | 11 | 3 | 150 | OPS | 30.63 | X | φ | | |
| 22 Oct | 2200 | 1537 | 5.79 | 610 | 9.6 | 152 | 11 | .79 | .1 | 11 | 4 | 147 | OPS | 30.82 | X | φ | | AIRC VARY |
| 23 Oct | 0000 | 1488 | 749 | 597 | 9.6 | 152 | 11 | 1.29 | .1 | 11 | 4 | 143 | OPS | 30.25 | X | φ | | |
| 23 Oct | 0200 | 1494 | 9.08 | 607 | 9.5 | | | | | | | 150 | OPS | 30.62 | X | φ | | |
| 23 Oct | 0400 | 1510 | 8.94 | 615 | 9.5 | 154 | 11 | 1.54 | .1 | 10 | 8 | 145 | OPS | 30.60 | X | φ | | |
| 23 Oct | 0600 | 1521 | 749 | 620 | 9.7 | | | | | | | 146 | OPS | 30.6 | X | φ | | |

M M M M S S M M

TABLE 21.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | MK V | | | | PORTABLE | | | | ATM CONTROL | | | | | |
|--------|------|----------------------|-----------------------|----------------------|------------------------|----------------------|-----------|----------------------|---------------------|----------------------------|-----------------------|----------------------|---------|-------------|---------------------|---------------------|-------|----------------------|-----|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | Fi ₂ ppm/ppm | Fi ₂ mm | O ₂ mm | LOC | BARO | S ₂ % | S ₂ % | BLEED | VENTILATION | |
| | | | | | | | | | | | | | | in | | | CFH | COM | SEC |
| 2300CT | 0800 | 153.5 | 6.11 | 596 | 9.6 | 156 | 15 | 1.02 | .1 | 12 | 13 | 146 | UL OPS. | 30.33 | X | X | 110 | | |
| 2300CT | 1000 | 154.4 | 5.40 | 601 | 9.9 | | | | | | 148 | UL OPS | 30.33 | X | X | 110 | | BEING REDUCED TO 100 | |
| 2300CT | 1800 | 150.1 | 6.62 | 593 | 9.4 | 150 | 11 | 1.12 | .15 | 28 | 4 | 145 | UL OPS | 30.60 | X | X | | | |
| 2300CT | 2000 | 157.2 | 5.54 | 602 | 9.7 | | | | | | 145 | UL OPS | 30.33 | | | | | | |
| 2300CT | 0300 | 153.5 | 6.93 | 613 | 9.3 | 153 | 14 | 1.2 | .15 | 36 | 8 | 155 | UL OPS | 31.05 | X | X | AFT | 50 | |
| 2400CT | 0600 | 147.1 | 7.25 | 595 | 10.9 | 150 | 15 | 1.1 | .1 | 20 | 1 | 145 | UL OPS | 31.05 | X | X | AFT | 50 | |
| 2400CT | 0800 | 147.7 | 6.34 | 402 | 10.3 | 150 | 15 | 1.1 | .1 | 20 | 1 | 142 | UL OPS | 30.34 | X | X | 50 | AFT | |
| 2400CT | 1000 | 148.1 | 5.32 | 608 | 10.1 | | | | | | | 142 | UL OPS | 30.56 | X | X | 50 | AFT | |
| 2400CT | 1300 | 145.0 | 5.66 | 611 | 10.3 | 152 | 11 | 1.59 | .1 | 11 | 6 | 142 | UL OPS | 30.68 | X | X | | | |
| 2400CT | 1700 | 143.8 | 6.13 | 594 | 10.2 | 155 | 10 | 1.1 | .15 | 12 | 10 | 145 | UL OPS | 29.82 | X | X | 150 | AFT | |
| 2400CT | 2200 | 147.5 | 7.94 | 609 | 10.3 | 150 | 11 | 1.48 | .15 | 12 | 28 | 148 | UL OPS | 30.68 | X | X | 130 | AFT | |
| 2500CT | 0000 | 149.4 | 8.80 | 617 | 10.1 | 153 | 15 | 2.4 | .1 | 24 | 18 | 144 | UL OPS | 31.10 | X | X | 130 | AFT | |
| 2500CT | 0200 | 145.4 | 7.66 | 598 | 10.3 | | | | | | | 138 | UL OPS | 30.16 | X | X | AFT | 130 | |
| 2500CT | 0400 | 140.2 | 6.50 | 582 | 10.5 | 143 | 12 | 1.12 | .1 | 8 | 0 | 140 | UL OPS | 29.82 | X | X | AFT | 150 | |
| 2500CT | 0715 | 147.3 | 5.78 | 595 | 10.0 | 145 | 15 | 1.1 | .1 | 24 | 4 | 146 | UL OPS | 30.00 | | | | | |
| 2500CT | 1100 | 150.3 | 6.16 | 606 | 10.9 | 154 | 11 | 1.1 | .1 | 40 | 6 | 145 | UL OPS | 30.50 | X | X | | | |

SWITCH IN
SUITABLE POSITION

TABLE 2j.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

TEST SWITCH LEFT IN TP-16 POSITION
BY MISTAKE

MASS SPEC MK V PORTABLE ATM CONTROL

| DATE | TIME | O ₂ | | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | FIR | O ₂ | LCC | BAPO | C ₂ H ₂ | C ₂ H ₄ | BLEED | | VENTILATION | |
|-------------------|-----------------|------------------|-----------------|-----------------|----------------|------------------|----------------|---------------|-----------------|----------------|---------------|----------------|----------------|------------------|-------------------------------|-------------------------------|----------------|----------------|-------------|----|
| | | mm | mm | | | | | | | | | | | | | | mm | mm | mm | mm |
| 25 OCT | 1200 | 151.4 | 6.07 | 6.12 | 612 | 48.8 | 154 | 11 | 6.1 | 10 | 20 | 18 | 147 | 30.86 | X | X | AFT | 120 | | |
| 25 OCT | 1200 | 151.4 | 6.07 | 6.12 | 612 | 48.8 | 154 | 11 | 6.1 | 10 | 20 | 18 | 147 | 30.86 | X | X | AFT | 120 | | |
| 25 OCT | 1400 | 146.0 | 4.95 | 5.90 | 590 | 49.3 | | | | | | | 145 | 29.72 | X | X | AFT | 120 | | |
| 25 OCT | 1600 | 147.0 | 4.71 | 5.94 | 594 | 48.7 | 150 | 12 | .83 | 10 | 23 | 4 | 145 | 29.90 | X | X | AFT | 120 | | |
| 25 OCT | 1800 | 148.8 | 5.20 | 6.00 | 600 | 49.2 | | | | | | | 140 | 30.28 | X | X | AFT | 120 | | |
| 25 OCT | 2000 | 149.8 | 6.02 | 6.05 | 605 | 49.3 | 153 | 10 | 1.1 | .15 | 12 | 8 | 141 | 30.54 | X | X | AFT | 120 | | |
| 25 OCT | 2200 | 146.2 | 7.11 | 5.87 | 587 | 49.3 | | | | | | | 138 | 29.60 | X | X | AFT | 120 | | |
| 26 OCT | 0000 | 146.5 | 7.52 | 5.86 | 586 | 50.1 | 190 | 13 | 1.38 | .1 | 10 | 24 | 141 | 29.67 | X | X | AFT | 120 | | |
| 26 OCT | 0400 | 149.3 | 5.63 | 5.97 | 597 | 51.1 | 155 | 13 | 1 | .15 | 8 | 9 | 151 | 30.20 | X | X | AFT | 120 | | |
| 26 OCT | 0600 | 150.1 | 5.46 | 6.02 | 602 | 50.2 | | | | | | | 146 | 30.40 | X | X | AFT | 100 | | |
| 26 OCT | 0800 | 151.9 | 6.70 | 6.10 | 610 | 50.6 | 155 | 14 | 1.2 | .10 | 18 | 19 | 150 | 30.83 | X | X | AFT | 100 | | |
| 26 OCT | 1000 | 145.2 | 6.43 | 5.86 | 586 | 50.3 | | | | | | | 142 | 29.60 | X | X | AFT | 100 | | |
| 26 OCT | 1200 | 147.5 | 6.50 | 5.93 | 593 | 50.6 | 70 | 12.5 | 2.55 | .9 | 120 | 200 | 142 | 30.00 | X | X | AFT | 150 | | |
| 26 OCT | 1400 | 149.4 | 5.27 | 5.99 | 599 | 51.0 | | | | | | | 145 | 30.05 | X | X | AFT | 150 | | |
| 26 OCT | 1600 | 150.8 | 5.8 | 6.05 | 605 | 50.5 | 155 | 12 | 1 | .1 | 15 | 11 | 145 | 30.56 | X | X | AFT | 60 | | |
| 26 OCT | 1800 | 145 | 5.09 | 6.13 | 613 | 51.6 | 190 | 12 | .9 | .15 | 20 | 4 | 145 | 30.85 | X | X | AFT | 60 | | |
| 26 OCT | 1900 | 152 | 5.01 | 6.15 | 615 | 51.4 | 155 | 12 | .85 | .1 | 18 | 4 | 152 | 30.94 | X | X | AFT | 60 | | |

M

TABLE 2k.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | | | | | MK V | | | PORTABLE | | | ATM CONTROL | | |
|---------|------|----------------|-----------------|----------------|------------------|----------------|------|-----------------|----------------|-----|--------|----------------|-----|------|----------------|----------------|-------------|-------------|--|--|
| | | O ₂ | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | FiI | FiR | O ₂ | LOC | BARO | O ₂ | O ₂ | VENTILATION | | | |
| | | mm | mm | mm | mm | ppm | % | % | ppm | mm | mm | mm | mm | in | mm | mm | mm | mm | | |
| 26 OCT. | 0200 | 146.9 | 9.46 | 605 | 51.6 | 150 | 1.78 | 1.5 | 52 | 145 | UL OPS | 30.44 | X | | | | | | | |
| 27 OCT. | 0400 | 147.9 | 9.60 | 610 | 52.4 | | | | | 146 | UL OPS | 30.69 | X | | | | | | | |
| 27 OCT. | 0600 | 150.2 | 9.44 | 617 | 52.6 | | | | | 145 | UL OPS | 31.08 | | | | | | | | |
| 27 OCT. | 0800 | 145.2 | 8.55 | 550 | 52.0 | 149 | 1.6 | 1.31 | 34 | 138 | UL OPS | 29.78 | X | | | | | | | |
| 27 OCT. | 1000 | 149.2 | 7.55 | 597 | 52.0 | | | | | 142 | UL OPS | 30.12 | X | | | | | | | |
| 27 OCT. | 1100 | 144 | 6.11 | 530 | 52.0 | REG-15 | 1.1 | 1.20 | 13 | 144 | UL OPS | 29.27 | X | | | | | | | |
| 27 OCT. | 1520 | 145.5 | 5.85 | 587 | 52.1 | 150 | 1.1 | 1.23 | 8 | 141 | UL OPS | 29.57 | X | | | | | | | |
| 27 OCT. | 1800 | 147.2 | 6.87 | 594 | 52.4 | 148 | 1.20 | 1.0 | 12 | | | 29.94 | X | | | | | | | |
| 27 OCT. | 2000 | 148.7 | 7.94 | 600 | 53.1 | | | | | | | 30.30 | X | | | | | | | |
| 27 OCT. | 2200 | 151.2 | 7.46 | 605 | 52.9 | 152 | 1.28 | 1.0 | 33 | 15 | | 30.56 | X | | | | | | | |
| 28 OCT. | 0000 | 152.8 | 6.73 | 610 | 53.1 | 158 | 1.2 | 1.25 | 15 | 15 | | 30.82 | X | | | | | | | |
| 28 OCT. | 0200 | 154.6 | 6.3 | 617 | 53.3 | | | | | 158 | UL OPS | 30.16 | X | | | | | | | |
| 28 OCT. | 0400 | 146.9 | 5.68 | 587 | 52.8 | 150 | 1 | 1.25 | 15 | 15 | UL OPS | 29.60 | X | | | | | | | |
| 28 OCT. | 0600 | 148.3 | 5.51 | 593 | 52.8 | | | | | 157 | UL OPS | 29.95 | X | | | | | | | |
| 28 OCT. | 0800 | 151.5 | 6.16 | 606 | 53.0 | 155 | 1.1 | 1.20 | 6 | 55 | UL OPS | 30.66 | X | | | | | | | |
| 28 OCT. | 1200 | 153.7 | 7.4 | 614 | 53.2 | 155 | 1.31 | 1.0 | 37 | 163 | UL OPS | 31.08 | X | | | | | | | |

TABLE 2.L - COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | MK V | | | | PORTABLE | | | ATM CONTROL | | | | | |
|--------|------|-------------------|--------------------|-------------------|---------------------|-------------------|--------|-------------------|------------------|----------|---------|-------------------|-------------|----------|--------------------------------|-----------|-----------------|-----|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | Fill ppm | FiR ppm | O ₂ mm | LOC | BAIRD in | CO ₂ H ₂ | BLEED CFH | VENTILATION COM | SEC |
| 28 OCT | 1400 | 148 | 7.83 | 592 | 54.7 | | | | | | | UL OPS | 29.94 | X | | AFT 110 | | |
| 28 OCT | 1600 | 149.1 | 8.13 | 596 | 54.7 | 155 | 1.51 | .10 | 10 | 62 | 160 | UL OPS | 30.18 | X | | AFT 110 | | |
| 28 OCT | 1900 | 150.3 | 7.30 | 602 | 55.1 | | | | | | 755 | UL OPS | 30.52 | X | | AFT 110 | | |
| 28 OCT | 2000 | 150.8 | 6.69 | 606 | 55.3 | 155 | 1.2 | .10 | 10 | 19 | 155 | UL OPS | 30.60 | X | | AFT 110 | | |
| 28 OCT | 2200 | 152.6 | 6.22 | 611 | 55.2 | | | | | | 155 | UL OPS | 30.83 | X | | AFT 110 | | |
| 29 OCT | 0130 | 145.1 | 6.56 | 587 | 54.8 | 150 | 1 | .1 | 28 | 10 | 148 | UL OPS | 29.88 | X | X | 100 | | |
| 29 OCT | 0600 | 148.2 | 6.7 | 599 | 55.7 | 157 | | | | | 152 | UL OPS | 29.76 | X | X | 120 AFT | | |
| 29 OCT | 0800 | 150.4 | 7.28 | 608 | 56.0 | 151 | 1.3 | .1 | 22 | 19 | 153 | UL OPS | 30.69 | X | | 120 AFT | | |
| 29 OCT | 1000 | 144.0 | 7.36 | 588 | 55.8 | | | | | | 143 | UL OPS | | X | X | 140 AFT | | |
| 29 OCT | 1200 | 146.1 | 8.20 | 590 | 56.5 | | | | | | 150 | UL OPS | 29.80 | X | X | 140 AFT | | |
| 29 OCT | 1530 | 152.7 | 4.09 | 593 | 55.1 | 155 | 0.68 | .1 | 28 | 8 | 154 | UL OPS | 30.66 | X | X | 140 AFT | | |
| 29 OCT | 2230 | 156.1 | 3.09 | 589 | 56.2 | 165 | 0.5 | .1 | 30 | 4 | 165 | UL OPS | 29.99 | X | | | | |
| 30 OCT | 2400 | 155.8 | 4.37 | 593 | 59.2 | 160 | 0.7 | .13 | 26 | 1 | 158 | UL OPS | 30.11 | X | X | 100 AFT | | |
| 30 OCT | 0200 | 155.2 | 5.78 | 592 | 87.0 | | | | | | 158 | UL OPS | 30.11 | X | X | 100 AFT | | |
| 30 OCT | 0400 | 156.6 | 3.62 | 589 | 56.8 | 160 | 0.62 | .1 | 22 | 6 | 160 | UL OPS | 30.11 | X | X | 100 AFT | | |
| 30 OCT | 0600 | 152.2 | 4.75 | 574 | 56.7 | | | | | | 155 | UL OPS | 29.25 | X | X | 100 AFT | | |

TABLE 2m.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | MK V | | | | PORTABLE | | | ATM CONTROL | | | | |
|--------|------|----------------------|-----------------------|----------------------|------------------------|----------------------|-----------|----------------------|---------------------|------------------------|----------------------|-----|-------------|---------------------|--------------|------------------------|---|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | FI ₂ ppm | O ₂ mm | LOC | BARO in | O ₂ % | BLEED CFH | VENTILATION COM SEC | |
| 30 Oct | 0840 | 155 | 6.8 | 588 | 56.9 | 166 | 11 | 1.25 | 0.12 | 21 | 34 | 155 | UL OPS | 30.06 | X | AFT 100 | M |
| 30 Oct | 1000 | 152.6 | 7.0 | 588 | 56.5 | | | | | | | 155 | UL OPS | | X | AFT 100 | M |
| 30 Oct | 1100 | 151.2 | 11.67 | 589 | 58.2 | 152 | 11 | 2.1 | .1 | 10 | 13 | 153 | UL OPS | 30.08 | X | AFT 70 | M |
| 30 Oct | 2200 | 147.5 | 7.90 | 578 | 58.4 | 150 | 11 | 1.41 | .10 | 22 | 28 | 156 | UL OPS | 29.30 | XX | AFT 100 | M |
| 31 Oct | 0000 | 153.3 | 7.03 | 599 | 58.5 | 165 | 11 | 1.25 | .1 | 22 | 26 | 154 | UL OPS | 30.32 | X | AFT 100 | M |
| 31 Oct | 0400 | 149.0 | 5.33 | 583 | 58.2 | 153 | 11 | .95 | .1 | 10 | 3 | 153 | UL OPS | 29.46 | X | AFT 100 | M |
| 31 Oct | 0600 | 150.4 | 5.4 | 590 | 58.7 | 152 | 9 | 1.0 | .1 | 21 | 11 | 154 | UL OPS | 29.85 | X | AFT 120 | M |
| 31 Oct | 0800 | 153.7 | 5.42 | 594 | 58.9 | 152 | 6 | 1.08 | .1 | 9 | 7 | 156 | UL OPS | 29.89 | X | AFT 120 | M |
| 31 Oct | 0940 | 156.6 | 2.51 | 589 | 59.0 | | | | | | | 165 | UL OPS | 29.95 | X | AFT 100 | M |
| 31 Oct | 1160 | 158.0 | 3.99 | 598 | 59.1 | 190 | 6 | .7 | .1 | 24 | 8 | 165 | UL OPS | 29.99 | X | AFT 100 | M |
| 31 Oct | 1400 | 158.3 | 5.56 | 604 | 59.8 | | | | | | | 165 | UL OPS | 30.70 | X | AFT 100 | M |
| 31 Oct | 2000 | 160.1 | 6.56 | 611 | 59.6 | 192 | 8 | 1.2 | .1 | 24 | 24 | 165 | UL OPS | 31.04 | X | AFT 100 | M |
| 31 Oct | 2200 | 160.9 | 8.06 | 615 | 59.7 | | | | | | | 164 | UL OPS | 31.30 | X | AFT 100 | M |
| 1 Nov | 0100 | 163.0 | 8.00 | 625 | 60.1 | 187 | 8 | 1.48 | .13 | 40 | 63 | 165 | UL OPS | 31.10 | XX | AFT 100 | M |
| 1 Nov | 0400 | 156.6 | 6.71 | 604 | 60.4 | 160 | 12 | 1.20 | .1 | 12 | 22 | 160 | UL OPS | 30.00 | X | AFT 100 | M |
| 1 Nov | 1200 | 154.2 | 6.63 | 594 | 61.5 | 165 | 12 | 1.2 | .1 | 10 | 24 | 150 | UL OPS | 30.40 | X | AFT 100 | M |

TABLE 2n.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

NOTE: TEST SWITCH LEFT IN ^{OFF} 10 BY MISTAKE

| DATE | TIME | MASS SPEC | | | | MK V | | | | PORTABLE | | | | ATM CONTROL | | | | | |
|-------|------|----------------------|-----------------------|----------------------|------------------------|----------------------|-----------|----------------------|---------------------|------------------------|------------------------|----------------------|-------|-------------|---------------------|----------------------|--------------|--------------------|-----|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | F ₁₁ ppm | F ₁₂ ppm | O ₂ mm | LOC | BARO in | H ₂ % | CO ₂ % | BLEED CFH | VENTILATION COM | SEC |
| 1 Nov | 1600 | 155.4 | 4.50 | 573 | 61.7 | 160 | 8 | .8 | .02 | 26 | 18 | UL OPS | 30.20 | X | X | AFT | | | |
| 1 Nov | 2000 | 153.8 | 7.69 | 591 | 62.0 | | | | | | | UL OPS | 30.10 | | | | | | |
| 1 Nov | 2375 | 152.1 | 9.75 | 594 | 61.8 | | | | | | | UL OPS | 30.25 | | | | | | |
| 1 Nov | 2400 | 152.5 | 10.11 | 597 | 62.1 | 70 | 113 | 2.3 | .25 | 112 | 100 | UL OPS | 30.15 | X | X | φ | | | |
| 2 Nov | 0240 | 152.5 | 7.23 | 604 | 62.0 | | | | | | | UL OPS | 30.15 | X | X | φ | | | |
| 2 Nov | 0440 | 152.4 | 6.20 | 609 | 61.9 | | | | | | | UL OPS | 30.26 | X | X | φ | | | |
| 2 Nov | 0600 | 145.6 | 5.80 | 584 | 62.5 | | | | | | | UL OPS | 29.42 | X | X | φ | | | |
| 2 Nov | 1000 | 147.2 | 5.72 | 592 | 62.4 | 147 | 11 | 1.8 | .1 | 20 | 8 | UL OPS | 29.83 | X | X | 0 | | | |
| 2 Nov | 1000 | 147.5 | 6.64 | 598 | 62.8 | | | | | | | UL OPS | 30.18 | X | X | 0 | | | |
| 2 Nov | 1400 | 144.0 | 5.58 | 585 | 62.8 | 154 | 8 | 1% | .18 | 38 | 0 | UL OPS | 29.43 | X | X | AFT | 130 | | |
| 2 Nov | 1800 | 145.2 | 5.95 | 582 | 62.7 | | | | | | | UL OPS | 29.34 | X | X | AFT | 120 | | |
| 2 Nov | 2000 | 146.4 | 6.52 | 588 | 63.7 | 143 | 12 | 1.2 | .1 | 10 | 6 | UL OPS | 29.98 | X | X | AFT | 120 | | |
| 3 Nov | 0000 | 155.7 | 1.67 | 584 | 62.4 | 150 | 3 | .3 | .1 | 9 | 5 | UL OPS | 29.68 | X | X | AFT | 120 | | |
| 3 Nov | 0200 | 155.9 | 4.35 | 573 | 63.7 | | | | | | | UL OPS | 30.19 | X | X | AFT | 100 | | |
| 3 Nov | 0400 | 157 | 6.16 | 594 | 63.6 | 196 | 5 | 1.12 | .12 | 27 | 8 | UL OPS | 30.52 | X | X | AFT | 100 | | |
| 3 Nov | 0800 | 158.2 | 6.72 | 605 | 64.0 | 148 | 7 | 1.8 | .1 | 27 | 3 | UL OPS | 30.80 | | | | | | |

M N M

9

TABLE 20. - COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

NOTE TEST SWITCH BY MISTAKE.
MASS SPEC LEFT IN JP-16 MK V

| DATE | TIME | PORTABLE | | | | | | | | | | ATM CONTROL | | | | | | |
|-------|-------|----------------------|-----------------------|----------------------|------------------------|----------------------|-----------|----------------------|---------------------|-------------------------|------------------------|----------------------|-----|------------|----------------|----------------|--------------|------------------------|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | H ₂ O ppm | FI ₂ ppm | O ₂ mm | LOC | BARO in | h ₂ | h ₂ | BLEED CFH | VENTILATION COM SEC |
| NOV 3 | 0930 | 148 | 4.60 | 559 | 637 | 148 | 2 | .8 | .1 | 27 | 3 | 153 | ops | 28.48 | X | X | NFT 120 | |
| NOV 3 | 1123φ | 149.6 | 6.52 | 570 | 644 | 152 | 3 | 1.2 | .1 | 10 | 3 | 155 | ops | 28.75 | X | X | 120 AFT | |
| NOV 3 | 160φ | 148.7 | 9.19 | 568 | 617 | 150 | 3 | 1.19 | .15 | 40 | 16 | 156 | ops | 29.03 | X | X | 120 AFT | |
| 3 NOV | 164φ | 154.1 | 4.78 | 582 | 642 | 155 | 3 | .81 | .1 | 15 | 1 | 157 | ops | 29.64 | X | X | 0 | |
| 3 NOV | 200φ | 153.3 | 6.24 | 588 | 652 | 155 | 8 | 1.2 | .1 | 14 | 8 | 157 | ops | 29.90 | X | X | 0 | |
| 3 NOV | | 145.6 | 8.06 | 671 | 653 | 175 | 8.5 | 1.48 | .1 | 12 | 4 | 157 | ops | 29.0 | X | X | | |
| 4 NOV | 080φ | 151 | 5.63 | 589 | 655 | 190 | 8 | 1.01 | .1 | 20 | 11 | 158 | ops | 29.80 | X | X | 130 AFT | |
| 4 NOV | 120φ | 1425 | 6.67 | 574 | 658 | 160 | 8 | 1.1 | .1 | 9 | 3 | 150 | ops | 29.08 | X | X | 130 AFT | |
| 4 NOV | 144φ | 146.8 | 6.67 | 577 | 666 | | | | | | | 151 | ops | 29.37 | X | X | 130 AFT | |
| 4 NOV | 160φ | 151 | 7.02 | 586 | 659 | 148 | 8 | 1.25 | .1 | 20 | 15 | 154 | ops | 29.76 | X | X | 130 AFT | |
| 4 NOV | 220φ | 153.7 | 3.35 | 582 | 660 | 155 | 2 | .6 | .15 | 10 | 3 | 157 | ops | 29.64 | | | | |
| 4 NOV | 240φ | 152.9 | 5.19 | 584 | 661 | 152 | 8 | 1. | .1 | 42 | 8 | 160 | ops | 29.74 | X | X | 110 AFT | |
| 5 NOV | 020φ | 152.6 | 5.96 | 580 | 661 | | | | | | | 160 | ops | 29.70 | | | | |
| 5 NOV | 040φ | 152.5 | 7.43 | 584 | 665 | 152 | 8 | 1.3 | .1 | 24 | 32 | 160 | ops | 29.70 | X | X | 110 AFT | |
| 5 NOV | 2600 | 147.2 | 6.95 | 562 | 666 | 150 | 8 | 1.28 | .33 | 30 | 24 | 154 | ops | 28.88 | X | X | 110 AFT | |
| 5 NOV | 080φ | 152.5 | 7.4 | 583 | 667 | | | | | | | 155 | ops | 29.75 | X | X | 110 AFT | |

TABLE 2p.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | | | | | MK V | | | | PORTABLE | | | ATM CONTROL | |
|-------|------|----------------|------|-----------------|----------------|------------------|----------------|------|-----------------|----------------|-----------------|-----------------|----------------|-------|------|-----------------|-----------------|-------|-------------|--|
| | | O ₂ | | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | Fi ₂ | Fi ₂ | O ₂ | LOC | BARO | CO ₂ | CO ₂ | BLEED | VENTILATION | |
| | | mm | mm | mm | mm | mm | ppm | % | % | ppm | ppm | mm | | in | % | % | CFH | COM | SEC | |
| 5 NOV | 1500 | 153.2 | 8.47 | 587 | 10.6 | 151 | 11 | 1.5 | 0.9 | 11 | 50 | 156 | 4L OPS | 29.74 | X | X | 110 | AFT | | |
| 5 NOV | 1700 | 153.1 | 7.20 | 590 | 9.7 | | | | | | | 156 | 0PS | 30.08 | | | | | | |
| 5 NOV | 1800 | 153.3 | 7.32 | 584 | 9.6 | 150 | 8 | 1.35 | 1.38 | 18 | 157 | 0PS | 29.76 | X | X | 124 | AFT | | | |
| 5 NOV | 1840 | 155.1 | 7.91 | 592 | 9.7 | | | | | | | 160 | 0PS | 29.65 | X | X | 124 | AFT | | |
| 5 NOV | 1940 | 156.2 | 8.26 | 599 | 9.9 | 153 | 6 | 1.42 | 0.8 | 16 | 13 | 160 | 0PS | 29.74 | X | X | 130 | AFT | | |
| 5 NOV | 2200 | 158.1 | 7.25 | 605 | 9.9 | | | | | | | 162 | 0PS | 29.68 | X | X | 110 | AFT | | |
| 6 NOV | 0200 | 159.5 | 6.37 | 610 | 10.2 | 157 | 8 | 1.19 | 1.14 | 28 | 13 | 162 | 0PS | 29.80 | X | X | 110 | AFT | M | |
| 6 NOV | 0230 | 153.6 | 5.71 | 569 | 9.9 | | | | | | | 158 | 0PS | 29.94 | X | X | 110 | AFT | | |
| 6 NOV | 0400 | 154.7 | 6.44 | 573 | 9.9 | 150 | 8 | 1.17 | 1.20 | 13 | 13 | 156 | 0PS | 29.80 | X | X | 100 | AFT | M | |
| 6 NOV | 0600 | 155.6 | 7.16 | 591 | 9.4 | 155 | 4 | 1.28 | 1.8 | 28 | 22 | 160 | 0PS | 30.50 | X | X | 100 | AFT | | |
| 6 NOV | 0800 | 157 | 7.95 | 608 | 10.1 | 150 | 12 | 1.30 | 1.35 | 20 | 13 | 160 | 0PS | 30.80 | X | X | 110 | AFT | | |
| 6 NOV | 1000 | 151.1 | 7.70 | 589 | 11.1 | 150 | 12 | 1.30 | 1.35 | 20 | 13 | 159 | 0PS | 29.76 | X | X | 110 | AFT | | |
| 6 NOV | 1200 | 150.9 | 7.28 | 591 | 10.5 | 149 | 15 | 1.29 | 1.65 | 10 | 18 | 160 | 0PS | 30.00 | X | X | 125 | AFT | | |
| 6 NOV | 1600 | 158.1 | 6.29 | 612 | 10.1 | 155 | 13 | 1.1 | 2.25 | 4 | 11 | 165 | 0PS | 30.04 | X | X | 125 | AFT | M | |
| 6 NOV | 1800 | 152.6 | 5.60 | 590 | 10.1 | 140 | 12 | 1.01 | 1.61 | 10 | 155 | 0PS | 29.92 | X | X | 90 | AFT | M | | |

TABLE 2q. - COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | MK V | | | | | | PORTABLE | | | | ATM CONTROL | | | |
|-------|------|-----------------|----------------|------------------|----------------|-----|-----------------|----------------|-----|-----|----------------|-------|------|----------------|----------------|-------|-------------|-------------|--|--|--|
| | | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | Fil | Fil | O ₂ | LOC | BARO | H ₂ | O ₂ | BLEED | VENTILATION | | | | |
| | | mm | mm | mm | mm | ppm | % | % | ppm | ppm | mm | | in | | | CFH | COM | SEC | | | |
| 6 Nov | 0000 | 154.4 | 6.85 | 578 | 7.9 | | | | | 157 | UL OPS | 30.41 | X | | 110 | AFT | | | | | |
| 6 Nov | 0200 | 155.6 | 7.37 | 603 | 7.7 | 12 | 1.25 | 15 | 23 | 158 | UL OPS | 30.67 | X | X | 110 | AFT | | | | | |
| 6 Nov | 0000 | 157.2 | 6.65 | 610 | 10.2 | | | | | 62 | UL OPS | 31.00 | | | | | | | | | |
| 7 Nov | 0200 | 158 | 6.28 | 614 | 10.1 | 11 | 1.08 | 19 | 20 | 162 | UL OPS | 31.12 | X | X | 110 | AFT | | | | | |
| 7 Nov | 0400 | 150.9 | 6.67 | 586 | 10.0 | | | | | 154 | UL OPS | 29.67 | | | | | | | | | |
| 7 Nov | 0600 | 150.9 | 6.22 | 585 | 9.8 | | | | | 157 | UL OPS | 29.59 | X | X | 40 | AFT | | | | | |
| 7 Nov | 0800 | 154.9 | 3.30 | 586 | 9.6 | 6 | .68 | 2 | 17 | 162 | UL OPS | 29.76 | X | | 100 | AFT | | | | | |
| 7 Nov | 1400 | 153.8 | 8.46 | 608 | 9.9 | | | | | 160 | UL OPS | 31.00 | X | X | 100 | AFT | | | | | |
| 7 Nov | 1600 | 155.2 | 7.27 | 597 | 9.8 | 12 | 1.3 | 1.1 | 10 | 158 | UL OPS | 30.34 | X | X | 100 | AFT | | | | | |
| 7 Nov | 2000 | 155.3 | 3.13 | 587 | 9.3 | 12 | 1.32 | 1.1 | 10 | 157 | UL OPS | 29.78 | | | | | | | | | |
| 7 Nov | 2200 | 156.2 | 5.22 | 594 | 9.8 | 0 | .7 | 4.1 | 10 | 160 | UL OPS | 30.26 | | | | | | | | | |
| 8 Nov | 0325 | 153.9 | 7.07 | 582 | 9.6 | 0 | .9 | .5 | 5 | 166 | UL OPS | 29.62 | X | X | AFT | 120 | | | | | |
| 8 Nov | 0400 | 154.5 | 7.13 | 583 | 9.7 | 12 | .96 | 4.6 | 4 | 164 | UL OPS | 29.77 | X | X | AFT | 120 | | | | | |
| 8 Nov | 0600 | 157.4 | 6.49 | 596 | 9.8 | | | | | 160 | UL OPS | 30.30 | X | X | 95 | AFT | | | | | |
| 8 Nov | 0800 | 158.9 | 6.01 | 599 | 10.0 | 8 | .8 | 1.2 | 4 | 161 | UL OPS | 30.51 | X | X | 95 | AFT | | | | | |
| 8 Nov | 1600 | 150.7 | 3.20 | 570 | 9.6 | 5 | .4 | .3 | 10 | 155 | UL OPS | 28.95 | X | X | 100 | AFT | | | | | |

TABLE 2r.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | | | | | MK | | | | | PORTABLE | | | | ATM CONTROL | | |
|--------|------|----------------------|-----------------------|----------------------|------------------------|----------------------|-----------|----------------------|---------------------|------------------------|------------------------|----------------------|-------|------------|----------------|----------------|--------------|------------------------|--|--|-------------|--|--|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | FI ₁ ppm | FI ₂ ppm | O ₂ mm | LOC | BARO in | H ₂ | O ₂ | BLEED CFH | VENTILATION COM SEC | | | | | |
| 8 NOV | 1800 | 155 | 4.61 | 589 | 9.4 | 180 | | | | | 162 | UL ORS | 29.85 | X | | | | | | | | | |
| 8 NOV | 2000 | 156 | 4.05 | 590 | 9.4 | 180 | 6 | .55 | 4 | 10 | 165 | UL ORS | 29.85 | X | | | | | | | | | |
| 9 NOV | 2200 | 158 | 5.06 | 600 | 9.2 | 184 | 8 | .9 | 15 | 10 | 166 | UL ORS | 29.80 | X | | | | | | | | | |
| 9 NOV | 0000 | 159.3 | 6.76 | 607 | 9.3 | 164 | 8 | .9 | 15 | 10 | 163 | UL ORS | 30.82 | X | | | | | | | | | |
| 9 NOV | 0200 | 160.2 | 6.35 | 610 | 9.4 | 160 | | | | | 160 | UL ORS | 31.00 | X | | | | | | | | | |
| 9 NOV | 0400 | 161.1 | 6.90 | 614 | 9.4 | 163 | 8 | .9 | 15 | 5 | 165 | UL ORS | 31.20 | X | | | | | | | | | |
| 9 NOV | 0600 | 153.1 | 6.64 | 584 | 9.2 | 155 | 6 | .9 | 3 | 5 | 157 | UL ORS | 29.92 | X | | | | | | | | | |
| 9 NOV | 0800 | 158.0 | 3.30 | 595 | 9.1 | 160 | 3 | .45 | 3 | 10 | 161 | UL ORS | 30.20 | X | | | | | | | | | |
| 9 NOV | 1200 | 158 | 5.75 | 601 | 9.2 | 160 | 3 | .78 | 25 | 9 | 166 | UL ORS | 30.52 | X | | | | | | | | | |
| 9 NOV | 1400 | 159.8 | 7.34 | 609 | 9.2 | 159 | | | | | 168 | UL ORS | 30.54 | X | | | | | | | | | |
| 9 NOV | 1600 | 158.3 | 7.87 | 606 | 9.4 | 159 | 12 | 1.0 | 15 | 4 | 167 | UL ORS | 30.50 | X | | | | | | | | | |
| 9 NOV | 2000 | 157.7 | 2.27 | 590 | 8.8 | 160 | 3 | .25 | 1 | 4 | 161 | UL ORS | 29.91 | X | | | | | | | | | |
| 10 NOV | 0000 | 159.0 | 6.19 | 607 | 9.5 | 163 | 8 | .8 | 3 | 4 | 162 | UL ORS | 29.85 | X | | | | | | | | | |
| 10 NOV | 0300 | 160.8 | 8.16 | 613 | 9.6 | 162 | 8 | .95 | 15 | 4 | 162 | UL ORS | 31.20 | X | | | | | | | | | |
| 10 NOV | 0500 | 161 | 8.21 | 620 | 9.8 | 162 | | 1% | | | 163 | UL ORS | 31.50 | X | | | | | | | | | |
| 10 NOV | 0600 | 157.7 | 7.75 | 606 | 9.4 | 160 | | 1% | | | 165 | UL ORS | 31.42 | X | | | | | | | | | |

TABLE 2s. - COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | MK V | | | | PORTABLE | | | | ATM CONTROL | | | |
|--------|------|-------------------|--------------------|-------------------|---------------------|-------------------|--------|-------------------|------------------|----------|---------|-------------------|-------|-------------|------------------|-----------|---------------------|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | Fil ppm | Fil ppm | O ₂ mm | LOC | BARO in | O ₂ % | BLEED CFH | VENTILATION COM SEC |
| 10 Nov | 0800 | 154.6 | 7.87 | 600 | 9.5 | 154 | 8 | .95 | .2 | 3 | 55 | UL OPS | 30.38 | X | AFT 574 | 100 | |
| 10 Nov | 1200 | 155.5 | 5.12 | 576 | 7.3 | 157 | 3 | .7 | .33 | 9 | 21 | UL OPS | 30.16 | X | AFT 100 | | |
| 10 Nov | 1400 | 160.3 | 1.33 | 576 | 7.8 | | | | | | | UL OPS | 30.17 | X | AFT 100 | | |
| 10 Nov | 1600 | 160.7 | 1.05 | 577 | 6.9 | 162 | 8 | .1 | .1 | 9 | 3 | UL OPS | 30.18 | X | AFT 40 | | |
| 10 Nov | 2200 | 164 | 2.50 | 613 | 8.4 | 165 | 6 | .3 | .8 | 8 | 8 | UL OPS | 31.0 | X | AFT 40 | | |
| 10 Nov | 2400 | 163.9 | 4.05 | 617 | 9.1 | 165 | 10 | .55 | .1 | 2 | 13 | UL OPS | 31.12 | X | AFT 40 | | |
| 11 Nov | 0200 | 164.5 | 5.53 | 623 | 9.0 | | | | | | | UL OPS | 31.26 | X | AFT 40 | | |
| 11 Nov | 0400 | 159 | 6.87 | 607 | 9.2 | 162 | 8 | .8 | .1 | 3 | 64 | UL OPS | 31.21 | X | AFT 40 | | |
| 11 Nov | 0600 | 159.1 | 8.22 | 608 | 8.7 | | | | | | | UL OPS | 31.10 | X | AFT 80 | | |
| 11 Nov | 0800 | 160.5 | 7.79 | 618 | 9.2 | 163 | 8 | .9 | .1 | 10 | 82 | UL OPS | 31.32 | X | AFT 80 | | |
| 11 Nov | 1000 | 161.7 | 11.02 | 623 | 9.2 | | | | | | | UL OPS | 31.60 | X | AFT 80 | | |
| 11 Nov | 1300 | 155 | 7.53 | 601 | 9.4 | 157 | 8 | .85 | .1 | 3 | 162 | UL OPS | 30.17 | X | AFT 80 | | |
| 11 Nov | 1600 | 157.7 | 7.81 | 613 | 10.0 | 160 | 8 | .92 | .1 | 10 | 111 | UL OPS | 31.00 | X | AFT 80 | | |
| 11 Nov | 2000 | 158 | 8.99 | 599 | 8.9 | 160 | 4 | .49 | .3 | 9 | 13 | UL OPS | 30.99 | X | φ | | |
| 11 Nov | 2200 | 159.7 | 2.41 | 598 | 8.2 | | | | | | | UL OPS | 31.90 | X | φ | | |

TABLE 2t.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | MK ∇ | | | | PORTABLE | | | ATM CONTROL | | |
|--------|------|-----------------|------|------------------|----------------|----|-----------------|----------------|-----------------|-----------------|----------------|----------|------|-----------------|-------------|-------------|-----|
| | | CO ₂ | CO | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | F ₁₁ | F ₁₂ | O ₂ | LOC | BARO | CO ₂ | BLEED | VENTILATION | |
| | | mm | ppm | mm | mm | % | % | ppm | ppm | mm | mm | UL OPS | in | ppm | CFH | COM | SEC |
| 12 NOV | 0000 | 160 | 1.94 | 8.2 | 162 | 1 | .13 | 4 | 3 | 161 | UL OPS | 30.35 | X | | | | |
| 12 NOV | 0400 | 160.5 | 1.4 | 8.2 | 162 | 5 | .13 | 5 | 3 | 162 | UL OPS | | X | | | | |
| 12 NOV | 0630 | 159 | 2.73 | 8.3 | 163 | 2 | .38 | 6 | 8 | 162 | UL OPS | 30.28 | X | | | | |
| 12 NOV | 0900 | 163.8 | 4.90 | 9.5 | 165 | 2 | .18 | 3 | 3 | 165 | UL OPS | 31.00 | X | | | | |
| 12 NOV | 1100 | 163 | 6.60 | 10 | 165 | 3 | .78 | 2 | 7 | 165 | UL OPS | 31.35 | X | | | | |
| 12 NOV | 1200 | 155.2 | 7.05 | 9.5 | 157 | 3 | .78 | 3 | 8 | 162 | UL OPS | 31.35 | X | | | | |
| 12 NOV | 1400 | 155.4 | 7.6 | 9.6 | | | | | | 1103 | UL OPS | 30.30 | X | | | | |
| 12 NOV | 1600 | 157.7 | 7.75 | 10.0 | 160 | 6 | .94 | 2 | 17 | 166 | UL OPS | 30.95 | X | | | | |
| 12 NOV | 2000 | 155.2 | 7.75 | 10.0 | 160 | 8 | .87 | 2 | 17 | 157 | UL OPS | 30.82 | X | | | | |
| 13 NOV | 0000 | 153.2 | 11.2 | 10.2 | 154 | 7 | 1.1 | 10 | 16 | 154 | UL OPS | 30.72 | X | | | | |
| 13 NOV | 0300 | 154.0 | 11.2 | 10.3 | 157 | 8 | 1.0 | 3 | 16 | 157 | UL OPS | 31.04 | X | | | | |
| 13 NOV | 0500 | 155.4 | 9.8 | 9.7 | 157 | 5 | 1.0 | 10 | 16 | 157 | UL OPS | 31.30 | X | | | | |
| 13 NOV | 0600 | 153.5 | 9.12 | 9.8 | | | | | | 161 | UL OPS | 31.06 | X | | | | |
| 13 NOV | 0800 | 151.0 | 8.74 | 9.9 | 152 | 8 | .85 | 2 | 58 | 160 | UL OPS | 30.44 | X | | | | |
| 13 NOV | 1100 | 152.2 | 8.01 | 9.8 | 157 | 8 | .85 | 2 | 44 | 161 | UL OPS | 31.02 | X | | | | |
| 13 NOV | 2200 | 148.7 | 7.16 | 9.8 | 150 | 9 | .65 | 1 | 13 | 158 | UL OPS | 30.09 | X | | | | |

TABLE 2u.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | MK V | | | | PORTABLE | | | ATM CONTROL | | |
|--------|------|-----------------|------|------------------|----------------|-----------------|----------------|------|-----|----------------|-------|----------|-----------------|-----------------|-------------|-------------|-----|
| | | CO ₂ | CO | H ₂ O | O ₂ | CO ₂ | H ₂ | FiI | FiR | O ₂ | LOC | BARO | CO ₂ | CO ₂ | BLEED | VENTILATION | |
| | | mm | ppm | mm | mm | % | % | ppm | ppm | mm | | in | H ₂ | H ₂ | CFH | COM | SEC |
| 13 NOV | 1600 | 154.3 | 1.58 | 595 | 160 | .18 | .1 | 3 | 4 | 162 | 30.12 | X | | | | | |
| 13 NOV | 1900 | 154.7 | 4.32 | 590 | 157 | .6 | .2 | 5 | 5 | 157 | 29.87 | X | | | | | |
| 13 NOV | 2200 | 154.7 | 6.21 | 588 | 158 | .8 | .15 | 10 | 13 | 158 | 30.20 | X | | | | | |
| 13 NOV | 2400 | 152.4 | 7.72 | 595 | 157 | .9 | .15 | 3 | 24 | 156 | 30.22 | X | | 100 | AFT | | |
| 13 NOV | 0200 | 151.3 | 8.12 | 595 | | | | | | 155 | 30.08 | X | | 100 | AFT | | |
| 14 NOV | 0400 | 152.4 | 7.18 | 599 | 154 | .8 | .1 | 9 | 13 | 156 | 30.21 | X | | 100 | AFT | | |
| 14 NOV | 0600 | 150.5 | 6.94 | 589 | | | | | | 155 | 29.92 | X | | 110 | AFT | | |
| 14 NOV | 0800 | 152.8 | 6.18 | 600 | 155 | .78 | .19 | 19 | 15 | 155 | 30.88 | X | | 110 | AFT | | |
| 14 NOV | 1000 | 155 | 7.29 | 608 | | | | | | 155 | 30.68 | X | | 110 | AFT | | |
| 14 NOV | 1530 | 147.8 | 7.36 | 581 | 150 | .8 | .1 | 25 | 26 | 160 | 29.30 | X | | 110 | AFT | | |
| 14 NOV | 1800 | 159.1 | 1.89 | 594 | 162 | .19 | .45 | 3 | 8 | 166 | 29.69 | X | | 110 | AFT | | |
| 14 NOV | 2100 | 158.9 | 1.77 | 594 | 162 | .25 | .45 | 14 | 8 | 167 | 30.00 | X | | 110 | AFT | | |
| 15 NOV | 0000 | 159.2 | 4.57 | 596 | 160 | .5 | .29 | 29 | 14 | 159 | 30.14 | X | | 110 | AFT | | |
| 15 NOV | 0330 | 157.4 | 1.48 | 595 | 162 | .15 | .5 | 3 | 3.5 | 163 | 30.26 | X | | 110 | AFT | | |
| 15 NOV | 1200 | 155.8 | 3.61 | 587 | 159 | .45 | .1 | 14 | 22 | 160 | 30.08 | X | | 110 | AFT | | |
| 15 NOV | 1630 | 158.2 | 7.23 | 605 | 159 | .72 | .1 | 22 | 22 | 162 | 30.68 | X | | 110 | AFT | | |

TABLE 2V.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | | | | | MK V | | | | PORTABLE | | | ATM CONTROL | | |
|--------|------|----------------------|-----------------------|----------------------|------------------------|----------------------|-----------|----------------------|---------------------|-------------|-----------------------|----------------------|-----|------------|---------------------|----------|-------------|--|-------------|--|--|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | Fill ppm | Fi ₂ mm | O ₂ mm | LOC | BARO in | O ₂ % | BLEED | VENTILATION | | | | |
| | | | | | | | | | | | | | | | CFH | COM | SEC | | | | |
| 15 Nov | 2000 | 152.6 | 7.4 | 587 | 7.3 | 155 | 0 | 0 | .13 | 11 | 8 | 155 | ops | 29.72 | X | 100 | AFT | | | | |
| 16 Nov | 0400 | 153.9 | 6.05 | 588 | 9.0 | 152 | 8 | 7 | .15 | 18 | 173 | 155 | ops | 29.80 | X | 40 | AFT | | | | |
| 16 Nov | 1600 | 157 | 8.38 | 608 | 8.8 | 159 | 6 | 8 | .1 | 12 | 160 | 161 | ops | 29.81 | XX | 80 | AFT | | | | |
| 16 Nov | 1230 | 153.6 | 8.12 | 596 | 7.6 | 158 | 5.5 | 7.3 | .1 | 17 | 157 | 156 | ops | 30.24 | | | | | | | |
| 16 Nov | 1600 | 160.7 | 1.33 | 599 | 6.9 | 160 | 0 | .15 | .13 | 3 | 163 | 163 | ops | 30.26 | X | | | | | | |
| 16 Nov | 2100 | 161.1 | 3.54 | 609 | 8.7 | | 13 | 0 | | | 164 | 164 | ops | 30.80 | | | | | | | |
| 17 Nov | 0000 | 161.4 | 5.94 | 619 | 9.5 | | | | | | 164 | 164 | ops | 31.50 | X | 80 | AFT | | | | |
| 17 Nov | 0200 | 161.9 | 7.57 | 624 | 9.8 | | | | | | 165 | 165 | ops | 31.58 | X | 50 | AFT | | | | |
| 17 Nov | 0400 | 159.9 | 8.81 | 617 | 9.6 | | | | | | 161 | 161 | ops | 31.40 | XX | 80 | AFT | | | | |
| 17 Nov | 0600 | 152.7 | 7.74 | 610 | 9.3 | | | | | | 160 | 160 | ops | 30.86 | XX | 80 | AFT | | | | |
| 17 Nov | 0800 | 159.3 | 4.65 | 617 | 10.1 | 157 | 5.5 | 6.2 | .1 | 19 | 160 | 160 | ops | 31.22 | XX | 80 | AFT | | | | |
| 17 Nov | 2000 | 161.7 | 1.28 | 598 | 7.1 | 160 | 5 | .A | .15 | 10 | 165 | 165 | ops | 30.84 | X | 80 | AFT | | | | |
| 17 Nov | 2200 | 163.6 | 3.41 | 607 | 9.0 | | | | | | 167 | 167 | ops | 30.76 | X | 80 | AFT | | | | |
| 18 Nov | 0000 | 164.7 | 5.20 | 613 | 9.7 | 163 | 5 | .5 | .4 | 20 | 162 | 162 | ops | 31.22 | X | 80 | AFT | | | | |
| 18 Nov | 0600 | 160.3 | 9.10 | 603 | 10 | 160 | 8 | .7 | .15 | 21 | 162 | 162 | ops | 30.75 | X | 80 | AFT | | | | |
| 18 Nov | 1000 | 164 | 6.78 | 621 | 10 | 162 | 8 | .6 | .3 | 12 | 166 | 166 | ops | 31.58 | XX | 80 | AFT | | | | |

TABLE 2w. - COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | MK V | | | | | PORTABLE | | | ATM CONTROL | | VENTILATION |
|--------|------|-----------------|----------------|------------------|----------------|-----|-----------------|----------------|------|-----|----------------|----------|------|-----------------|-----------------|-------|-------------|
| | | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | Fill | FiR | O ₂ | LOC | BARO | CO ₂ | CO ₂ | BLEED | |
| | | mm | mm | mm | mm | ppm | % | % | ppm | ppm | mm | | in | % | % | CFH | |
| 18 NOV | 1200 | 157.9 | 598 | 9.8 | 157 | 10 | .6 | .57 | 10 | 244 | 160 | 30.33 | X | X | 80 | | M |
| 18 NOV | 1400 | 158.8 | 604 | 9.5 | | | | | | | 161 | 30.61 | X | X | 80 | | M |
| 18 NOV | 1600 | 159.8 | 609 | 10.5 | 157 | 10 | .62 | .27 | 244 | 162 | 30.90 | 30.90 | X | X | 80 | | M |
| 18 NOV | 1800 | 161.3 | 616 | 10.4 | | | | | | | 164 | 31.88 | X | X | 80 | | |
| 18 NOV | 2000 | 161.7 | 620 | 10.4 | 162 | 11 | .6 | .21 | 200 | 163 | 31.48 | 31.48 | X | X | 80 | | |
| 19 NOV | 0700 | 157.2 | 589 | 9.5 | 155 | 15 | .49 | .20 | 244 | 159 | | | X | X | 80 | | |
| 19 NOV | 0800 | 158.3 | 595 | 9.5 | 157 | 12 | .55 | .19 | 240 | 160 | | | X | X | 80 | | |
| 19 NOV | 1000 | 161.5 | 618 | 10.4 | 159 | 12 | .6 | .1 | 244 | 165 | | | X | X | 80 | | |
| 19 NOV | 1500 | 153.6 | 593 | 10.4 | 153 | 12 | .5 | .1 | 244 | 156 | | | X | X | 80 | | |
| 19 NOV | 1800 | 157 | 604 | 10.6 | 155 | 15 | .45 | .1 | 244 | 158 | | | X | X | 80 | | |
| 20 NOV | 0200 | 160.3 | 595 | 8.6 | 157 | 17 | .2 | .35 | 16 | 42 | 164 | 30.18 | X | X | 80 | | M |
| 20 NOV | 2600 | 160.2 | 594 | 8.8 | | | | | | | 162 | 30.11 | X | X | 80 | | |
| 20 NOV | 0800 | 160.8 | 594 | 8.5 | 160 | 20 | .15 | .1 | 15 | 36 | 163 | 30.11 | X | X | 80 | | |
| 21 NOV | 0000 | 164.5 | 612 | 9.5 | 163 | 12 | .30 | .2 | 30 | 244 | 165 | 31.16 | X | X | 80 | | M |
| 21 NOV | 0630 | 159.9 | 601 | 9.8 | 160 | 12 | .55 | .2 | 12 | 244 | 161 | 30.62 | X | X | 80 | | |
| 21 NOV | 0930 | 161.1 | 604 | 10.3 | 162 | 12 | .65 | .15 | 31 | 244 | 163 | 31.00 | X | X | 80 | | G |

TABLE 2x.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

$\bar{V}_z = (1.26)$

| DATE | TIME | MASS SPEC | | | | MK V | | | | PORTABLE | | | | ATM CONTROL | | | | | |
|--------|------|-----------------|-----|----------------|------------------|----------------|------|-----------------|----------------|----------|---------|----------------|-------|-------------|-----------------|-----------------|-------|-------------|-----|
| | | CO ₂ | CO | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | FIL | FIR | O ₂ | LOC | BARIO | CO ₂ | CO ₂ | BLEED | VENTILATION | |
| | | mm | ppm | mm | mm | mm | ppm | % | % | ppm | ppm | mm | | in | | | CFH | COM | SEC |
| 21 Nov | 1200 | 162 | 12 | 613 | 10.6 | 162 | 55 | 1 | 7 | 240 | 165 | UL OPS | 31.25 | XX | XX | 80 AFT | | | |
| 21 Nov | 1400 | 155.2 | | 592 | 9.8 | 151 | 95 | | | | 153 OPS | 30.04 | XX | XX | 80 AFT | | | | M |
| | 1600 | 156.5 | | 598 | 9.8 | 154 | 44 | 2 | 17 | 240 | 159 OPS | 32.20 | XX | XX | 80 AFT | | | | M |
| 21 Nov | 2000 | 154.2 | | 586 | 9.7 | 155 | 49 | 1.5 | 9 | 200 | 156 OPS | 32.63 | XX | XX | 40 AFT | | | | |
| 21 Nov | 2200 | 157.1 | | 585 | 8.2 | | | | | | 160 OPS | 32.76 | XX | XX | 110 AFT | | | | |
| 22 Nov | 0400 | 156.9 | | 600 | 9.6 | 170 | 5 | 2.9 | 9 | 240 | 158 OPS | 30.35 | | | | | | | |
| 22 Nov | 0800 | 151.0 | | 588 | 9.7 | 152 | 6 | 1.5 | 9 | 240 | 154 OPS | 29.54 | XX | XX | 80 AFT | | | | |
| 22 Nov | 1000 | 152.9 | | 591 | 9.5 | | 5.5 | | | | 155 OPS | 29.92 | XX | XX | 730 AFT | | | | M |
| 22 Nov | 1200 | 154.6 | | 596 | 9.4 | 155 | 11.5 | 1.3 | 11 | 200 | 155 OPS | 30.15 | XX | XX | 120 AFT | | | | M |
| 22 Nov | 1400 | 155.2 | | 599 | 9.4 | 155 | | | | | 156 OPS | 30.32 | XX | XX | 120 AFT | | | | |
| 22 Nov | 2300 | 162.1 | | 606 | 8.3 | 163 | 16 | 1 | 10 | 240 | 163 OPS | 30.35 | XX | XX | 120 AFT | | | | |
| 22 Nov | 2400 | 163.1 | | 611 | 9.3 | 163 | 5 | 1 | 20 | 240 | 165 OPS | 30.84 | XX | XX | 80 AFT | | | | M |
| 22 Nov | 0200 | 157.9 | | 591 | 9.7 | | | | | | 159 OPS | 30.75 | XX | XX | 80 AFT | | | | M |
| 23 Nov | 0400 | 159.1 | | 597 | 9.8 | 159 | 15 | 1 | 18 | 240 | 163 OPS | 30.34 | XX | XX | 100 AFT | | | | M |
| 23 Nov | 0800 | 160.6 | | 595 | 7.3 | 160 | 3 | 1 | 19 | 240 | 163 OPS | 30.38 | XX | XX | 100 AFT | | | | |
| 23 Nov | 1100 | 168 | | 626 | 9.5 | 168 | 4.5 | 1.3 | 22 | 240 | 170 OPS | 31.90 | | | | | | | |

TABLE 2y. - COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | MK V | | | | PORTABLE | | | | ATM CONTROL | | | | |
|--------|------|-------------------|--------------------|-------------------|---------------------|-------------------|--------|-------------------|------------------|-----------------------|-------------------|--------|---------|-------------------|-------------------|-----------|-----------------|-----|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | FI/F ₂ ppm | O ₂ mm | LOC | BARO in | CO ₂ % | CO ₂ % | BLEED CFH | VENTILATION COM | SEC |
| 23 Nov | 0000 | 161.6 | 2.99 | 601 | 8.9 | 162 | 13 | .25 | .15 | 34 | 165 | UL OPS | | X | | 70 AFT | | |
| 24 Nov | 0000 | 163.3 | 6.98 | 611 | 9.2 | 165 | 11 | .43 | .1 | 300 | 164 | OPS | 31.13 | X | | 90 AFT | | |
| 24 Nov | 0200 | 164.7 | 7.41 | 616 | 9.4 | 165 | 12 | .48 | .1 | 300 | 165 | OPS | 31.43 | X | | 90 AFT | | |
| 24 Nov | 0600 | 163 | 7.04 | 613 | 9.3 | 165 | 12 | .48 | .1 | 300 | 165 | OPS | 31.24 | X | | 90 AFT | | |
| 24 Nov | 0900 | 160 | 1.03 | 607 | 9.2 | 166 | 11 | .45 | .1 | 300 | 161 | OPS | 30.8 | X | | 90 AFT | | |
| 24 Nov | 1100 | 163 | 6.97 | 617 | 9.4 | 166 | 11 | .45 | .1 | 300 | 165 | OPS | 30.3 | X | | 90 AFT | | |
| 24 Nov | 1800 | 159.2 | 6.9 | 605 | 9.1 | 160 | 8 | .44 | .1 | 200 | 161 | OPS | 30.72 | X | | 90 AFT | | |
| 24 Nov | 2200 | 162 | 7.78 | 621 | 9.2 | 162 | 8 | .49 | .1 | 200 | 164 | OPS | 31.99 | X | | 90 AFT | | |
| 25 Nov | 0000 | 155.7 | 7.35 | 591 | 9.3 | 158 | 8 | .5 | .1 | 200 | 158 | OPS | 1.55 | X | | 90 AFT | | |
| 25 Nov | 0200 | 155 | 7.13 | 480 | 9.6 | 158 | 8 | .5 | .1 | 200 | 158 | OPS | 30.33 | X | | 90 AFT | | |
| 25 Nov | 0400 | 157 | 7.12 | 607 | 9.4 | 157 | 8 | .49 | .1 | 200 | 160 | OPS | 29.72 | X | | 90 AFT | | |
| 25 Nov | 1500 | 156.6 | 6.18 | 595 | 10.1 | 157 | 8 | .4 | .1 | 200 | 156 | OPS | 30.2 | X | | 90 AFT | | |
| 25 Nov | 1700 | 157.8 | 7.56 | 601 | 10.1 | 157 | 8 | .4 | .1 | 200 | 157 | OPS | | X | | 90 AFT | | |
| 25 Nov | 1800 | 151 | 8.37 | 580 | 10.1 | 154 | 6 | .45 | .15 | 280 | 153 | OPS | 29.15 | X | | 90 AFT | | |
| 25 Nov | 2200 | 151.3 | 6.56 | 597 | 10.2 | 156 | 7 | .38 | .15 | 200 | 156 | OPS | 30.20 | X | | 90 AFT | | |
| 25 Nov | 2400 | 148 | 5.46 | 577 | 10.3 | 156 | 7 | .38 | .15 | 200 | 150 | OPS | 30.19 | X | | 90 AFT | | |

TABLE 2z.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | | | | | MK V | | | | PORTABLE | | | | ATM CONTROL | | | | | | | | | | |
|--------|------|----------------|------|-----------------|------|----------------|------|------------------|----|----------------|-----|-----------------|--------|----------------|----|----------|----|----------------|----|-------------|----|------|----|-----------------|----|-------|-----|-------------|-----|--|
| | | O ₂ | | CO ₂ | | N ₂ | | H ₂ O | | O ₂ | | CO ₂ | | H ₂ | | FiR | | O ₂ | | LCC | | BARO | | CO ₂ | | BLEED | | VENTILATION | | |
| | | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | mm | in | in | mm | mm | CFH | COM | SEC | |
| 26 NOV | 0600 | 150.6 | 6.4 | 577 | 7.1 | 152 | 8 | 4 | 13 | 11 | 200 | 152 | UL OPS | 27.9 | | | | | | | | | | | | | | | | |
| 26 NOV | 0800 | 152.7 | 8.2 | 590 | 10 | 155 | 11 | 49 | 12 | 20 | 208 | 154 | UL OPS | 27.7 | | | | | | | | | | | | | | | | |
| 26 NOV | 1300 | 155.8 | 8.4 | 600 | 10.3 | 162 | 11 | 48 | 15 | 10 | 202 | 158 | UL OPS | 30.58 | | | | | | | | | | | | | | | | |
| 26 NOV | 1500 | 155.0 | 6.9 | 577 | 9.7 | | | | | | | 153 | UL OPS | 29.22 | | | | | | | | | | | | | | | | |
| 26 NOV | 2000 | 161.6 | 8.13 | 603 | 9.6 | 162 | 12 | 35 | 20 | 15 | 198 | 166 | UL OPS | 30.09 | | | | | | | | | | | | | | | | |
| 27 NOV | 0000 | 156.1 | 7.17 | 583 | 7.9 | 157 | 8 | 4 | 14 | 10 | 200 | 157 | UL OPS | 28.73 | | | | | | | | | | | | | | | | |
| 27 NOV | 0400 | 157.5 | 8.40 | 591 | 6.7 | | | | | | | 159 | UL OPS | 30.13 | | | | | | | | | | | | | | | | |
| 27 NOV | 0404 | 158.9 | 7.97 | 596 | 6.63 | 160 | 11.5 | 4 | 4 | 15 | 200 | 159 | UL OPS | 30.39 | | | | | | | | | | | | | | | | |
| 27 NOV | 0600 | 159.4 | 7.74 | 596 | 6.53 | 160 | 6 | 26 | 1 | 21 | 200 | 162 | UL OPS | 30.60 | | | | | | | | | | | | | | | | |
| 27 NOV | 0830 | 162 | 7.86 | 611 | 6.63 | | | | | | | 165 | UL OPS | 31.1 | | | | | | | | | | | | | | | | |
| 27 NOV | 1000 | 154 | 6.90 | 580 | 6.66 | 151 | 12 | 26 | 1 | 21 | 200 | 157 | UL OPS | 31.2 | | | | | | | | | | | | | | | | |
| 27 NOV | 1200 | 156 | 6.07 | 587 | 6.6 | 157 | 12 | 34 | 2 | 10 | 200 | 158 | UL OPS | 31.75 | | | | | | | | | | | | | | | | |
| 27 NOV | 1800 | 163.7 | 4.85 | 609 | 6.67 | 165 | 9 | 35 | 3 | 15 | 200 | 164 | UL OPS | 31.09 | | | | | | | | | | | | | | | | |
| 27 NOV | 2000 | 165.2 | 5.65 | 616 | 6.72 | | | | | | | 165 | UL OPS | 31.42 | | | | | | | | | | | | | | | | |
| 27 NOV | 2200 | 158.5 | 5.8 | 595 | 6.67 | 162 | 8 | 32 | 1 | 22 | 200 | 161 | UL OPS | 30.30 | | | | | | | | | | | | | | | | |
| 27 NOV | 0000 | 158.9 | 6.32 | 595 | 6.72 | 158 | 8 | 32 | 1 | 22 | 200 | 161 | UL OPS | 30.38 | | | | | | | | | | | | | | | | |

TABLE 2aa. - COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | MK V | | | | PORTABLE | | | ATM CONTROL | | | | |
|--------|------|----------------------|-----------------------|----------------------|------------------------|----------------------|-----------|----------------------|---------------------|------------------------|------------------------|----------------------|-------------|------------|----------------------------------|---------------------------------|--------------|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | F ₁₁ ppm | F ₁₂ ppm | O ₂ mm | LOC | BARO in | S _{CO₂} % | S _{O₂} % | BLEED CFH |
| 28 Nov | 0200 | 160.1 | 6.78 | 601 | 10.7 | 160 | 6 | 0.86 | 0.2 | 30 | 600 | UL OPS | 30.63 | X | N | 100 AFT | |
| 29 Nov | 0400 | 162.4 | 1.05 | 610 | 10.6 | | | | | | | UL OPS | 31.05 | X | X | 100 AFT | |
| 28 Nov | 0600 | 164.2 | 6.32 | 613 | 10.4 | | | | | | | UL OPS | 30.54 | X | X | 100 AFT | 90 |
| 28 Nov | 0800 | 166.2 | 3.38 | 624 | 10.9 | 167 | 6 | 0.2 | 1.1 | 10 | 169 | UL OPS | 31.74 | X | X | 100 AFT | |
| 28 Nov | 1200 | 162.8 | 2.65 | 606 | 10 | 165 | 16 | 0.2 | 1.3 | 10 | 164 | UL OPS | 30.82 | X | X | 100 AFT | |
| 28 Nov | 1400 | 163.6 | 5.51 | 613 | 10.4 | | | | | | | UL OPS | 31.21 | X | X | 100 AFT | |
| 28 Nov | 1600 | 164 | 7.61 | 621 | 10 | 165 | 18 | 0.49 | 1.3 | 10 | 164 | UL OPS | 31.39 | X | X | 100 AFT | |
| 28 Nov | 2000 | 150.1 | 3.79 | 563 | 9.8 | 152 | 11 | 0.0 | 0.3 | 10 | 152 | UL OPS | 28.59 | X | X | 100 AFT | |
| 29 Nov | 2200 | 152.0 | 6.02 | 575 | 10.6 | | | | | | | UL OPS | 30.48 | X | X | 100 AFT | |
| 29 Nov | 0100 | 146.4 | 7.45 | 565 | 9.9 | 147 | 8 | 0.48 | 1.1 | 17 | 160 | UL OPS | 30.54 | X | X | 100 AFT | |
| 29 Nov | 0400 | 148.4 | 7.54 | 569 | 10.7 | 150 | 8 | 0.50 | 1.1 | 15 | 148 | UL OPS | 28.94 | X | X | 100 AFT | |
| 29 Nov | 0534 | 149.5 | 6.54 | 569 | 10.3 | 152 | 11 | 0.49 | 1.1 | 9 | 147 | UL OPS | 29.12 | X | X | 100 AFT | |
| 29 Nov | 0600 | 150.4 | 5.64 | 579 | 10.7 | | | | | | | UL OPS | 29.32 | X | X | 100 AFT | |
| 29 Nov | 1900 | 152.8 | 6.05 | 588 | 10.6 | 155 | 11 | 0.4 | 1.1 | | 169 | UL OPS | 29.79 | X | X | 100 AFT | |
| 29 Nov | 1700 | 153.3 | 6.64 | 591 | 10.9 | | | | | | | UL OPS | 29.94 | X | X | 100 AFT | 1009 1305 |
| 29 Nov | 1400 | 157.7 | 3.24 | 592 | 10.0 | 160 | 15 | 0.2 | 0.35 | 20 | 26 | UL OPS | 30.08 | X | X | 0 | |

TABLE 2bb.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | MK V | | | | PORTABLE | | | ATM CONTROL | | |
|--------|------|-----------------|----------------|------------------|----------------|------|-----------------|----------------|-----|----------------|-------|----------|----------------------|-------|-------------|-----|--|
| | | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | FIR | O ₂ | LOC | BARO | Sp. H ₂ O | BLEED | VENTILATION | | |
| | | mm | mm | mm | mm | ppm | % | % | ppm | mm | | in | | CFH | COM | SEC | |
| 29 Nov | 1600 | 157.5 | 573 | 599 | 7.5 | | | | 160 | UL OPS | 30.32 | | | | | | |
| 29 Nov | 1800 | 156.9 | 7.0 | 604 | 7.9 | | | | 158 | UL OPS | 30.00 | X | | 90 | 2150 | | |
| 29 Nov | 2000 | 157.6 | 7.7 | 612 | 10.8 | 0.49 | 25 | 184 | 159 | UL OPS | 30.72 | X | | 60 | | | |
| 30 Nov | 0400 | 158.4 | 5.1 | 601 | 10 | 0.38 | 1 | 30 | 157 | UL OPS | 30.61 | X | | 70 | | | |
| 30 Nov | 0800 | 158.4 | 7.48 | 610 | 10.5 | | | | 160 | UL OPS | 30.76 | X | | 20 | | | |
| 30 Nov | 0900 | 160.4 | 8.45 | 619 | 10.4 | 0.45 | 17 | 200 | 161 | UL OPS | 31.15 | X | | 70 | | | |
| 30 Nov | 0600 | 158.9 | 9.2 | 614 | 10.2 | | | | 159 | UL OPS | 31.44 | X | X | 0 | | | |
| 30 Nov | 0800 | 155.3 | 7.12 | 607 | 10.0 | 0.8 | 3 | 10 | 156 | UL OPS | 30.70 | X | X | 0 | | | |
| 30 Nov | 1100 | 153.3 | 3.62 | 583 | 10.4 | | | | 154 | UL OPS | 29.56 | X | | 0 | 0955 | | |
| 30 Nov | 1200 | 154.4 | 6.16 | 597 | 10.0 | 0.4 | 3 | 10 | 162 | UL OPS | 30.24 | X | | 0 | | | |
| 30 Nov | 1400 | 151.4 | 1.92 | 593 | 10.0 | | | | 159 | UL OPS | 29.99 | X | | 60 | | | |
| 30 Nov | 1600 | 152.1 | 9.76 | 578 | 10.8 | 0.5 | 15 | 5 | 160 | UL OPS | 30.31 | X | | 40 | | | |
| 30 Nov | 1800 | 152.4 | 9.77 | 603 | 10.7 | | | | 154 | UL OPS | 30.49 | X | | 40 | | | |
| 30 Nov | 2000 | 151.7 | 2.55 | 567 | 11 | 0.2 | 1 | 22 | 154 | UL OPS | 28.86 | X | X | | | | |
| 1 DEC | 0030 | 152.1 | 7.10 | 584 | 10.2 | | | | 154 | UL OPS | 29.72 | X | X | 120 | | | |
| 1 DEC | 0230 | 153.8 | 8.30 | 593 | 11.4 | | | | 154 | UL OPS | 30.70 | | | 120 | | | |

TABLE 2dd.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | MK V | | | | PORTABLE | | | ATM CONTROL | | |
|-------|------|-----------------|----------------|------------------|----------------|-----|-----------------|----------------|------|-----------------|----------------|----------|------|----------------------|-------------|-------------|-----|
| | | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | Fill | Fi ₂ | O ₂ | LOC | BARO | CO ₂ | BLEED | VENTILATION | |
| | | mm | mm | mm | mm | ppm | % | % | ppm | ppm | mm | | in | H ₂ | CFH | COM | SEC |
| 2 DEC | 1400 | 157.3 | 596 | 11.0 | 157 | 8 | 0.2 | 15 | 85 | 158 | UL OPS | 30.29 | X | ∅ | 1558 | 1452 | |
| 2 DEC | 1800 | 154.3 | 606 | 10.6 | | | | | | 157 | UL OPS | | X | ∅ | | | M |
| 2 DEC | 2000 | 154.7 | 616 | 11.7 | 157 | 3 | 0.4 | 2 | 800 | 157 | UL OPS | 31.3 | X | ∅ | | | M |
| 2 DEC | 2200 | 149.3 | 594 | 11.2 | | | | | | 151 | UL OPS | 30.09 | X | 100% 100% 100% | | | M |
| 2 DEC | 2342 | 151.4 | 604 | 10.9 | 157 | 2.5 | 0.38 | 1 | 35 | 153 | UL OPS | 30.49 | | | 2340 | 0026 | |
| 3 DEC | 0025 | 160.5 | 612 | 10.6 | 160 | 10 | 0.19 | 0.11 | 9 | 162 | UL OPS | 31.08 | X | | | | |
| 3 DEC | 0446 | 161.4 | 628 | 10.2 | 162 | 10 | 0.33 | 0.1 | 7 | 161 | UL OPS | 31.78 | X | X | | | |
| 3 DEC | 0630 | 161.6 | 637 | 10.7 | 162 | 6 | 0.0 | 0.15 | 10 | 162 | UL OPS | 32.10 | X | X | | | |
| 3 DEC | 0800 | 152.3 | 606 | 10.7 | | | | | | 153 | UL OPS | 30.60 | X | | | | |
| 3 DEC | 1000 | 158.5 | 638 | 12.5 | 157 | 3 | 0.0 | 0.25 | 10 | 157 | UL OPS | 32.00 | X | X | | | |
| 3 DEC | 1200 | 151.3 | 614 | 12.2 | 152 | 2 | 0.35 | 0.25 | 10 | 151 | UL OPS | 30.91 | X | 100% 100% | 1330 | 1410 | |
| 3 DEC | 1400 | 151.5 | 600 | 10.5 | | | | | | 160 | UL OPS | 31.02 | X | ∅ | | | |
| 3 DEC | 1600 | 154.8 | 612 | 11.1 | 158 | 8 | 0.42 | 0.2 | 5 | 157 | UL OPS | 31.02 | X | ∅ | | | M |
| 3 DEC | 1800 | 156.8 | 641 | 11.0 | 157 | 5 | 0.4 | 0.1 | 6 | 158 | UL OPS | 31.41 | X | ∅ | | | |
| 3 DEC | 2000 | 156.1 | 604 | 11.1 | | | | | | 157 | UL OPS | 30.52 | X | X | | | |
| 3 DEC | 2200 | 148.2 | 600 | 11.0 | 150 | 5 | 0.38 | 0.12 | 2 | 150 | UL OPS | 30.15 | X | X | X | | |

TABLE 2ee. - COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | | | | | | PORTABLE | | | | ATM CONTROL | | | |
|-------|------|----------------|------|-----------------|----------------|------------------|----------------|-----------------|-----------------|----------------|----------------|----------|----------------|--------|-------|-----------------|-------|-------------|------|
| | | O ₂ | | CO ₂ | N ₂ | H ₂ O | O ₂ | CO | CO ₂ | H ₂ | FI | FR | O ₂ | LOC | BARO | CO ₂ | BLEED | VENTILATION | |
| | | mm | mm | mm | mm | mm | ppm | % | % | ppm | ppm | mm | ops | in | in | CFH | COM | SEC | |
| 4 DEC | 0030 | 160.7 | 2.76 | 610 | 11.0 | | | | | | | | 163 | UL OPS | 30.00 | | | 2320 | 0015 |
| 4 DEC | 0230 | 152.3 | 4.88 | 592 | 11.4 | 153 | 9 | 0.2 | 1 | 10 | 114 | | 154 | UL OPS | 30.00 | | | | |
| 4 DEC | | 153.2 | 6.4 | 602 | 10.7 | 153 | 9 | 0.2 | 1 | 10 | 114 | | | | | | | | |
| 4 DEC | 0400 | 153.7 | 6.4 | 602 | 10.7 | | | | | | | | 154 | UL OPS | 30.97 | | | | |
| 4 DEC | 0600 | 153.7 | 6.74 | 610 | 10.3 | | | | | | | | 155 | UL OPS | 30.97 | | | | |
| 4 DEC | 0800 | 157.8 | 6.97 | 597 | 10.8 | 155 | 0 | 0.35 | 2 | 14 | 000 | | 152 | UL OPS | 30.05 | | | 0800 | 0340 |
| 4 DEC | 1000 | 152.7 | 4.77 | 588 | 10.8 | | | | | | | | 155 | UL OPS | 29.8 | | | | |
| 4 DEC | 1235 | 152.9 | 7.33 | 612 | 11.7 | 155 | 2.5 | 0.4 | 1 | 3 | 000 | | 154 | UL OPS | 31.01 | | | | |
| 4 DEC | 1600 | 151.0 | 8.16 | 617 | 11.6 | 157 | 2.5 | 0.32 | 2 | 20 | 000 | | 151 | UL OPS | 30.82 | | | | |
| 4 DEC | 1900 | 149.6 | 6.68 | 600 | 11.7 | | | | | | | | 152 | UL OPS | 30.80 | | | | |
| 4 DEC | 2200 | 155.5 | 4.10 | 598 | 10.7 | | | | | | | | 154 | UL OPS | 30.16 | | | 1848 | 2005 |
| 5 DEC | 0030 | 147.0 | 4.00 | 576 | 9.6 | 148 | 7 | 0.3 | 3 | 135 | 2 | 000 | 152 | UL OPS | 23.8 | | | | |
| 5 DEC | 0200 | 148.9 | 4.09 | 578 | 9.7 | | | | | | | | 147 | UL OPS | 29.17 | | | | |
| 5 DEC | 0500 | 150.3 | 4.18 | 584 | 9.8 | 148 | 6 | 0.35 | 4 | 2 | 000 | | 147 | UL OPS | 29.96 | | | | |
| 5 DEC | 0600 | 150 | 4.16 | 585 | 9.5 | 148 | 8 | 0.38 | 3 | 3 | 000 | | 148 | UL OPS | 28.78 | | | | |
| 5 DEC | 0745 | 143.0 | 4.17 | 559 | 9.9 | 140 | 2.5 | 0.4 | 2 | 3 | 000 | | 140 | UL OPS | 26.20 | | | | |

TABLE 2ff.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | MK.V | | | | | PORTABLE | | | ATM CONTROL | | | |
|-------|------|----------------------|-----------------------|----------------------|------------------------|----------------------|-----------|----------------------|---------------------|----------|------------|------------|----------------------|-------|-------------|----------------------|-------------|--------------------|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | Ha mm | Fil ppm | Fil ppm | O ₂ mm | LOC | BARO in | CO ₂ % | BLEE CFH | VENTILATION SEC |
| 5 Dec | | 152 | 470 | 513 | 10.5 | 153 | 14 | .19 | .32 | 9 | 00 | 157 | OPS | 27.95 | X | 0 | 0952 | 1034 |
| 5 Dec | 1300 | 153.7 | 496 | 600 | 10.7 | 153 | 6 | .38 | .3 | 9 | 00 | 153 | OPS | 30.31 | X | | | |
| 5 Dec | 1500 | 155.6 | 5.19 | 608 | 10.9 | | | | | | | 154 | OPS | 30.76 | | | | |
| 5 Dec | 1700 | 151.7 | 5.21 | 594 | 10.9 | 148 | 6 | 000 | .6 | 10 | 00 | 149 | OPS | 30.06 | | 100 | | |
| 5 Dec | 1800 | 152.3 | 5.28 | 596 | 10.9 | | | | | | | 150 | OPS | 30.10 | X | 130 PD | | M |
| 5 Dec | 1734 | 153.5 | 5.43 | 601 | 11.0 | 151 | 0 | .32 | .29 | 9 | PH | 152 | OPS | 30.34 | X | 130 PD | | M |
| 5 Dec | 2000 | 156.6 | 5.77 | 614 | 11.1 | | | | | | | 154 | OPS | 30.80 | X | 100 PD | | |
| 6 Dec | 0000 | 147.9 | 5.53 | 561 | 10.5 | 145 | 2.5 | .05 | .11 | 2 | 00 | 147 | OPS | 29.37 | X | 110 | | |
| 6 Dec | 0345 | 153.3 | 5.75 | 606 | 10.9 | 152 | 5 | .05 | .33 | 4 | 00 | 155 | OPS | 30.60 | X | 0 | NOTE CO | |
| 6 Dec | 0700 | 158.5 | 6.17 | 616 | 10.7 | 153 | 6 | 00 | .75 | 11 | 00 | 152 | OPS | 31.02 | X | 0 | | |
| 6 Dec | 0900 | 161 | 6.12 | 598 | 10.5 | | | | | | | 151 | OPS | 30.70 | | 0 | | |
| 6 Dec | 1100 | 151.1 | 6.23 | 602 | 10.5 | 153 | 15 | 000 | .7 | 11 | 00 | 159 | OPS | 30.32 | X | 0 | 0955 | 1133 |
| 6 Dec | 1200 | 152.6 | 6.39 | 608 | 10.6 | 155 | 12 | .25 | .25 | 4 | PH | 159 | OPS | 30.40 | X | 0 | | |
| 6 Dec | 1400 | 152.6 | 6.39 | 608 | 10.6 | | | | | | | 158 | OPS | 29.40 | X | 0 | | |
| 6 Dec | 1600 | 152.6 | 6.39 | 608 | 10.6 | 147 | 0 | .3 | .2 | 10 | PH | 159 | OPS | 29.57 | X | 0 | | |
| 6 Dec | 1600 | 0 | 0 | 0 | 0 | 148 | 5.5 | .39 | .1 | 9 | PH | 153 | OPS | 29.72 | X | 100 | AFT | |

TABLE 2gg.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

AT 101 Johnson discovered that the mass spectrometer had zero flow rate indicated. A blow of size was discovered in the supply to the sample pump. The placement fuses blew with switch (pump) placed in On. Mass Spec. considered OOC; however, appears to be reading correct atmosphere.

TABLE 2hh.- COMPARISON DATA RECORDED BY U.S.S. HAMMERHEAD'S CREW

| DATE | TIME | MASS SPEC | | | | | MK. V | | | | | DOO | | | | PORTABLE | | | ATM CONTROL | | | |
|-------|------|----------------------|-----------------------|----------------------|------------------------|----------------------|-----------|----------------------|---------------------|-----|------------------------|----------------------|-----|------|----------------------|--------------|--------------------|-----|-------------|--|--|--|
| | | O ₂ mm | CO ₂ mm | N ₂ mm | H ₂ O mm | O ₂ mm | CO ppm | CO ₂ % | H ₂ % | FI | FI ₂ ppm | O ₂ mm | LOC | BANO | CO ₂ % | BLEED CFH | VENTILATION COM | SEC | | | | |
| 6 Dec | 2200 | | | | 8.0 | 149 | 8 | 32.1 | 70 | 60 | 44 | UL | in | | X | 120 | | | | | | |
| 7 Dec | 0000 | | | | | 096 | 0 | 2.7 | | 147 | JU | | | X | 120 | | | | | | | |

PERFORMANCE SUMMARY

Final system calibration prior to deployment of the Undersea Atmospheric Sensor was achieved at 1001 on 6 October 1970. A specially prepared gas mixture from Matheson Scientific, consisting of 74.8% nitrogen, 20.11% oxygen, 2.13% carbon dioxide and 3.0% argon, was utilized both during calibrations for the five-day sea trial and the final sea deployment. This standard mix now resides at NASA/Langley Research Center. Figure 1 is a graphical presentation of the system calibrations before and after the 62-day deployment of the Undersea Atmospheric Sensor. Also, as a correlation to the system calibration, standard ships air data before and after the deployment is also presented. Water vapor calibration, originally achieved utilizing relative humidity and temperature at the sample point, has remained stable during the sea deployment.

The variations in torr between theoretical output and system outputs before and after sea deployment are as follows:

| Calibration Before Sea Deployment | Gas Constituent | Calibration After Sea Deployment |
|--------------------------------------|--------------------|-------------------------------------|
| 0 | N ₂ | +5.22 |
| +0.5 | O ₂ | +0.98 |
| +0.04 | CO ₂ | -0.12 |
| +0.65 | A _r | +0.53 |
| 0 | H ₂ O | 0 |
| +1.24 | P _{TOTAL} | +6.92 |

Comparing the instrument calibration prior to the deployment with the calibration verification after deployment, it can be seen that all parameter, except argon, have remained within 1.0% of their theoretical values; and argon, a low-level constituent, has shifted less than 0.6% from the original calibration during the 62-day test.

Figures 3a through 3e are graphical presentations of the carbon dioxide output from the Undersea Atmospheric Sensor during sea deployment. Absolute carbon dioxide outputs, rather than delta values, are plotted in an effort to draw correlation between shipboard activities and levels of carbon dioxide experienced. In these figures it is possible to correlate the reduction in CO₂ scrubbers, ventilation, snorkel, and various air bleeds.

It has been reported that the operating efficiency of shipboard CO₂ scrubbers improved as the CO₂ level increases above a nominal 1%, such that the higher the carbon dioxide level the more efficient the scrubbers. Final system calibration, however, showed that CO₂ was within 1% of the theoretical value. This results in a ± 0.11 torr error out of a nominal 7.6 torr control level. Clearly this level of uncertainty is of little or no concern relating to a decision to start CO₂ scrubbers or not.

Figures 4a through 4e are comparisons of the Mark V carbon dioxide outputs with the Perkin-Elmer Undersea Atmospheric Sensor taken as a reference. A $\pm 5.7\%$ full-scale error band or ± 0.95 torr based on 16.74 torr full scale has been shown for purposes of data evaluation. This error band relates to previous theoretical worst case analysis performed on the mass spectrometer attributed to system performance characteristics occurring simultaneously.

Figures 5a through 5e represent the summation of the partial pressure outputs during sea deployment of the Undersea Atmospheric Sensor. Since nitrogen, oxygen, carbon dioxide and water vapor were the four atmospheric constituents displayed and argon constitutes approximately 1% of the nominal atmosphere, 7.4 torr has been added to the recorded data outputs for the contribution of argon, in order to more closely approximate the ambient pressure. From 0200, 25 October through 1000, 5 November, and from 0200, 27 November through 2400, 27 November, the Test Mode Select switch was left in a position other than H₂O output. During these periods 10.0 torr, a nominal water vapor output level, has been added to the summation of the data outputs recorded. The resulting summation of the partial pressures is assumed to be within $\pm 0.25\%$ of the actual data outputs during the periods when the Test Mode Select switch was in the wrong position. During the remainder of the data it is expected that the summation of the partial pressures is within $\pm 0.1\%$ of the actual summation if argon would have been recorded.

In Figure 5a, from 13 October through 23 October, the summation of the partial pressures increased approximately 9.5 torr from a nominal level which was 6.5 torr lower than the barometric pressure. This positive system drift is basically attributed to system thermal stability. Factors contributing to long term system instability are thermal coefficients of system electronic components and the pressure transducer utilized. The Undersea Atmospheric Sensor experienced an ambient temperature change from 54°F to 28°F, or a delta change of 26°F during sea deployment. Relating this thermal variation to system drift, a 0.37 torr/°F system thermal shift was observed. The Undersea Atmospheric Sensor was originally developed for use in a relatively constant ambient atmosphere, and thus rigorous consideration of thermal variation was not made. Future applications would be addressed to satisfying more stringent ambient conditions.

A $\pm 1\%$, 7.6 torr error band has been added for purposes of system evaluation, to Figure 5a through 5e, about a nominal +2.5 torr assumed thermally stable system partial pressure summation. It can be seen that the majority of the summation of the Undersea Atmospheric Sensor outputs fall well within this $\pm 1\%$ error band.

Major perturbations below the nominal partial pressure summation are attributed to possible reduction of inlet leak conductance. Since Engineering data was not recorded during the periods when the summation of the partial pressures were significantly below the barometric reading, it is impossible to positively identify the source of the nonconformance. If, however, one were to sum the partial pressures recorded in Tables 1a, 1b, and 1c on 17 November at 1400 hours, one would see that the summation of the partial pressures was equal to 736 torr, while the pressure transducer output was 771 torr - a variation of -35 torr. As the inlet valve conductance is reduced, the anode current increases in an attempt to maintain stable closed-loop operation. Closed-loop operation of the Undersea Atmospheric Sensor has been limited to 35.0 microamperes anode current. If the system demands more anode current - due to change in inlet leak conductance - than can be supplied by the preset limitation of 35 microamperes anode current, the sum of the partials will start to deviate from the barometric pressure. In all instances this will be a negative deviation from the barometric pressure.

The reason for inlet leak conductance variation is not obvious from the data available. A thermal transient in the inlet valve, or a vibration transient, seems to be the most reasonable explanation. However, a discrete contaminant or particulant water in the inlet leak area might also provide similar behavior. The data does, however, indicate that inlet conductance does recover to the original value within two hours.

Major perturbations above the nominal partial pressure summation are attributed in most cases to barometric pressure reading error. It should be noted that in many instances when the summation of the partial pressures is higher than the barometric pressure, the deviation between the two is approximately 1 inch of mercury. While this does not appear indicative of all the barometric pressure readings, it was determined that accurate reading of the available barometer does require more familiarity than should normally be required. This conclusion was reached during predeployment calibration when members of the ship's crew were asked for barometric pressure readings and, when once taken, were found to be in error by a significant amount.

A comparison of the major perturbations in the summation of partial pressures, Figures 5a through 5e, and the oxygen partial pressure output, Figures 2a through 2e, reveals no correlation. If the major perturbations in the summation of partial pressures are accepted for what they are, it is found that the maximum error in Figure 5 is approximately +6% of 760 torr. This same error related to the oxygen output or carbon dioxide output would produce approximately +9.6 torr error in oxygen based on 160 torr nominal level and +0.67 torr error in carbon dioxide based on 11.1 torr maximum level. Negative perturbations can be eliminated in the future by thermal and vibrational isolation of the inlet leak valve, additional sample filtration and higher anode current set point. Positive errors can be eliminated through additional care in observing and recording barometric pressure readings.

SAMPLE TRANSPORT PUMP FAILURE

On 6 December, between 0900 hours and 1100 hours, the sample transport pump, a commercially procured auxiliary pump utilized to move the air sample through the sample line, ceased functioning. This failure resulted in the absence of new gas sample at the inlet to the mass spectrometer. The transport pump has been replaced and a failure analysis performed. The results of the failure analysis are shown in Figures 6 and 7. The Gast pump, Model 0531-102347, has four carbon vanes inserted into a steel rotor hub which, in turn, is mounted in a cast housing (with eccentric cavity). These carbon vanes are designed such, that, as the rotor rotates, the vanes slide in and out of the rotor hub following the contour of the eccentric housing and moving a gas sample through the system. The mechanical fit between the rotor and vanes is such that abrasion of the vane results. Continuous operation of this pump eventually produces failure of the vane which, when fractured, tends to jam the motion of the rotor and subsequently causes the pump fuse to blow. Future systems utilizing a similar sample transport scheme must provide access to the rotor area of the pump and adequate replacement vane spares to insure considerably longer system operation.

CONCLUSIONS

The Undersea Atmosphere Sensor System, described in Appendix 3, Operation and Maintenance Manual, Undersea Atmospheric Sensor, has proven to be a reliable, automatic and accurate tool for monitoring the atmospheric constituents of nitrogen, oxygen, carbon dioxide, water vapor and argon under in-use undersea environment conditions.

Virtually maintenance-free operation was provided during 62 continuous days of sea deployment. Calibrations were performed at the beginning and end of sea deployment, the results of which indicate unusually stable long term operation. During the period of sea deployment, the readings provided by the Undersea Atmospheric Sensor were in close agreement with other atmospheric analyzing equipment, both fixed and portable. By observing atmospheric trends, carbon dioxide scrubber operation, ventilation and air bleed modes, the operational reliability of the Undersea Atmospheric Sensor can be verified. This close agreement with other on-board equipment and the stability of system calibration indicate that, with the addition of other monitoring functions - not presently included in the Undersea Atmospheric Sensor, - the requirement for total atmospheric constituent monitoring can be provided in a compact, reliable, automatic configuration.

THIS PAGE INTENTIONALLY LEFT BLANK

APPENDIX 1

RESULTS OF FIVE DAY SEA TRIAL
ON THE UNDERSEA
ATMOSPHERIC SENSOR
UPTITE

October 1970

Perkin-Elmer SPO 30006
NASA Contract Number NAS1-9469

Prepared for
NATIONAL AERONAUTICS AND SPACE ADMINISTRATION
LANGLEY RESEARCH CENTER
Langley Station
Hampton, Virginia 23365

RESULTS OF FIVE DAY SEA TRIAL
ON THE UNDERSEA
ATMOSPHERIC SENSOR
UPTITE

October 1970

Approved By: J. P. Stuart Date 10-19-70
J. P. Stuart, Project Engineer

Approved By: M. R. Ruecker Date 10/19/70
M. R. Ruecker,
Manager Space Physics

Approved By: B. F. Bicksler Date 10/19/70
B. F. Bicksler, Project Manager

Perkin-Elmer SPO 30006
NASA Contract Number NAS1-9469

Prepared for
NATIONAL AERONAUTICS AND SPACE ADMINISTRATION
LANGLEY RESEARCH CENTER
Langley Station
Hampton, Virginia 23365

RESULTS OF FIVE DAY SEA TRIAL
ON THE UNDERSEA ATMOSPHERIC SENSOR

UPTITE

INTRODUCTION

Under the general scope of Contract NAS1-9469, the Perkin-Elmer Corporation, Aerospace Division, undertook as Contract Item II, the effort necessary to modify a Government-Furnished Two Gas Atmospheric Sensor System for continuous use as an Undersea Atmospheric Sensor System. The intent of the experiment was to monitor the atmospheric constituents of nitrogen, oxygen, carbon dioxide, water vapor, and argon over an extended period of time, in an undersea cabin atmosphere. The operation of the Undersea Atmospheric Sensor was to be essentially automatic with visual display of the nitrogen, oxygen, carbon dioxide, and water vapor atmosphere constituents. Ease of maintenance and troubleshooting and reliable long-term operation of the instrument were the key requirements.

On 25 September 1970, the Undersea Atmospheric Sensor System, UPTITE, was delivered, placed in operation, and calibrated for a preliminary five day sea trial. Table 1-1a, 1-1b, and 1-1c are copies of the engineering data recorded during the five day sea trial by a member of the Naval Research Laboratory staff. This data was taken periodically for purposes of monitoring the system's internal performance by Perkin-Elmer.

Table 1-2a, 1-2b, 1-2c are copies of data forms maintained during the five day sea trial by the same NRL staff member that recorded the data on Table 1-1a through 1-1c. The data in Table 1-2a through 1-2c would normally be recorded by members of the crew during each watch period.

As the instrument was initially calibrated on a special Perkin-Elmer gas mixture of 74.8% nitrogen, 20.11% oxygen, 2.13% carbon dioxide and 3.0% argon, each readout parameter was within 2.0 torr of the theoretical value. The summation of the partial pressures was within 5.0 torr of the barometric pressure. The performance of the system on standard air was such that the readout parameters were within 1.5 torr of theoretical values, and the summation of the partial pressures was within 2.0 torr of the barometric pressure. Throughout the five day sea trial the system performed steadily, retaining the level of calibration initially observed. Calibration verification performance at the conclusion of the five day sea trial showed that: 1) nitrogen was within 2.84 torr of the 563.17 torr theoretical value; 2) oxygen was within 2.4 torr of the 151.4 torr theoretical value; 3) carbon dioxide was within 0.14 torr of the 16.04 torr theoretical value; 4) argon was within 0.62 torr of the 22.58 torr theoretical value; and 5) the summation of the partial pressures was within 0.82% of the barometric pressure.

PERKIN-ELMER
AEROSPACE DIVISION

TABLE 1-2a
MASS SPECTROMETER SENSOR DATA FORM #1

| DATE | TIME | N ₂ | O ₂ | CO ₂ | H ₂ O | ΣPP | REMARKS | DATA READ BY |
|------|--------------|----------------|----------------|-----------------|------------------|-----------------|---|--------------|
| 9-26 | 2200 | 579. 581. | 155.5 156.0 | 0.71 1.19 | 11.9 12.6 | | fluctuating (35% RH) 763 torr | RSD |
| 9-26 | 2345 | 572. | 158.8 158.9 | 0.92 0.95 | 12.8 12.9 | | " " " | RSD |
| 9-27 | 0045 | 572. | 158.7 158.8 | 0.97 1.03 | 12.8 13.0 | | " " " | RSD |
| 9-27 | 0220 4420 | 591. 592. | 158.7 158.8 | 0.94 1.00 | 12.7 12.8 | | " (34% RH) 762 torr | RSD |
| 9-27 | 0430 | 591. | 158.6 158.7 | 0.43 0.76 | 12.0 12.1 | | " (33% RH) 761 torr | RSD |
| 9-27 | 0630 | 591. | 158.6 158.5 | 0.89 0.91 | 11.7 11.8 | | " (32 1/2% RH) 760 torr | RSD |
| 9-27 | 0800 | 590. | 158.4 158.5 | 0.65 0.89 | 11.8 11.9 | | " (3 1/2% RH) 760 torr | RSD |
| 9-27 | 0927 | 586 | 157.0 | 1.03 | 12.5 | -3.57 756.53 | 81% (33% RH) 760 Torr | RSD |
| 9-27 | 1300 | 581 | 157.7 | 1.15 | 10.1 | 749.95 | 79°F | RSD |
| 9-27 | 1450 | 582 | 157.9 | 1.20 | 10.2 | 751.30 | 80°F ^{Bar} Sample Inlet Barom 29.9" | RSD |
| 9-27 | 1700 | 583 | 158.6 | 0.99 | 9.7 | 752.29 | 80°F Samp. Inlet Barom 29.98 | RSD |
| 9-27 | 2103 | 585 | 158.5 | 1.32 | 9.9 | 754.72 | 78°F Bar. 30.04 | RSD |
| 9-27 | 2325 | 585 | 158.3 | 1.58 | 9.5 | 754.38 | 76°F Bar. 30.04 | RSD |
| 9-28 | 0105 | 585 | 158.3 | 1.55 | 9.9 | 754.75 | 76°F Bar. 30.05 | RSD |
| 9-28 | 0200 | 585 | 158.3 | 1.49 | 9.8 | 754.59 | 76°F Bar. 30.05 | RSD |
| 9-28 | 0838 | 585 | 158.4 | 1.43 | 10.2 | 755.03 | 76°F Bar. 30.07 | RSD |
| 9-28 | 0935 | 585 | 158.7 | 1.28 | 10.3 | 755.28 | 80°F Surface ventilate note lower CO ₂ | RSD |
| 9-28 | 1035 | 586 | 158.8 | 1.21 | 10.3 | 756.31 | 80°F Bar. 30.07 | RSD |
| 9-28 | 1155 | 585 | 158.7 | 1.16 | 10.2 | 755.06 | 79°F Bar. 30.07 | RSD |
| 9-28 | 1255 | 584 | 158.4 | 1.22 | 10.1 | 753.23 | 78°F Moderate pitching + rollings. (readings steady) 7600 ft. | RSD |

TABLE 1-2b
MASS SPECTROMETER SENSOR DATA FORM #1

PERKIN-ELMER
AEROSPACE DIVISION

| DATE | TIME | N ₂ | O ₂ | CO ₂ | H ₂ O | ΣPP | REMARKS | DATA READ BY |
|------|------|----------------|----------------|-----------------|------------------|--------|--|--------------|
| 9-28 | 1410 | 584 | 158.3 | 1.15 | 10.0 | 753.45 | 29.87 Bar 742.34 Torr 761.24 78°F | RSD |
| 9-28 | 1800 | 584 | 158.4 | 1.15 | 9.7 | 753.23 | 29.94 Bar - 761.75 Torr 78°F | RSD |
| 9-28 | 1900 | 580 | 156.9 | 1.51 | 9.7 | 748.11 | 29.77 Bar. 756.16 Submerged 1848 77°F On recirculate air 1823 | RSD |
| 9-28 | 2000 | 601 | 161.0 | 2.84 | 9.5 | 774.34 | 30.77 Bar 782.07 Torr. 76°F Surfaced 1923 1910 | RSD |
| 9-28 | 2400 | 583 | 152.5 | 5.48 | 8.8 | 749.78 | 29.76 Bar 755.90 76°F Resubmerged 2030 | RSD |
| 9-28 | 0118 | 587 | 152.2 | 6.62 | 8.9 | 754.72 | 29.98 Bar 761.49 76°F | RSD |
| 9-28 | 0330 | 593 | 151.8 | 8.07 | 8.6 | | 30.20 75°F | RSD |
| 9-28 | 0430 | 596 | 151.8 | 8.63 | 8.6 | | 30.34 75°F | RSD |
| 9-28 | 0530 | 591 | 150.1 | 9.05 | 8.6 | 758.75 | 30.09-764.89 Torr 75°F | RSD |
| 9-28 | 0730 | 577 | 144.7 | 10.01 | 9.8 | | 29.40- 80°F (0720) Start CO ₂ scrubber | RSD |
| 9-28 | 1125 | 592 | 144.8 | 8.80 | 11.3 | 756.93 | 30.02-762.51 82°F | RSD |
| 9-29 | 1415 | 596 | 145.2 | 6.66 | 10.1 | | 30.12- 84°F opened hatch approx. 1300 | RSD |
| 9-29 | 1530 | 586 | 145.0 | 6.64 | 9.4 | | 30.03 81°F surface ventilate. (1525) | RSD |
| 9-29 | 1705 | 591 | 154.4 | 2.93 | 8.9 | | 30.12- 80°F | RSD |
| 9-29 | 2156 | 603 | 158.1 | 4.68 | 8.8 | | 30.78 76°F recirculate air 1923 DUE 1935 | RSD |
| 9-30 | 0237 | 591 | 152.4 | 8.04 | 9.0 | | 30.16 77°F | RSD |
| 9-30 | 0550 | 573 | 146.2 | 8.74 | 9.5 | | 29.22- 79°F | RSD |
| 9-30 | 0948 | 589 | 152.6 | 6.08 | 10.0 | | 30.88 83°F | RSD |
| 9-30 | 1145 | 593 | 149.9 | 8.58 | 9.8 | | 30.18 80°F | RSD |
| 9-30 | 0015 | 595 | 149.6 | 7.13 | 9.5 | | 30.33 80°F | RSD |

F-351

PERFORMANCE SUMMARY

Final system calibration was achieved just prior to 0927, 27 September 1970 on a Perkin-Elmer supplied calibration gas sample containing 74.8% nitrogen, 20.11% oxygen, 2.13% carbon dioxide and 3.0% argon. Calibration of the water vapor output was achieved using the existing relative humidity and temperature at the sample point. The system outputs were as follows in torr (MM Hg).

| | |
|--------------------|-------|
| N ₂ | 565 |
| O ₂ | 153.5 |
| CO ₂ | 16.38 |
| Ar | 23.3 |
| H ₂ O | 5.5 |
| P _{TOTAL} | 763.7 |

Theoretical outputs in torr are as follows:

| | |
|------------------|-------|
| N ₂ | 563.6 |
| O ₂ | 151.5 |
| CO ₂ | 16.05 |
| Ar | 22.6 |
| H ₂ O | 5.5 |
| P _{BAR} | 759 |

The variations between theoretical output and the instrument outputs (error) are as follows:

| | |
|------------------------|------------------------|
| N ₂ + 1.4 | Torr out of 563.6 torr |
| O ₂ + 2.0 | Torr out of 151.5 torr |
| CO ₂ + 0.33 | Torr out of 16.05 torr |
| Ar + 0.70 | Torr out of 22.6 torr |
| H ₂ O 0.00 | Torr -- |
| P _{TOTAL} 4.7 | Torr out of 759 torr |

Figure 8 is a graphical representation of the data presented in Tables 1-2a, 1-2b, and 1-2c. For clarity, it should be noted that on 30 September 1970, shortly after 1145 and prior to 1334 hours, a line of data was taken with a time of 0015 hours. It has been assumed that the time was actually 1215 hours. It should also be noted that at 1143, 2 October 1970, a line of data was taken which would indicate that 1584.4 torr of oxygen was present. Since the digital voltmeter can physically display only 3 1/2 digits, 199.9 being the largest, it has been assumed that the oxygen data was entered on the log sheet incorrectly.

The first and last sets of data presented on the graph, Figure 8, were taken by Perkin-Elmer personnel and are circled for identification.

At 1029 on 3 October 1970, the Undersea Atmospheric Sensor System was recalibrated on Perkin-Elmer calibration gas. The calibration sample was the same that was used previously; 74.8% nitrogen, 20.11% oxygen, 2.13% carbon dioxide, and 3.0% argon. The system outputs were as follows:

| | |
|--------------------|--------|
| N ₂ | 566 |
| O ₂ | 153.8 |
| CO ₂ | 16.18 |
| Ar | 23.2 |
| H ₂ O | 6.1 |
| P _{TOTAL} | 765.28 |

Theoretical outputs in torr are as follows:

| | |
|------------------|--------|
| N ₂ | 563.17 |
| O ₂ | 151.4 |
| CO ₂ | 16.04 |
| Ar | 22.58 |
| H ₂ O | 6.1 |
| P _{BAR} | 759 |

The variations between theoretical outputs and actual outputs after the five day sea trial are as follows:

| | | |
|-------------|--------|-------------------------|
| N_2 | + 2.84 | Torr out of 563.17 torr |
| O_2 | + 2.4 | Torr out of 151.4 torr |
| CO_2 | + 0.14 | Torr out of 16.04 torr |
| Ar | + 0.62 | Torr out of 22.6 torr |
| H_2O | 0.00 | Torr -- |
| P_{TOTAL} | 6.28 | Torr out of 759 torr |

Comparing the instrument calibration prior to the five day sea trial with the calibration verification after the sea trial, it can be seen that the instrument performance has remained within 1% on all parameters.

APPENDIX 2

EVALUATION OF A FOUR GAS MASS SPECTROMETER
USED FOR ATMOSPHERIC CONTROL DURING THE
NINETY DAY TEST

By Michael R. Ruecker

SUMMARY

The design and performance of a Mass Spectrometer Atmospheric Sensor System which was utilized for monitoring and control of the atmosphere of a manned space station simulator during a 90 day test is reviewed. The instrument was a modified Two Gas Atmosphere Sensor System which was operated with a new closed loop electronics control system for improved long term stability. Based upon calibration verification data taken during the 5-day and 90-day runs the instrument was demonstrated to hold its calibration within one percent for nitrogen, two percent for oxygen and three percent for carbon dioxide for a period of 132 days. It also monitored water vapor partial pressure. The output signal from the oxygen channel was employed by an atmosphere control system for maintaining the oxygen partial pressure of the Space Station Simulator. The instrument demonstrated its ability to perform reliably and its potential value as equipment for ECS applications.

INTRODUCTION

In 1965 a phased program was initiated with Langley Research Center aimed at the development of a mass spectrometer system for monitoring the major constituents of a buffered two gas atmosphere as well as the primary metabolic products of respiration. The Two Gas Atmosphere Sensor System, a single focusing magnetic sector mass spectrometer, evolved from this program, which was capable of continuously monitoring nitrogen, oxygen, carbon dioxide, and water vapor. An engineering test model and four prototype units were fabricated on this program. One of these units is shown in Figure 9. These instruments have been employed in several applications for atmospheric and respiratory measurements in various laboratories, a space station simulator and two undersea habitats. One of these applications was in conjunction with the 60-Day Manned Space Cabin Simulator Test in 1968. At that time the Two Gas Atmosphere Sensor System was operated externally to the Space Cabin Simulator with a laboratory vacuum system, and sampled the cabin atmosphere through a capillary-bypass inlet system.

In the most recent application, the subject of this paper, one of the original instruments was refurbished, equipped with updated electronics, repackaged with a close coupled ion pump and a direct entry sample inlet system, and mounted inside the Space Station Simulator where it monitored the partial pressures of oxygen, nitrogen, carbon dioxide, and water vapor. The output signal of the mass spectrometer's oxygen channel was provided as the input to the atmospheric control system which controlled the oxygen partial pressure. The performance of the combined system was more than adequate to hold the oxygen partial pressure within the limits required for constant physiological functioning of the crew.

PRINCIPLES OF OPERATION

The Two Gas Atmosphere Sensor System is a single focusing magnetic sector mass spectrometer that is designed to provide four simultaneous outputs that are proportional to the partial pressures of N_2 , O_2 , CO_2 and H_2O . The fundamentals of the operation of the mass spectrometer are diagrammed in Figure 10. A small quantity of the gas sample to be analyzed is continuously introduced to the mass spectrometer through a molecular inlet leak. The characteristics of this leak allow each constituent of the sample to flow through the leak independent of the other components. The resulting partial pressures within the ion source are proportional to the corresponding partial pressures in the sample environment.

The ion source performs the function of ionizing part of the gas to form charged particles which are then acted upon by the electrostatic and magnetic fields within the instrument. Ionization is accomplished by bombardment of an electron beam which is derived from a hot wire filament. The ions are repelled from the ionizing region, focused by an electrostatic lens and passed through the ion source exit slit into the magnetic sector. A permanent magnet provides a uniform magnetic field through which the ion beam passes within the vacuum envelope. The ions are deflected into circular arcs by the magnetic field, their radii being proportional to the square root of the mass to charge ratios of the ions. Since all of the ions of interest are singly charged, the radii are proportional to $m^{1/2}$. Consequently, the ions are dispersed as they leave the magnetic field and collected by four Faraday cage type collectors located along a focal plane. The collectors are attached to single pin feedthroughs that pass the current through the vacuum envelope to four electrometer amplifiers that amplify the small currents to provide output voltages which are proportional to the ion currents. The output signals are therefore proportional to their respective partial pressures. The internal vacuum necessary for operation of the analyzer is maintained by a suitable high vacuum pump which is connected to the mass spectrometer by means of a pump tube.

The Two Gas Atmosphere Sensor mass spectrometer analyzer assembly is shown in Figure 11. The vacuum envelope, permanent magnet, single pin feedthroughs and the pump tube are clearly visible. The multipin feedthroughs that are visible in the ion source housing provide the voltages that operate the ion source filament and focusing electrodes.

REQUIREMENTS FOR THE 90-DAY TEST

The requirements for the 90-Day Space Station Simulator (SSS) application were to monitor the partial pressures of nitrogen, oxygen, carbon dioxide, and water vapor using a modified Two Gas Atmosphere Sensor. The principal requirements are summarized in Table 2-1.

The instrument was to be located within a specified volume within the SSS and to give continuous outputs within a specified tolerance for the entire 90-day period without requiring recalibration. In order to provide information for engineering evaluation of the instrument's performance, a method of making calibration verification was required. The instrument was to be provided with the necessary support equipment to maintain its internal vacuum through a power failure. Dual outputs were necessary for internal signal monitoring by meters as well as voltage outputs to be monitored by the computer system external to the SSS.

TABLE 2-1

Requirements for 90-Day SSS Atmospheric Sensor

| | |
|-----------------------|--|
| Monitored Species: | H ₂ O, N ₂ , O ₂ and CO ₂ |
| Monitored Masses: | m/e 18, m/e 28, m/e 32, and m/e 44 |
| Full Scale Ranges: | 20 torr, 500 torr, 200 torr and 120 torr respectively |
| Total Pressure: | Nominally 10 lbf/in ² abs or 517 torr |
| Configuration: | Limited size consistent with available space and commercial support components |
| Operating Controls: | None for normal operation. Inlet system valving for initial setup and calibration verification. Power on-off, ion pump and emission current adjusts. |
| Maintenance: | None |
| Outputs: | Internal: four meters Remote: four buffered, linear, zero to 5 volt |
| Performance Monitors: | Anode current, ion pump current, and battery voltage |
| Sample Inlet: | Sample transport line w/3 inch H ₂ O head |
| Nominal Accuracy: | +2% of full scale for N ₂ and O ₂ +3% of full scale for CO ₂ +5% of full scale for H ₂ O |
| Environment: | Compatible with operation in the Space Station Simulator |

SYSTEM DESCRIPTION

The description of the 90-Day SSS Atmospheric Sensor is facilitated by considering its major system components which are: first, the sample inlet and calibration inlet system; second, the mass spectrometer subsystem including its vacuum pump; and third, the support electronics subsystem which includes the electronics required to operate the analyzer and ion pump as well as the output circuits. A block diagram of the system is shown in Figure 12 and can be used for reference in the following discussion. The sample and calibration inlet system shown in the upper left hand corner of Figure 12 is shown in greater detail in Figure 13 with the front panel of the system shown in Figure 14.

To simplify the system description, each part of the system referred to has been assigned an index number as shown in Figures 13 and 14. Sample gases enter the inlet system from the one-eighth inch sample line at the sample inlet point (1). The sample gas then passes through a needle flow control valve on the front panel of the instrument (2). After passing through the flow controlling valve the sample gas goes to the mode selector valve (3) which determines the mode of operation, that is, operating in the calibration mode or the normally operating sample mode. After passing through the mode selector valve the gas is filtered by a two stage inline filter (4). On the sample outlet corresponding sets of filters (5) are present. The gas mixture then passes through a sample flowmeter (6), which measures the rate of gas flow through the instrument, and therefore, allowing the pressure drops through the inlet system to be checked. After passing through the flowmeter the sample travels past a total pressure transducer (16) and out the sample vent (17). Between the double filters the gas passes through the variable leak valve (7). This variable leak valve is fitted with a temperature control system. The heater switch (8) controls the heater for the inlet valve. The calibration gas mixture is stored in a pressure tank (9) with regulator (10). Passing through the regulator the sample gas goes to a protection shutoff valve (14), and then to a needle flow control valve (15), then to the selector valve (3). Part of the gas to be sampled (either in the sample gas or the calibration gas) passes through the restriction in the leak valve and through a small diameter line (11) into the mass spectrometer (12), and finally to the ion pump (13). There is a roughing valve (25) located within the analyzer chassis for initially pumping down the instrument. The conductance of the variable leak valve can be adjusted by means of a slotted screw adjustment (28) on the front panel. The ion currents coming out of the mass spectrometer are detected and amplified by four electrometers. The electrometer outputs go to the output meters (26) and also to buffered outputs. The zero levels of the electrometers can be checked by pressing the press to test button (27) which cuts off the ion beams. The main power to the mass spectrometer is provided by a 28 Vdc power supply, requiring a 115 volt ac input.

There is a front panel switch (18) that controls the operation of the 28 volt power supply. Other items on the front panel are the ion pump meter switch (19) and the ion pump current meter (20). These monitor the current flowing to the ion pump from the high voltage supply, and therefore, indirectly monitor the analyzer pressure. There are two other meters on the front panel of the Analyzer Control Module. One of these is the battery voltage indicator (21). This indicates the stage of readiness of the emergency batteries that are used for powering the ion pump in a power off situation. The other front panel meter is the anode current meter (22). This meter measures the anode current and gives an indication of the sensitivity at which the source is being operated. The anode current may be adjusted only in open-loop mode by the anode current adjustor potentiometer (23). The mode of operation, open or closed loop, is controlled by a selector switch (24) on the front panel.

An important feature of the system is the closed loop mode of operation which automatically compensates for any common mode variations. The four electrometer outputs are scaled to provide signals that are all proportional to pressure with the same volt per torr sensitivity and summed to give a "total pressure" signal. Since the four components of interest comprise essentially all of the atmosphere this signal can be compared with the output of a total pressure transducer, which is reflecting the true pressure seen by the sample inlet system. The resulting error signal represents the sensitivity error of the mass spectrometer. This signal is fed back to the emission regulator that controls the level of ionizing electron current in the ion source and, thereby, the levels of the ion currents which are detected at the collectors. In this way the summation of the partial pressures is held at the prevailing ambient pressure level and, consequently, common mode variations due to such factors as changes in the inlet leak conductance or ion source sensitivity are eliminated. This method of operation represents a significant improvement in mass spectrometer design and allows a high level of accuracy to be maintained for a long period of time.

Other elements of the system, which are shown in Figure 12, are the power supplies that provide voltages to the ion source, the ion pump and its high voltage power supply, the power supply system, the front panel control and monitoring functions, and the output buffer amplifiers.

The complete 90-Day SSS Atmospheric Sensor is shown in Figure 15 and the internal construction of the upper and lower bays is shown in Figures 16 and 17 respectively. The upper bay contains the mass spectrometer, ion pump, their support electronics, the sample inlet and calibration inlet systems, and the calibration gas supply. The lower module houses the main power supply, the battery pack and charger, the buffer amplifiers, output and monitoring meters, and the buffer amplifiers and pressure transducer power supplies.

CALIBRATION

The Atmospheric Sensor was calibrated using a calibrated gas mixture of N₂, O₂ and CO₂. In order to calibrate the instrument as close as possible to the expected operating conditions the calibration gas was admitted to a laboratory inlet system in which the pressure was held at 10 lbf/in² abs (psia) (517 torr) and from this reservoir it was introduced into the mass spectrometer inlet system. The volt per torr sensitivity at the electrometer amplifier output of each channel was computed and this information was utilized to adjust the summing resistors so that the current arriving at the summing junction of the summing operational amplifier from each channel has the same ampere per torr sensitivity. Then the gains of the buffer amplifiers are set so that the proper full scale value for each channel is achieved at five volts. During the dry gas calibration the pressure was exercised between 400 and 634 torr to verify that the channels were tracking pressure. This is a ± 22.6 percent pressure variation which is much greater than the expected variation of the Space Station Simulator atmospheric pressure. The final calibration data over this pressure range is shown in Table 2-2.

TABLE 2-2

Table of Calibration Errors at Final Calibration

| Pressure | Error | | |
|----------|----------------|----------------|-----------------|
| | N ₂ | O ₂ | CO ₂ |
| 400 torr | +0.37% | -0.85% | -2.86% |
| 517 torr | -0.01% | +0.01% | +0.31% |
| 634 torr | -0.29% | +0.58% | +2.87% |

The water vapor channel was calibrated by allowing the inlet system to sample laboratory air and comparing the H₂O electrometer output with the partial pressure, as computed from the relative humidity indicated by a wet bulb - dry bulb Mason's form hygrometer. Since the instrument was sampling at one atmosphere during this test the variable inlet leak valve was closed down to maintain the normal internal partial pressures. The water sensitivity was compared with the nitrogen sensitivity as determined from the air composition and this data was utilized to set the water vapor channel summing resistor and buffer amplifier gain.

OPERATIONAL RESULTS

During the 5-day test run the calibration of the Atmospheric Sensor was verified twice by admitting a calibration gas sample. During the first verification the gains of the buffer amplifiers for the dry gas channels were reset. The oxygen buffer amplifier gain had somehow shifted during the process of shipment, inspection and installation. The other buffer amplifiers were

very close to their proper values. The data from this calibration verification and a second one taken later during the run is shown in Table 2-3. In both cases the error is less than one percent on all dry channels.

During the 5-day test it was found that the water vapor channel was reading high compared with the Cambridge Dew Point Hygrometer. The dew point indicator was reading 11.3 torr while the mass spectrometer was reading 23.1 torr, with a buffer amplifier output of 4.68 volts. Just prior to the end of the run the buffer amplifier was adjusted to give an output of 2.82 volts based upon an 11.3 torr partial pressure and a desired system gain of 20 torr/5 volts. This correction gave the proper water output, but it did not correct the summing resistor for the water channel, which was also apparently in error. One effect of this adjustment is seen in the data presented in Figure 18. This shows the error between the sum of the partial pressures as indicated by the mass spectrometer and the total cabin pressure as measured by the Wallace-Tiernan gauge. During the 5-day run the sum of the partial pressures agreed with the total pressure within one percent or better except for the final reading which was 2.9 percent low. This final value was taken after the water vapor channel was readjusted. This additional two percent error amounts to an error of 10.5 torr at 525 torr total pressure, which is very nearly equal to the 11.8 torr error that existed in the water output prior to adjustment of the buffer amplifier.

This error results from the action of the closed-loop, which makes up for the erroneously high H₂O electrometer amplifier output by dropping the ionizing current to achieve the correct total pressure. Reduction of the H₂O buffer gain led to a low value for the summation of the partial pressures.

TABLE 2-3

Results of Calibration Verification During the 5-Day Test

| | ERRORS, PERCENT | | | |
|---------|-----------------|----------------|-----------------|-------|
| | N ₂ | O ₂ | CO ₂ | PP |
| INITIAL | +0.23 | +0.23 | +0.47 | +0.23 |
| FINAL | -0.91 | -0.71 | -0.70 | -0.84 |

During the 90-day test the Atmospheric Sensor performed without malfunction. Five calibration verifications were run during the course of the test and the data from these is presented in Table 2-4. This data shows that the mass spectrometer maintained its calibration within very close tolerances. The last adjustment of the instrument was made on 30 April 1970. Therefore, the analyzer maintained its calibration on N₂, O₂, and CO₂ within 0.9, 2.1 and

2.7 percent respectively for a period of 132 days. The sum of the dry gas partial pressures remained within 0.84 percent during the same period. It is difficult to evaluate the performance of the instrument by any means other than the calibration verification data. Typical output data for the instrument is shown in Figures 19 through 21. This data was taken from the computer reduced data obtained on the MDAC Low Speed Data Acquisition System at 1800 hours on each day of the test. This time was selected because it was considered the most "normal", with the least unprogrammed activity, and should give a more representative picture of the cabin atmosphere from day to day.

Figure 19 shows the variations in the oxygen and nitrogen partial pressures; Figure 20, the variations in the carbon dioxide and water vapor partial pressures; and Figure 21, the variation the cabin pressure, the sum of the partial pressures and the difference between these two values. A cursory review of the data has indicated that the fluctuations in the partial pressures are usually accounted for by specific known events that occurred within the SSS. The oxygen partial pressure was controlled within a total variation of 2.9 torr or better than ± 1 percent. The total pressure variations are much wider primarily due to a lower gain in the nitrogen make up portion of the atmospheric control system. The variations in the carbon dioxide and water vapor partial pressures reflect changes in the status of the solid amine and molecular sieve CO₂ scrubber systems. The summation of the partial pressures is consistently low because of the incorrect summing resistor in the water channel as predicted by the last data point taken during the 5-day test run. This effect from the water channel is further substantiated by comparing the correlation between the water vapor output and the error in the summation of the partial pressures. Note that whenever the water vapor level goes up the summation of the partial pressures goes down and visa versa. This is exactly what was expected from a detailed analysis that was made of the interchannel effects of an incorrectly established summing resistor.

TABLE 2-4

Results of Calibration Verification During the 90-Day Test

| DATE | TIME | Error, Percent | | | |
|------|------|----------------|----------------|-----------------|------|
| | | O ₂ | N ₂ | CO ₂ | PP |
| 6-13 | 1315 | +1.1 | +0.1 | -1.5 | +0.4 |
| 6-18 | 1902 | +0.7 | -0.4 | 0.0 | 0.0 |
| 7-8 | 1135 | +0.9 | -0.6 | +0.8 | 0.0 |
| 7-13 | 1346 | +1.3 | -0.5 | +0.6 | 0.0 |
| 9-9 | 0250 | +2.1 | -0.3 | +2.7 | +0.1 |

During the course of the 90-day test it was found that the CO₂ output of the mass spectrometer was not in agreement with the infrared analyzer. Consequently, on 21 August 1970 a portion of the gas utilized to make the initial calibration of the mass spectrometer was sent to the laboratory that made the original mixture analysis. The results are shown in Table 2-5.

TABLE 2-5
Comparison of Calibration Gas Analyses

| Component | Mixture 7130P | |
|----------------|---------------|---------|
| | 1/30/70 | 8/21/70 |
| Nitrogen | 66.166 | 66.336% |
| Oxygen | 31.271 | 31.632% |
| Carbon Dioxide | 2.544 | 2.012% |

Note that the calibration for CO₂ changed by more than 20 percent. If the later calibration figures are utilized the agreement with the infrared analyzer is very close. There is no reason to suspect that the calibration gas changed during this period, and therefore, it must be concluded that the original calibration was in error.

At the conclusion of the 90-day test the Atmospheric Sensor was shut down and returned to Perkin-Elmer Aerospace Division where it is now being operated on laboratory ambient atmosphere for a period of 180 days.

CONCLUSIONS

The Atmospheric Sensor was shown to be a reliable and accurate instrument for monitoring nitrogen, oxygen, carbon dioxide, and water vapor during the course of the 90-Day Manned SSS Test. It demonstrated its capability not only to monitor these constituents but to provide outputs that could be utilized by an atmospheric control system for regulation of the primary atmospheric constituents of a closed environment. The closed-loop operating mode controlled the sensitivity of the mass spectrometer so that it could operate for a period of 132 days without a calibration. The accuracy of the outputs was affected by the initial calibration which was found to be in error because of a faulty calibration technique in the case of water vapor, and an inaccurate calibration gas mixture in the instance of carbon dioxide. These procedural matters have been rectified and should allow the Atmospheric Sensor to perform to its full capability in future applications.

The 90-Day Manned SSS Test was intended to prove out equipment for application to future space stations. It is therefore important to assess the feasibility of reducing the size, weight and power of the Atmospheric Analyzer to levels that are compatible with a flight program. A contract, under the direction of NASA Manned Spacecraft Center, is currently in progress for the development of a flight qualified Mass Spectrometer Atmospheric Sensor System (MASS) as well as a modified version to be used as a respiratory gas analyzer as part of the M-171 Metabolic Analyzer which will be used in Skylab in 1972. A photograph of the instrument is shown in Figure 22.

These units are fully self-contained, requiring only a sample inlet and bypass line, a small diameter vacuum line to outer space for initial roughing of the mass spectrometer, system power, and command functions for the ion pump mass spectrometer electronics, open-loop/closed-loop control, and selection of one of the dual ion source filaments. The system is fully protected against operation at excessive pressures and provides status indicator outputs on all important functions that can change during operation. The inputs and outputs are fully isolated and protected and the instrument has sample inlet heaters and ion source temperature control for improved performance.

The design is fully compatible with Apollo and Skylab environments including a 38 lbf/in² abs over-pressure requirement. The atmospheric monitoring version of this instrument measures the partial pressures of hydrogen, water vapor, nitrogen, oxygen, carbon dioxide, and hydrocarbons in the mass range 50 to 120 u (amu). The final configuration of this system weighs 21 pounds, requires 19 watts of power during normal operation and has a cylindrical enclosure with a diameter of 7.2 inches and a length of 12.5 inches. The first design verification test unit of the MASS is scheduled for delivery to NASA/MSC in November 1970.

APPENDIX 3

OPERATION AND MAINTENANCE MANUAL
UNDERSEA ATMOSPHERIC SENSOR

Revised
16 October 1970

Perkin-Elmer SPO 30006
NASA Contract Number NAS1-9469

Prepared for
NATIONAL AERONAUTICS AND SPACE ADMINISTRATION
LANGLEY RESEARCH CENTER
Langley Station
Hampton, Virginia 23365

DESCRIPTION AND OPERATION

The Undersea Atmospheric Sensor (Figure 23) is a mass spectrometer system which continuously monitors the following atmospheric constituents: water vapor, nitrogen, oxygen, argon, and carbon dioxide. A functional block diagram of the system is shown in Figure 24.

The following brief discussion of the system is given as an overall familiarization with the instrument. The sample to be monitored enters the sample inlet when the sample shutoff valve is opened. It passes through a double filter and then through the bypass line in an inlet leak valve. It continues through a flow meter, another double filter assembly, a flow control valve, and finally to the sample transport pump where it is exhausted to the atmosphere. A small portion of this sample is admitted to the mass spectrometer (called the analyzer) when the leak valve is open. This gas passes into the ion source where it is ionized and the resulting charged particles are resolved in the magnetic field of the analyzer magnet to form separate beams of ion current. The ion current strikes individual collectors located at the proper positions for the masses of interest. This instrument monitors water vapor at mass 18, nitrogen at mass 28, oxygen at mass 32, argon at mass 40, and carbon dioxide at mass 44. The ion current is converted to electron current in the leads attached to the collectors. These leads pass through single-pin feed-throughs to electrometer amplifiers. The amplifiers amplify the currents, which are in the range of 10^{-14} to 10^{-10} amperes, to provide full-scale output voltages of the order of -10 volts. The output signals are scaled by a voltage divider network and then displayed on digital voltmeters which are calibrated in torr (mm of Hg).

A central feature of the mass spectrometer is its closed loop electronic control system which automatically controls the gain of the mass spectrometer against any common-mode errors. This is accomplished by scaling each of the electrometer outputs to give an output voltage which is proportional to torr (i.e., each channel is scaled to have the same gain, since 1 torr of oxygen is the same as 1 torr of nitrogen, etc.). At this point the outputs are summed with an operational amplifier to give a total pressure signal. This signal is compared with the output of a total pressure transducer which is located in the sample inlet line, and the error signal is fed back to the emission regulator where it is used to control the filament drive and the resulting ionizing current.

Other components of the electronics subsystem are the power supply system consisting of the following: 1) an EMI filter and transformer; 2) an electrode bias supply and voltage divider network, which provides voltages to operate the ion source; and 3) the B+ and B- supply for the electrometer amplifiers, summing amplifier, and total pressure transducer amplifier.

In addition to the inlet system, analyzer, and electronics subsystem described above, there is a vacuum support system. Most of these components are located in the upper module, whereas the inlet system, analyzer, and electronics subsystem are located in the lower module. A 20 liter/second ion pump provides the primary vacuum required for the operation of the mass spectrometer. It is attached to the analyzer by means of a vacuum adapter which also interfaces the thermocouple gauge, the roughing line, and the ionization pressure gauge (for laboratory use only). The ion pump is operated by a 5 kV power supply, and is controlled by the output of the thermocouple gauge, so that it can only turn on when the pressure is below a preset level. The ion pump current, which indicates the internal pressure, is sensed by an overpressure protection circuit which inhibits the mass spectrometer electronics unless the ion pump pressure is less than 1.5×10^{-6} torr (500 μ A).

When the pressure is above the point at which the ion pump can turn on, it is necessary to rough down the system through the roughing valve. This is accomplished by a mechanical rough pump which is protected by a molecular sieve trap to prevent backstreaming oil vapors from contaminating the analyzer.

Also housed in the upper module is a temperature controller which drives the heaters on the inlet leak valve and the connecting line to the analyzer.

The Undersea Atmospheric Sensor is housed in three modules:

- a. Display Module: Contains four digital output meters.
- b. Upper Module: Contains vacuum support components and the inlet leak valve temperature control.
- c. Lower Module: Contains the analyzer, mass spectrometer electronics, inlet system, and ion pump.

The Upper and Lower Modules are covered with separate panel covers which may be removed with quarter-turn fasteners (Figure 25).

Table 3-1 along with Figures 26 (a-c) identify all system controls which will be discussed in this manual.

TABLE 3-1
Controls and Monitors

| NO. | CONTROL OR MONITOR | LOCATION | | TITLE | FUNCTION | REMARKS |
|-----|------------------------------|----------|-------------|------------------|---|--|
| | | MODULE | AREA | | | |
| 1. | Nitrogen Display Meter | Display | Upper Left | N ₂ | Displays the nitrogen partial pressure in torr. | |
| 2. | Oxygen Display Meter | Display | Lower Left | O ₂ | Displays the oxygen partial pressure in torr. | |
| 3. | Carbon Dioxide Display Meter | Display | Lower Right | CO ₂ | Displays the carbon dioxide partial pressure in torr. | |
| 4. | Water Vapor Display Meter | Display | Upper Right | H ₂ O | Displays the water vapor partial pressure in torr. | Also displays 15 other parameters as selected by monitor switch. |
| 5. | Ion Pump "On" Light | Upper | Upper Left | High Voltage | When lit, indicates that high voltage is being applied to the ion pump. | Lights when the adjacent circuit breaker is turned on. |
| 6. | Ion Pump "On" Breaker | Upper | Upper Left | (None) | Turns on the ion pump power supply. | |
| 7. | Ion Pump "Start" Light | Upper | Middle Left | Start | Indicates when the start-run switch is in the start mode. | Start mode allows maximum (150 μA current for starting. Run mode turns off the ion pump at or above 4 x 10 ⁻⁵ torr. (10 μA) |
| 8. | Ion Pump "Start-Run" Switch | Upper | Middle Left | Start-Run | Places the ion pump power supply in the start or run mode. | |

See Figure 26a

See Figure 26b

TABLE 3-1 (Cont)

| NO. | CONTROL OR MONITOR | LOCATION | | TITLE | FUNCTION | REMARKS |
|-----|--|----------|--------------|-------------------|---|---|
| | | MODULE | AREA | | | |
| 9. | Ion Pump Control Meter | Upper | Upper Center | None | Indicates ion pump current, pressure, or voltage as selected by the Range Selector Switch. | Ion pump current is proportional to the internal pressure. See Figure 27. |
| 10. | Ion Pump Meter Range Selector Switch | Upper | Upper Right | Range Selector | Selects ion pump current range, pressure or voltage. | |
| 11. | Thermocouple Gauge Readout Meter | Upper | Lower Left | Analyzer Pressure | Indicates the internal pressure in microns (black needle). | |
| 12. | Thermocouple Gauge Set Point Control Knob | Upper | Lower Left | None | Sets the trip point (red needle) below which the ion pump is allowed to turn on. | Set so that the trip point occurs when the indicated pressure is 10 μ . |
| 13. | Rough Pump On-Off Circuit Breaker | Upper | Lower Left | Rough Pump | Turns on the rough pump to initially evacuate the system. | Normally off. |
| 14. | Trap Heater On-Off Switch | Upper | Lower Center | Trap Heater | Turns on the molecular sieve trap for regeneration of the trap. | Disconnected. |
| 15. | Current Relay Power On Light | Upper | Lower Right | Power | Light turns on when the ion pump control is turned on. | Redundant to the high voltage light. |
| 16. | Ion Pump Current Set Point Selector Switch | Upper | Lower Right | None | Sets the ion pump current below which the mass spectrometer electronics are allowed to turn on. | Set at 500 μ A. Trips at about 440 μ A. |

See
Figure
26b

TABLE 3-1 (Cont)

| NO. | CONTROL OR MONITOR | LOCATION | | TITLE | FUNCTION | REMARKS |
|-----|-------------------------------------|----------|-------------------|----------------------|---|--|
| | | MODULE | AREA | | | |
| 17. | Ion Pump Current Relay Light | Upper | Lower Right | Relay Light Energize | Light turns on when ion pump current is below the relay set point. | Mass spectrometer electronics can be turned on when light is on. |
| | | Lower | Lower Left | None | Turning the double knurled knob CCM opens the inlet leak admitting sample gas to the mass spectrometer. | Caution must be used when adjusting valve. Monitor the ion pump current meter (200 μ A range). There will be no indication for a few turns but then the pressure rises very rapidly. |
| 19. | Sample Shut Off Valve | Lower | Upper Left Center | Sample Shut Off | When opened this valve allows sample gas to flow through the inlet system. | Use for shut off only. |
| 20. | Flow Control Valve | Lower | Upper Left Center | Flow Control | Adjusts the flow rate of sample gas through the sample inlet system. | Not to be used for shut off. |
| 21. | Flow Meter | Lower | Upper Center | None | Indicates the flow rate of sample gas through the inlet system in CFM. | Normally set at 0.015 CFM |
| 22. | Sample Transport Pump On-Off Switch | Lower | Center | Trans. Pump | Turns on the sample transport pump which pulls sample gas through the inlet system. | |
| 23. | Fuse Holder | Lower | Left Center | F2-3A | 3 amp fuse in sample transport pump power line | |

See Figure 26b

See Figure 26c

TABLE 3-1 (Cont)

| NO. | CONTROL OR MONITOR | LOCATION AREA | | TITLE | FUNCTION | REMARKS |
|-----|---|---------------|--------------------|-----------------|--|---|
| | | MODULE | AREA | | | |
| 24. | Fuse Holder | Lower | Right Center | F3-5A | 2 amp fuse in valve heater power lines | |
| 25. | Roughing Valve | Lower | Lower Center | Rough Valve | This valve is opened (CCW) when it is desired to initially evacuate the analyzer during a pump down cycle. | DO NOT TOUCH unless it is intended to rough down the analyzer. |
| 26. | System Power Circuit Breaker | Lower | Lower Center | System Power | Primary power circuit breaker for entire system. | The thermocouple gauge and two cooling fans are directly turned on by this breaker. |
| 27. | Mass Spectrometer Power "On-Off" Switch | Lower | Lower Right Center | M/S | Turns on the mass spectrometer electronics when not inhibited by the ion pump current relay. | |
| 28. | Filament Select Switch | Lower | Lower Left Center | Fil 1-Fil 2 | Selects which ion source filament the mass spectrometer is operating on. | Initially set on filament #2. Switch only if a filament malfunction develops. |
| 29. | Inlet Leak Valve Heater "On-Off" Switch | Lower | Lower Left Center | Heater | Turns on the inlet leak valve heater and temperature controller. | |
| 30. | Open Loop Emission Adjust Potentiometer | Lower | Lower Left Center | Emission Adjust | Adjusts the emission when in open loop mode and sets the maximum closed loop emission. | |

See
Figure
26C

TABLE 3-1 (Concluded)

| NO. | CONTROL OR MONITOR | LOCATION | | TITLE | FUNCTION | REMARKS |
|-----|---------------------------|----------|--------------------|--------------|--|---|
| | | MODULE | AREA | | | |
| 31. | Loop Mode Switch | Lower | Lower Left Center | Loop Mode | Selects system control mode - open or closed. | Normally set to closed. |
| 32. | Zero Check Push Button | Lower | Lower Left Center | Zero Check | Used only in open loop mode to check electrometer zeros. | |
| 33. | Test Mode Selector Switch | Lower | Lower Extreme Left | Test Monitor | Switches the H ₂ O digital panel meter to various test points throughout the system. See Table 3-2. | Normally set on H ₂ O. |
| 34. | Fuse Holder | Lower | Lower Left Center | 1/2 ASB | M/S electronics fuse. | |
| 35. | Ion Gauge | Lower | Lower Center | None | Ionization gauge for initial laboratory setup. Requires ionization gauge controls for operation. | Function not required for normal operation of system. |

See Figure 26C

UNDERSEA ATMOSPHERIC SENSOR OPERATING PROCEDURES

INITIAL PUMP DOWN PROCEDURE (Reference Paragraph 2-1)

- a. Remove the front cover panel.
- b. Check the following switches. They should be in the positions indicated:
 1. System power OFF
 2. M/S power OFF
 3. Ion pump control power OFF
 4. Rough pump OFF
 5. Loop mode OPEN
 6. Filament select FIL #1 or #2
 7. Ion pump, start-run START
 8. Ion pump range selector 200 mA
 9. Emission adjust Full CCW
 10. Test monitor switch H₂O
- c. Check the following controls. They should be in the positions indicated:
 1. Sample shut off Closed
 2. Variable leak valve Closed
 3. Roughing line valve Closed
- d. Set the SYSTEM POWER circuit breaker switch to ON. This energizes the fans and the thermocouple gauge control (analyzer pressure).
- e. Allow the Thermocouple Gauge to warm up for a few minutes and observe the reading. If the reading is less than or equal to 10 microns, the ion pump may be started immediately. (See Step i.) If the pressure remains above 10 microns, continue to the next step.
- f. Set the ROUGH PUMP circuit breaker to ON. This commences the pump-down of the roughing line.
- g. Wait 10 minutes for the line to pump down.
- h. Open the ROUGHING LINE VALVE. This begins the evacuation of the analyzer.

- i. Observe the Thermocouple Gauge, and when the pressure approaches 10 microns set the ION PUMP CONTROL power circuit breaker to ON. The ion pump control will remain inhibited until the pressure reaches 10 microns, at which time it will turn on automatically.

CAUTION

From the time the operator turns on the ion pump control power until the ion pump control is switched to RUN, the operator must remain in attendance.

- j. Observe the ion pump control meter and watch for a current indication on the 200 milliamper range. When the control unit switches ON there will be an indication on the meter. Also the power on (HIGH VOLTAGE) light will illuminate.
- k. When the ion pump turns ON as indicated above, immediately close the ROUGHING LINE VALVE.

CAUTION

Prolonged operation of the ion pump with the ROUGHING LINE VALVE open can damage the ion pump.

- l. Continue to observe the pumpdown of the analyzer on the Ion Pump Control Meter. As the pressure reduces, the ion pump current will go down and its progress can be observed by turning the RANGE SELECTOR switch as indicated in the table below. The ion pump pressure can be checked by referring to Figure 27 which gives the correlation between pressure and current.

| RANGE | RANGE SWITCH POINT | NEXT RANGE |
|-------------|--------------------|-------------|
| 200 mA | 2 | 20 mA |
| 20 mA | 2 | 2 mA |
| 2 mA | 2 | 200 μ A |
| 200 μ A | 2 | 20 μ A |
| 20 μ A | 2 | ---- |

- m. When the ion pump current reaches 200 μ A, the M/S power toggle switch may be turned ON. This turns on the mass spectrometer electronics. The ROUGH PUMP switch may now be turned OFF and the VALVE HEATER should now be turned ON.

- n. Wait 5 minutes for the electronics to warm up and then observe the outputs as indicated by the four digital meters. The outputs should be less than the values indicated in the table below. The Argon output can be checked by turning the TEST MONITOR switch to AR.

| CHANNEL | ACCEPTABLE ZERO LEVEL (In Digits) |
|-----------------------------------|--------------------------------------|
| Nitrogen (N ₂) | 001. |
| Oxygen (O ₂) | <u>±</u> 000.1 |
| Carbon Dioxide (CO ₂) | <u>±</u> 00.1 |
| Water Vapor (H ₂ O) | <u>±</u> 00.1 |
| Argon (AR) | <u>±</u> 00.1 |

- o. After the zero levels have been established the mass spectrometer background outputs can be established. Turn the TEST MONITOR switch to ANODE CUR, and then slowly turn the EMISSION ADJUST clockwise (CW), until a reading is seen on the TEST MONITOR meter. Continue turning the EMISSION ADJUST until output of the TEST MONITOR digital meter reads 20.0; this corresponds to 20 μ A anode current.
- p. When the emission current is set the outputs as read on the digital meters will indicate the background signal existing within the analyzer. As the analyzer warms up due to the heat input from the filament and the valve, the Water output may rise but it should go back down. After the valve heater has been on 1 hour, the background levels should be below the values shown in the following table. The VALVE TEMP position of the TEST MONITOR selector switch should indicate 20.0 to 22.0. Again Water and Argon are checked by operating the TEST MONITOR selector switch to the desired locations.

| OUTPUT | ACCEPTABLE BACKGROUND LEVELS (Torr) |
|-----------------------------------|---|
| Nitrogen (N ₂) | 000. |
| Oxygen(O ₂) | 000.2 |
| Carbon Dioxide (CO ₂) | 00.1 |
| Water Vapor (H ₂ O) | 00.5 |
| Argon (AR) | 00.1 |

- q. After establishing the background levels the sample may be introduced. Turn ON the sample TRANS PUMP and turn the SAMPLE SHUT OFF valve to full OPEN.
- r. Adjust the FLOW CONTROL valve until a flow of 0.015 CFM is read on the Flow Meter.
- s. Turn the Ion Pump Control RANGE SELECTOR switch to 200 μ A.
- t. Admit the sample to the mass spectrometer by opening the Inlet Leak Valve (Item 18, Figure 26c.). Turn the double lock knob counter-clockwise (CCW); 1 1/2 revolutions of the knob will be required before the valve starts to open. At this point turn the valve VERY CAUTIOUSLY to avoid over pressurization. Watch the ion pump control meter and bring the current up to 100 μ A.
- u. Turn the TEST MONITOR switch to the ANODE CUR position.
- v. Adjust the EMISSION ADJUST until the anode current is 35.0 on the Digital Panel Indicator (i.e., 35.0 μ A anode current).
- w. Place the LOOP MODE switch in the CLOSED position. The anode current should reduce to approximately 20 μ A.
- x. Readjust the Inlet Leak Valve until the anode current is 20.0 μ A. The ion pump current should be very nearly 100 μ A.
- y. Turn the TEST MONITOR selector switch to H₂O.
- z. The system is now operational and the outputs may be read on the Digital Display with the exception of Argon (AR).
- aa. Replace the cover panels and secure.

CALIBRATION VERIFICATION

The purpose of the Calibration Verification is to determine whether the Under-sea Atmospheric Sensor is in calibration. It should be realized that if a calibration gas mixture is used which does not have constituents which are known to a high degree of precision, or which is different from the mixture which was used to calibrate the instrument, then the indicated outputs may differ from the expected outputs. In this case it is necessary to consider the outputs only on a relative, day to day basis. To avoid this problem it is recommended that the calibration verification of the instrument be carried out using the gas mixture supplied by Perkin-Elmer.

- a. Attach the Calibration Verification Gas Mixture to be used to the Sample Inlet Point.

- b. Set a small positive pressure (less than 1 pound) on the pressure regulator and purge the inlet system for about 30 seconds by opening the FLOW CONTROL valve full CCW. Next, adjust the sample flow with the FLOW CONTROL valve to the normal flow rate of 0.015 CFM as seen on the Flow Meter. Caution must be used at this point not to over pressure this inlet system. Monitor the transducer pressure on the H₂O Digital Panel Meter by switching the TEST MONITOR Switch to PRESS (pressure). Do not exceed 80.0 (800 torr). It may be necessary to reduce the sample gas pressure regulator setting.
- c. Allow 2 to 5 minutes for the calibration gas to stabilize within the inlet system and the water vapor to be removed from the inlet system. The WATER OUTPUT can be monitored to determine when a minimum output level is reached (less than 3 torr).
- d. When stable outputs have been reached, compare them with predetermined values as indicated by the following computation. It is necessary to know the total applied pressure in torr (mm Hg) in order to make the comparison. This is obtained from the pressure transducer output. Set the TEST MONITOR Switch on PRESS and multiply by 10 (i.e., 76.0 = 760 torr). If the water output is significant (i.e., greater than 1 torr) when monitoring the dry gas mixture it is necessary to correct the Total Pressure by the following computation in order to correct Calibration Verification Values. Corrected Total Pressure (torr) = Cabin Pressure (torr) - H₂O Pressure (torr).

| COMPONENT | PERCENTAGE (%) | | CORRECTED TOTAL PRESSURE (torr) | = | PARTIAL PRESSURE (torr) |
|----------------|----------------|---|---------------------------------|---|-------------------------|
| Nitrogen | TBD* X 0.01 | X | _____ | = | _____ |
| Oxygen | TBD* X 0.1 | X | _____ | = | _____ |
| Carbon Dioxide | TBD* X 0.01 | X | _____ | = | _____ |
| Argon | TBD* X 0.01 | X | _____ | = | _____ |

*See percentage composition attached to calibration gas bottle.

- e. Then note the partial pressures as read on the Display Module and compare them with the computed values above.
- f. If the above comparison is acceptable, disconnect the Calibration Verification Gas Mixture and return to normal atmospheric sampling.

OPERATION

The Undersea Atmospheric Sensor will operate for extended periods of time without any monitoring or adjustment. The following procedure is given to allow an operator to check the system operation.

- a. Record the outputs as given by the Digital Display.
- b. Note any significant variations from the previous readings.
- c. Add the outputs as given by the Display for Nitrogen, Oxygen, Carbon Dioxide, and Water Vapor.
- d. Add 7 torr for the Argon partial pressure.
- e. Compare this total to the total pressure for agreement.
- f. If the outputs appear to be normal based upon these tests, then no further check is necessary.
- g. If anything appears to be abnormal, a further check of the operation can be made as indicated in the following section.

OPERATION VERIFICATION

- a. Verify that the sample transport line is attached to the SAMPLE INLET and that all external valving or plumbing is correct (i.e., not plugged or kinked).
- b. Remove the FRONT PANEL.
- c. Check to see that all switches and controls listed below are in the positions indicated.

| | |
|----------------------------------|------------------|
| 1. System Power Switch | ON |
| 2. M/S Power Switch | ON |
| 3. Ion Pump Control Power Switch | ON |
| 4. Rough Pump Switch | OFF |
| 5. Loop Mode Switch | CLOSED LOOP |
| 6. Filament Select Switch | FIL #2 |
| 7. Ion Pump Start-Run Switch | RUN |
| 8. Test Monitor Switch | H ₂ O |
| 9. Sample Shut Off Valve | Full CCW |

d. Check the following meter displays.

| | READING | TOLERANCE |
|---------------------|-------------|------------------|
| 1. Flow Meter | 0.015 | .010 - .020 |
| 2. Ion Pump Current | 100 μ A | 70 - 130 μ A |

e. Turn the TEST MONITOR selector switch through the first 16 positions and observe the values on the TEST MONITOR Digital Display. They should fall within the ranges specified in Table 3-2.

f. If any are out of tolerance, refer to the following section.

TABLE 3-2

| Position | Nominal Value | | Tolerance | | Function |
|------------|---------------|-------|-----------|--------|------------------------------|
| | Set 1 | Set 2 | High | Low | |
| 1 | +10.0 | | 20.0 | 2.0 | H ₂ O Output |
| 2 | +7.0 | | 15.0 | 5.0 | Argon Output |
| 3 | 127.3 | 131.0 | +133.0 | +125.0 | Regulated Elect B+ |
| 4 | 15.7 | | +17.0 | +15.5 | Unregulated Elect B+ |
| 5 | -133.4 | | -145.5 | -125.5 | Regulated Elect B- |
| 6 | -17.0 | | -17.9 | -16.5 | Unregulated Elect B- |
| 7 | +37.0 | +38.5 | +40.5 | +36.6 | Regulated Electrode Bias |
| 8 | +36.5 | | +38.8 | +35.8 | Unregulated Electrode Bias |
| 9 | +19.7 | | +21.0 | +19.5 | High Voltage Regulator |
| 10 | +18.3 | | +19.0 | +17.8 | Emission Regulator Positive |
| 11 | -19.6 | | -20.5 | -19.0 | Emission Regulator Negative |
| EMSN DR | +13.0 | | +19.0 | +10.0 | Emission Regulator Drive |
| ANODE CUR | 20.0 | | 30.0 | 10.0 | Anode Current |
| PRESS | 76.0 | | 80.0 | 70.0 | Total Pressure Transducer |
| VALVE TEMP | 21.0 | | 22.0 | 18.5 | Variable Leak Temperature |
| ACCELR CUR | 45.0 | | 90.0 | 20.0 | Electron Accelerator Current |

ELECTRONIC SUBSYSTEM OPERATION AND MAINTENANCE

GENERAL

The Undersea Atmospheric Sensor System should not be subdivided into independent mechanical or electrical subsystem operation but should be considered as an integral system. Each subcomponent relies upon another for proper performance. DURING SERVICE OF THE SYSTEM, NO POTENTIOMETER SHOULD BE MOVED OR READJUSTED. They have been previously set for system calibration, and readjustment would destroy calibration.

ELECTRICAL SUBSYSTEMS

The electrical subsystems can be classified into two basic categories: the electronic subsystem, and the ac power distribution system.

The electronic subsystem provides the necessary conditioned voltages and signals to power the atmospheric analyzer and provide data outputs.

The ac power distribution system provides the necessary power as well as a measure of protection for all other remaining subsystems, both mechanical and electrical. The ac power interlock protection system is primarily housed in the upper module of the Undersea Atmospheric Sensor. (Ref.: Figure 26c.)

A very useful aid, the Test Monitor Switch (Item 33, Figure 26c) is provided in the lower module. This switch allows the operator to quickly and positively identify a questionable area, should it exist, and lend to its correction with a minimum amount of effort.

The other aids for determining proper instrument operation are located in the upper module, Figure 26b.

Automatic Restart

The Undersea Atmospheric Sensor System has been carefully checked out, calibrated, and placed in stable operating condition. This instrument will remain in this stable, calibrated condition as long as uninterrupted 115 Vac, 60 cps line power is applied through the safety locked convenience outlet box provided. Should line power be interrupted for a short period - less than 30 minutes - no action should be required on the part of the operator. This instrument has been provided with adequate safeguards so that the filament will be turned off, the inlet valve closed, and the ion pump power supply turned off automatically without degrading the instrument performance or calibration. Once line power is restored, the instrument will restart automatically. In approximately 30 minutes after line power is restored, the digital outputs should again read as they did prior to the power shortage.

If for some reason, power to the instrument has been interrupted longer than 30 minutes or the thermocouple gauge readout meter (black indicator hand; Item 11, Figure 26b) reads greater than 10 microns, automatic restarting cannot be accomplished. Manual restart will be required. (Ref.: Page 79, Para 2.1, for restarting the system with the exception of 2.1c, 1 and 2.) If the procedure of Paragraph 2.1 is required, and after all switches, knobs, levels, etc., have been placed in the proper position for starting, a check of the three fuses should be made (Items 23, 24, and 34, Figure 26c). Once it has been determined that these fuses are good, the procedure on Pages 79 through 82, except c, c.1, and c.2 can be followed.

Electronic Component Failure

An undersea atmospheric analyzer engineering data book has been provided which should be maintained once a day. Should the data outputs on the data sheet vary beyond the tolerances provided in Table 3-2 and the summation of the partial pressures, barometric pressure, and pressure transducer output (position 14) not coincide any closer than 1% (7 mm Hg) some electronic component has possibly failed or is in the process of failing. Should this set of circumstances occur, the installed set of seven electronic circuit boards should be replaced in the following manner.

- a. Remove front covers
- b. Switch loop mode switch (Item 31, Figure 26c) into open loop.
- c. Turn emission current potentiometer (Item 30, Figure 26c) full CCW (counterclockwise). Monitor output with test monitor switch in ANODE CUR position. Output will be on the water digital voltmeter channel. This should go to 00.0.
- d. Turn mass spectrometer power switch off (Item 27, Figure 26c).

All seven electronic boards may now be replaced.

NOTE

The shield boards (clad on both sides)
do not require replacement.

Using the card ejector at the top of each card eject one card at a time replacing it with its counterpart from the spares provided. Once all cards have been replaced and are securely installed the mass spectrometer power switch can be switched to the on position, (Item 27, Figure 26c). Turn emission adjust potentiometer up until 35.0 is read on the ANODE CUR position of the TEST MONITOR switch. Switch the LOOP MODE switch (Item 31, Figure 26c) into the CLOSED position. The anode current should drop to approximately 20.0.

Allow 15 minutes for complete system warm up and record all data in Engineering Data Book. The system should once again be in calibration. Cross check the Engineering Data with the permissible tolerance band provided in Table 3-2, according to the board set being used.

CONTINGENCY PROCEDURES

a. Anode Current out of tolerance:

1. Check that ion pump current is $100 \mu\text{A} \pm 20 \mu\text{A}$. If not, readjust leak valve.

CAUTION

Extreme care should be exercised when adjusting this valve.

2. Check Electron Accelerator current. If out of tolerance, switch filaments.
3. If none of the above corrects the problem, switch all electronic circuit cards.

b. Ion Pump Current out of tolerance:


1. Check valve temperature. If out of tolerance, readjust valve setting to make ion pump current $100 \mu\text{A} \pm 20 \mu\text{A}$.
2. Check pressure transducer output and compare to ship's ambient pressure gauge. If they do not compare, check inlet system for proper flows and no obstructions.

c. Loss of electrometer output (H_2O , N_2 , O_2 , CO_2 or AR output)

1. This would be indicated by a no signal output (all zeros) or a greater than full-scale output; and therefore, the ΣPP would not equal the total pressure by a gross amount.
2. In this event, the suspect amplifier should be replaced with its spare counterpart.

CAUTION

Use extreme care in handling these electrometers, since they may be damaged by static charge.







FIGURE 2a








FIGURE 2d



UNDERSEA ATMOSPHERIC SENSOR OUTPUT
DURING SEA DEPLOYMENT
6 OCT 70 THRU 16 DEC 70

B.F. Bayliss

FIGURE 3a









FIGURE 4c









B.F. Bickley 11/4/71


FIGURE 5a



B.F. Bickler 11/17/71

FIGURE 5b









FIGURE 6.- Two Carbon Vanes From Gast Pump - Model 0531-102,347 -
Showing Abrasion During Operation




FIGURE 7.- Two Carbon Vanes From Gas Pump - Model 0531-102, 347 -
Showing Failure and Abrasion

K-E 10 X 10 TO THE CENTIMETER 46 1513
INCHES INCHES
 RUFFEL & KREIN CO.





FIGURE 8

FIGURE 9. Basic Two Gas ATM Sensor



A-105-A2

FIGURE 10. Principles of Mass Spectrometer Operation

FIGURE 11. Two Gas Atmosphere Analyzer Assembly





FIGURE 12. Atmospheric Sensor System Block Diagram



A-842

FIGURE 13. Sample and Calibration Inlet System Schematic Flow Diagram




FIGURE 14. Atmospheric Sensor Front Panel

FIGURE 15. 90-Day Space Station Simulator Atmospheric Sensor System




FIGURE 16. Atmospheric Sensor - Upper Bay

FIGURE 17. Atmospheric Sensor - Lower Bay




FIGURE 18. Comparison of the Summation of Partial Pressures to the Total Pressure During the Five Day Test




FIGURE 19. Oxygen and Nitrogen Partial Pressures During the 90 Day Test




FIGURE 20. Carbon Dioxide and Water Vapor Partial Pressures During the 90 Day Test




FIGURE 21. Comparison of the Summation of Partial Pressures to the Total Pressure During the 90 Day Test




FIGURE 22. Flight Qualified Mass Spectrometer Atmospheric Sensor System for Atmospheric and Respiratory Monitoring

FIGURE 23. Undersea Atmospheric Sensor (UAS)




FIGURE 24. Block Diagram - Undersea Atmosphere Sensor (UAS)

FIGURE 25. UAS (With Panel Covers Installed)




FIGURE 26a. Identification of System Controls

FIGURE 26b. Identification of System Controls




FIGURE 26c. Identification of System Controls




FIGURE 27
Pump Current (20 l/s Ultek Ion Pump)