NASA Technical Memorandum 102775

Laser Transit Anemometer Measurements on a Slender Cone in the Langley Unitary Plan Wind Tunnel

William M. Humphreys, Jr., William W. Hunter, Jr., Peter F. Covell, and Cecil E. Nichols, Jr.

December 1990

(NASA-TM-102775) LASER TRANSIT ANEMUMETER N91-15521
MEASUREMENTS ON A SLENDER CONE IN THE
LANGLEY UNITARY PLAN WIND TUNNEL (NASA)
CSCL 148
G3/35 0326554



Langley Research Center Hampton, Virginia 23665-5225

`

•

! .

Laser Transit Anemometer Measurements on a Slender Cone in the Langley Unitary Plan Wind Tunnel

SUMMARY

A Laser Transit Anemometry (LTA) system has been used to probe the boundary layer on a slender (5° half-angle) cone model in the Langley Unitary Plan Wind Tunnel. The anemometry system utilized a pair of laser beams with a diameter of $40\,\mu\text{m}$ (0.0016 in.) spaced $1230\,\mu\text{m}$ (0.0484 in.) apart to measure the transit times of ensembles of seeding particles using a cross-correlation technique. From these measurements boundary layer profiles around the model were constructed and compared with theory. The tunnel seeding system consisted of a small vibrated fluidized bed containing dry kaolin dust with a nominal particle size of $0.9\,\mu\text{m}$ (3.54 x 10^{-5} in.). The kaolin was injected directly into the boundary layer via nine 1.32-mm (0.052-in.) orifices (three rows of three each) in the model nose. Dry kaolin dust was used as the seeding material to eliminate the problems of condensation present among seeding systems using liquid carriers such as ethanol.

The measured boundary layer profiles representing the boundary layer velocity normalized to the freestream as a function of height above the model surface were collected for zero angle of attack, Mach numbers of 2.5 and 4.5, and Reynolds numbers of 3.3×10^6 / meter and 6.6×10^6 / meter (1.0×10^6) foot and 2.0×10^6 / foot) and were collected in a vertical plane that bisected the model's longitudinal center line at a location 635 mm (25 in.) from the tip of the forebody cone. The results indicated an excellent ability of the LTA system to make velocity measurements deep into the boundary layer. However, because of disturbances in the flow field caused by the onboard seeding system, premature transition to a turbulent boundary layer occured, implying that upstream seeding is mandatory if model flow field integrity is to be maintained.

INTRODUCTION

Programs such as the National AeroSpace Plane (NASP) are becoming increasingly dependent on having accurate flow diagnostic information in supersonic flows. Mie scattering-based laser anemometry, which utilizes a laser beam to nonintrusively probe the test medium, offers a method of measuring the flow velocity, flow angle, turbulence intensity, and shear stress around a test model under various flow conditions. Small probe volumes produce high laser power densities, allowing submicron seed particles to be used as the radiation scattering medium in the flow field as well as allowing measurements to be made very close to model surfaces (within $\approx 100~\mu m$). In support of programs such as NASP, it was desired to demonstrate the usefulness of a Mie scattering-based system in the supersonic Langley Unitary Plan Wind Tunnel. To this end, an experiment was conducted using a Laser Transit Anemometry (LTA) system to probe the boundary layer on a slender (5° half angle) cone model at Mach numbers of 2.5 and 4.5. The specific goals of this test were:

- 1. Evaluate the general performance of laser velocimetry based on Mie scattering to achieve accurate velocity measurements in the facility.
- 2. Determine the most appropriate type and location of seeding system to use with the LTA system so that particle tracking is acceptable while at the same time ensuring that the flow field under investigation is not disturbed.

SYMBOLS

The units for the physical quantities in this paper are given both in U.S. Customary Units and in the International System of Units (SI). In the text the SI units will be given first with the U.S. Customary Units following in parentheses.

b	background noise estimate in time domain correlogram
С	contrast estimate for time domain correlogram
DT	delay time between adjacent bins of time domain correlogram function, sec
E_{m}	error in mean velocity estimate, m/sec (ft/sec)
Es	error in velocity standard deviation estimate, m/sec (ft/sec)
excess	deviation of velocity space correlogram function from normal distribution
HW	half-width of self-adaptive triangular weighting filter
h	peak height of time domain correlogram function
m_k	k th moment of velocity domain correlogram function
n	number of bins in correlogram function
$r(\tau_i)$	number of events in i th channel of correlogram function in time domain
$r(\tau_i)'$	normalized number of events in ith channel of correlogram function in time domain
$r(v_i)$	number of events in ith channel of correlogram function in velocity domain
Q	total number of correlated events in time domain correlogram
q	total number of correlated events in velocity domain correlogram
r _{a,b}	individual channel data rates, pulses per sec
Re	unit Reynolds number per m (per ft)
s	beam separation, m (in.)
ST	integration time over which measurement is taken, sec

TI :	measured turbulence intensity normalized to velocity magnitude at sample point in flow field
u	boundary layer local velocity, m/sec (ft/sec)
ue	boundary layer edge velocity, m/sec (ft/sec)
. V	measured mean flow velocity magnitude at sample point, m/sec (ft/sec)
Vi	i th particle velocity measurement, m/sec (ft/sec)
$x(t_i)$	i th pulse occurring on data channel a
$y(t_i)$	i th pulse occurring on data channel b
x,y,z	model coordinates
Z _f	focal depth of system, m (in.)
β	empirical filter constant (=0.3)
Δz	height of sample volume above model surface, m (in.)
λ	laser wavelength, m (in.)
$ au_{\mathbf{i}}$	i th particle transit time between two beams, sec
ω_{o}	Gaussian waist radius of individual sample volume beam, m (in.)

DESCRIPTION OF EXPERIMENTAL APPARATUS

Laser Transit Anemometer (LTA)

In laser transit anemometry, the transit times of seed particles entrained in the flow field are measured as they cross the foci of a pair of parallel laser beams (Figure 1). The transit time of an individual particle, τ_i , in conjunction with the known beam separation, s, provides a measurement of the velocity, vi, of the particle in the plane normal to the optical axis of the system by

$$v_i = \frac{s}{\tau_i} \tag{1}$$

Assuming that the particle faithfully tracks the flow field, vi is a measure of the flow velocity in the measurement plane at the sample volume point. The beams are rotated about an axis that is equidistant from and parallel to the two beams. The ability to rotate the two beams allows for the determination of two dimensional flow angularity in the measurement plane.

The optics package used for this test was manufactured by Spectron Development Laboratories and is shown schematically in Figure 2. An argon-ion laser beam at a wavelength of 514.5 nm, circularly polarized, is sent through a Wollaston prism/lens assembly which splits the beam into a pair of parallel beams. These parallel beams are sent through a dove prism assembly mounted in a rotating ring controlled by a DC servo system to enable the two beams to be rotated precisely about a common optical axis. The beams then pass through a final field lens assembly and are focused to form two Gaussian waists at the measurement point in the flow. The system operates in a 180 degree backscatter mode with the scattered light from particles crossing the sample volume being collected around the annulus of the transmission optics. This scattered light passes through the dove prism assembly and beam stop (which blocks the focused image of each beam) and enters a fiber optic link connected to separate photomultiplier detectors. Each photomultiplier sends its output to a filter-discriminator / pulse-center detector circuit which produces a TTL level pulse whenever the signal exceeds a preset threshold level. Thus, as particles pass through the sample volume, a series of pulses are produced on two channels, each one corresponding to a separate beam.

The method used to extract transit time information for this test consists of computing the cross correlation function between the input pulse streams occurring on the two data channels. This discrete correlation function is given by

$$r(\tau_i) = \sum_{k=0}^{\infty} x(t_k) y(t_k + \tau_i)$$
 (2)

and is represented as a correlogram with 256 discrete channels or bins with each successive bin representing a delay time $\tau_i = i\Delta \tau$.

A block diagram of the LTA system used for this series of tests is shown in Figure 3. The LTA system is controlled by a MicroVAX II minicomputer system with 9 MB of main memory running at 0.9 MIPS. The computer is connected to an optical head controller which operates the image rotator as well as the discriminator / pulse-center detector circuits. The output from these detectors is sent to an event correlator which computes the cross correlation function in real time between the two channel data streams and relays this correlation function to the computer in the form of a correlogram which is a plot of the number of correlated events versus delay time, which is represented discretely in 256 channels or bins. LaRC-developed LTA data acquisition software manages all aspects of LTA system control and data acquisition as well as controlling the position of the scan rig containing the optics package. Figure 4 shows the LTA optics package mounted on the scan rig in front of test section 2 of the Langley Unitary Plan Tunnel.

Calibration of LTA System

The LTA focal point beam separation is a critical parameter, and any error in the measurement of this parameter directly affects system accuracy. The measurement of beam separation is performed by placing a $25-\mu m$ slit directly in front of a power meter detector at the focal point of the instrument. The focal length was approximately 900 mm for the optical

configuration used for this test. The slit/detector is aligned normal to the plane of the beams and scanned across both beams while monitoring laser power. A plot of laser power versus slit position allows the determination of both focal point beam diameter and beam separation. Using this technique, the beam diameter at the focal point was measured to be approximately $40\,\mu\text{m}$, while the beam separation was measured to be $1230\,\mu\text{m} \pm 10\,\mu\text{m}$. The depth of focus can be expressed using the Rayleigh range criteria [1] by:

$$z_{\rm r} \approx \pm \frac{\pi \, \omega_{\rm o}^2}{\lambda} \tag{3}$$

where z_r is the focal depth, ω_0 is the beam radius, and λ is the laser wavelength. By using equation (3), the depth of focus for the instrument is computed as \pm 2.44 mm. The active measurement volume is defined by the focal depth, beam separation, and beam waist cross section.

Wind Tunnel

The Langley Unitary Plan Wind Tunnel is a closed-circuit, continuous-flow, pressure tunnel with two 1.22- by 1.22- by 2.13-m (4- by 4- by 7-ft) test sections. The major elements of the facility are a 74.6-MW (100 000-hp) drive system, a dry air supply and evacuating system, a cooling system, and interconnecting ducting to produce the proper airflow through either of two test sections. The tunnel overall volume is 4642 m³ (163,922 ft3). The tunnel circuit is designed to operate at pressure from near-vacuum to 10 atm (1 atm = 101.3 kPa). The settling chambers provide a large volume which results in low-velocity flow and smooth transition from the circular tunnel duct work to the rectangular nozzles and test section. The settling chamber for test section 2 used in this test is a cylindrical duct 3.66 m (12.0 ft) in diameter and 7.47 m (24.5 ft) long followed by a 1.07-m (3.5-ft) long transition section from circular to 1.22- by 2.67-m (4- by 8.75-ft) rectangular. The nozzle is of the asymmetric sliding-block type such that the nozzle throat-to-test-section area ratio can be varied to provide continuous variation of Mach number. With this arrangement, the Mach number in test section 2 is variable from 2.30 to 4.63 while the unit Reynolds number per m (per ft) is variable from 1.64 x 10⁶ to 26.23 x 10⁶ (0.50 x 10⁶ to 8.00 x 10⁶). [2]

Model

The model employed for this test consisted of a 5° half angle cone forebody, a cylindrical midbody, and a 9° truncated cone afterbody. Figure 5 gives the dimensions of these individual components making up the overall model. The model was attached to the model support system in test section 2 which provided for forward and aft travel of 0.921 m (36.25 in.) in the tunnel x direction and traverse travel of ± 0.508 m (± 20 in.) in the tunnel y direction. The model was maintained at a zero degree angle-of-attack. By traversing the LTA system on its scan rig while at the same time moving the model support system, approximately 80 percent of the model's surface was accessible for surveys. Figure 6 shows the mounted model in test section 2.

Seeding System

One of the challenges to using laser anemometry in supersonic facilities is the ability to introduce the proper type of seeding material into the flow field of interest without disturbing the flow. The seed material used must be sufficiently small that particulate flow field tracking error is acceptable to the researchers. Thus the choice of a proper seeding system and seed material constituted one of the goals of this test.

Several seeding systems were tried during the course of the experiments with varying degrees of success. The first two systems utilized polystyrene latex (PSL) microspheres, measured to be $0.6 \,\mu\text{m}$ ($2.36 \, \text{x} \, 10^{-5}$ in.) in diameter, suspended in water and then suspended in pure ethanol. The PSL and liquid carriers were injected through six $0.64 \, \text{mm}$ ($0.025 \, \text{in.}$) orifices on the nose of the model, with the flow field across the orifices shearing off the liquid as it left, thus atomizing it. However, both systems were unsuccessful due to the incomplete atomization of the liquid carrier, causing the liquid to run down the sides of the model. The failure of the carrier to be atomized completely is due in part to the lower velocities present in the boundary layer of the model.

Due to these problems with using seeding materials in a liquid carrier, various dry seeding systems were investigated. The first dry seeding system consisted of injecting incense smoke through the same orifices on the model nose; however, the concentration of smoke was not sufficient to obtain good data rates with the LTA system. It was decided at this point to use dry kaolin [3] as the seeding material. A small vibrated fluidized bed (Figure 7) was used to create a cloud of dry kaolin particles which were injected into the flow via nine 1.32-mm (0.052-in.) orifices (three rows of three each) in the model nose (Figures 8 and 9). The kaolin used was Engelhard ASP 200 coated with 0.5 percent Degussa R972 to minimize caking and prevent agglomeration of particles. The nominal aerodynamic particle size of the kaolin was measured using a TSI model 3300 aerodynamic particle sizer to be $0.9 \,\mu\mathrm{m}$ (3.54 x 10^{-5} in.), with 96 percent of the particles being less than or equal to $1.7 \,\mu\text{m}$ (6.69 x 10^{-5} in.) in diameter. A pressure differential of approximately 724.0 mm Hg (14 psi) was maintained across the fluidized bed during particle injection by leaving the intake of the seeding system at atmospheric pressure and allowing the low static pressure present in the tunnel provide the necessary pressure differential. Adequate data rates (greater than 1000 pulses/sec per channel) were obtained with the LTA system using this seeding arrangement.

LTA DATA ACQUISITION

The data acquisition procedure used for this series of tests allows the determination of mean velocity, mean flow angle, and the turbulence intensity along this angle. The basic procedure, illustrated in Figure 10, consists of collecting correlograms at between five and nine discrete angular positions preselected by the system operator. From this set of correlograms a measurement of the mean flow angle is made. Next, the beam orientation is set to the mean flow angle so that an additional correlogram containing the mean velocity and turbulence intensity information can be obtained.

Angular Search Procedure

For each correlogram obtained with the system, a contrast quantity defined [4] as

$$C \equiv \frac{h - b}{\sqrt{b}} \tag{4}$$

is computed, where h is the number of events occurring in the peak channel of the correlogram and b is the average level of background noise present in the correlogram (due to spurious correlations and flare light from model surfaces). A plot is produced of these contrast quantities versus angular position through which a least squares fit of a parabolic equation is performed. The abscissa of the parabolic vertex is taken to be the mean flow angle or best angle. Figure 11 shows an example of this process. The standard deviation of the least squares fit represents the error in the mean flow angle estimate. Once the mean flow angle is determined, the two beams are repositioned at this angle and a velocity magnitude measurement is performed. Note that this procedure assumes a constant particulate concentration present in the flow. If for some reason the particulate concentration should vary during data acquisition, the correlograms used to determine the mean flow angle can be normalized by

$$r(\tau_i)' = \frac{r(\tau_i)}{(\sqrt{\frac{r_a^2 + r_b^2}{r_a^2 + r_b^2}})(ST)}$$
 (5)

where $r(\tau_i)'$ is the number of events in the ith channel of the normalized correlogram, $r(\tau_i)$ is the number of events in the ith channel of the original correlogram, r_a , r_b are the individual data rates from each channel, and ST is the integration time over which the raw correlogram is collected. The normalized functions are then used in the least squares fit routine to determine the mean flow angle.

DATA ANALYSIS

Filtering and Noise Removal

The first step in the data analysis process (Figure 12) is the removal of noise from the correlogram through filtering and background removal. To filter the correlogram, a self-adaptive digital filter incorporating a Parzen triangular weighting function is used. The average background level for the function, b, is first estimated by averaging the correlogram between the third channel and the channel nearest to 0.6 times the peak channel. The total data in the correlogram, Q, is determined by adding up all the events in all the channels of the correlogram. The half-width of the filter is then computed by

$$HW = ROUND \left(\frac{\beta(Q - nb)}{2(h - b)}\right) \tag{6}$$

where the operation ROUND makes HW an integer. β is an empirical constant which provides control over the broadening induced by the filter while n is the number of bins in the correlogram (normally 256 for this instrument). For this series of tests, β was set to 0.3. Having computed HW, the filter is now used on the original correlogram. Each data point Y(k) in the function is replaced by a triangularly weighted average over the set of data points from locations k-HW to k + HW using

$$Y(k) \leftarrow \frac{1}{HW} \sum_{j=-1 \text{IW}}^{j=+HW} \left(1 - \frac{|k-j|}{HW}\right) Y(j) \qquad (7)$$

After filtering, a new estimate of the background noise is made by computing the average channel height over the first 20 percent of the bins in the correlogram and over the last 20 percent of bins. The greater of these two averages is then taken to be the background estimate which is then subtracted from the correlogram. Note that this background removal process assumes that the nonspurious correlations in the correlogram do not occupy more than the central 50 percent of the correlogram, thereby limiting the effective turbulence intensity that can be measured. Higher turbulence intensities can be measured, but an estimate of the background noise must be made by visual inspection of the correlogram and then subtracted out.

The final step in the noise removal process is to truncate the outliers to remove any stray background signals that remain beyond the bounds of the central correlogram distribution. The truncation limits are defined to be the channels on either side of the peak of the function which first fall below the average background estimate. The correlogram is zeroed out beyond these limits. Through visual inspection these truncation limits may be changed to correct failures in the automatic truncation limit algorithm due to unusual background features which sometimes appear in correlograms. An example of the noise reduction process can be seen in Figure 13 for a typical function.

Transformation to Velocity Space

Before the mean and higher order moments are extracted from the final correlogram acquired after the angle search, the nonlinearity between the time domain (also referred to as tau space) and velocity domain (or velocity space) must be taken into account. Note that this nonlinearity is a first order inverse relationship between time and velocity as represented by equation (1):

$$v_i = \frac{s}{\tau_i}$$

Before any flow parameters are extracted, the function is transformed from tau space to

velocity space through the application of the Jacobian function [5], τ^2 /s, as shown below:

$$r(v_i) \leftarrow \frac{\tau_i^2}{s} r(\tau_i) \tag{8}$$

where $r(\tau_i)$ represents the number of events in the ith discrete channel of the tau space correlogram and $r(v_i)$ is the number of events in the ith discrete channel of the velocity space correlogram.

Extraction of Flow Parameters

For each correlogram transformed to velocity space, the total number of nonspurious correlated events in the correlogram is computed by

$$q = \sum_{i=0}^{n} r(v_i) \qquad . \tag{9}$$

The mean velocity is computed by

$$V = \frac{\sum_{i=0}^{n} v_i r(v_i)}{q} , \qquad (10)$$

while the higher-order moments are computed by

$$m_{k} = \frac{\sum_{i=0}^{n} (v_{i} - V)^{k} r(v_{i})}{q} \qquad (11)$$

The turbulence intensity along the direction of the mean flow angle is defined as

$$TJ. = \frac{\sqrt{m_2}}{V} \qquad , \tag{12}$$

and is usually expressed as a percentage. The standard error in the mean velocity measurement can be determined from

$$E m = \frac{\sqrt{m_2}}{\sqrt{q}} \qquad , \tag{13}$$

while the standard error of the standard deviation can be determined from

$$Es = (\sqrt{m_2})(\frac{1}{\sqrt{2q}})(\sqrt{1 + \frac{\text{excess}}{2}})$$
 , (14)

where excess represents the deviation of the correlogram from a normal distribution and is defined as

excess = kurtosis - 3.0 =
$$\frac{m_4}{(m_2)^2}$$
 - 3.0 (15)

Note that excess is zero for a purely normal distribution.

MEASUREMENT RESULTS

Boundary layer surveys were conducted in the vertical x-z plane at a longitudinal center line distance of 635 mm (25 in.) from the model apex. The surveys were obtained at free stream Mach numbers of 2.5 and 4.5 for Reynolds numbers of 3.3×10^6 /m and 6.6×10^6 /m (1.0×10^6 /ft and 2.0×10^6 /ft). Table I lists the LTA measured flow quantities as well as the tunnel free stream conditions for each measurement in the test. Each survey was started approximately 4-5 mm (0.157-0.197 in.) from the model surface. The two beams were lowered to the surface in discrete increments with a velocity measurement made at each step. Note that angles searches were not performed for these surveys because the flow angle was assumed to be parallel to the model surface over the range of profile measurement positions. Thus, the beams were oriented to be parallel to the model surface and only a single correlogram collected at each measurement point (see figure 10, alternate acquisition procedure). Data acquisition time was reduced by a factor of five to ten using this procedure.

The correlograms collected for the data points listed in Table I are shown in Figure 14, and several observations are in order. First, as the beams were lowered into the boundary layer, the measured turbulence intensity increased from a low value of 1.0 - 1.5 percent in the near free stream region to a high of 20.0 percent near the model surface. This effect can be seen in the turbulence intensity profiles shown in Figures 15-18. Second, the background noise rose steadily as the beams were lowered into the boundary layer. This effect, observed in similar tests, is not surprising given the increase in flare light as the model surface is approached. Finally, during the course of the surveys an occasional large surge would occur in the measured data rate and the correlogram would display unusual characteristics similar to the one shown in Figure 14 - File 21. At these times agglomerated particles were probably passing through the sample volume of the system.

An interesting effect observed from the collected correlograms is the nonsymmetric nature of the background noise present in many of the plots of Figure 14, especially for correlograms collected near the model surface. The background noise level is higher toward the longer transit time portions of the correlograms, corresponding to the lower velocity regions of the velocity space correlograms. The most likely cause of this effect is model vibrations during the tests. Such vibrations can cause integration of velocity regions in the boundary layer due to the finite size of the LTA sample volume. During the course of the tests, the model could be seen to vibrate several millimeters, especially with free stream Mach

numbers of 4.5. These vibrations can be reduced or eliminated with a stronger model support system.

VELOCITY PROFILES

Boundary layer velocity profiles corresponding to the measurements listed in Table I are shown in Figures 19-22. Also shown are theoretical predictions based on the method described in [6]. At Mach 2.5, the LTA data corresponds very nearly to the turbulent boundary layer prediction. While the LTA data at Mach 4.5 exhibits a turbulent boundary layer profile trend, it does not correspond as well with the turbulent prediction. Based on these comparisons, it is believed that the all the boundary layers are fully turbulent. The descrepancies between the LTA data and the theoretical turbulent boundary layer predictions are most likely due to the previously mentioned model vibration problems.

The development of a turbulent boundary layer condition on the model is not surprising even though the test Reynolds numbers were generally below previously measured cone transition Reynolds numbers. In [7], onset and end of transition Reynolds numbers of 11.5 x 10^6 /m and 19.7×10^6 /m (3.5 x 10^6 /ft and 6.0 x 10^6 /ft), respectively, were measured on a highly polished 10-degree cone. Several factors have likely contributed to premature transition at a lower Reynolds number in the present study. Foremost is the disturbance due to the ejection of seeding particles directly into the boundary layer. Shadowgraph photographs of the model during the course of the tests (Figure 23) show a shock at the seeding orifices on the model. Also, a slight buildup of kaolin was observed at the ejection orifices which would also contribute disturbances sufficient to create an observable weak shock even for the case of having the seeding system shut off. The disturbances caused by these two factors are sufficient to cause transition into a turbulent boundary layer. Obviously if upstream seeding were employed the direct disturbance to the boundary layer would be eliminated. However, due to the tendency of kaolin to stick on surfaces (the tunnel walls and model supports contained a dusting of kaolin, although this could have been caused by the proximity of the seeding orifices), it is still possible to disrupt the boundary layer even with upstream seeding due to the accumulation of kaolin on the model. It is also possible that surface roughness caused premature transition since the model was not highly polished.

Finally, note that in Figure 19, part of the LTA data was corrected. A height correction of 0.5 mm (1.969 x 10^{-2} jn.) was necessary to remove a bias present in the scan rig readout of position due to mechanical errors early in the test.

CONCLUSIONS

The results of the surveys indicate that indeed Mie scattering-based laser velocimetry systems are capable of yielding significant flow information which would be difficult to obtain otherwise in the Unitary Plan wind tunnel. The area of greatest education from this test concerned the choice of which seeding method to use. It is obvious that onboard seeding as

was used to conduct this test has severe consequences in terms of obtaining accurate information about laminar boundary layers on models, due to the effect of premature transition from disturbances caused by injection of particles into the boundary layer. A more appropriate seeding system would consist of upstream injection of particles into the flow, preferably in the settling chamber of the tunnel. However, given the turbulent boundary layers present on the model, the LTA system performed adequately in surveying the boundary layer and was able to make measurements to within several microns of the model surface.

REFERENCES

- 1. Kingslake, R., Optical System Design, Academic Press, 1983.
- 2. Jackson, C.M. Jr., Corlett, W.A., and Monta, W.J., Description and Calibration of the Langley Unitary Plan Wind Tunnel, NASA TP-1905, 1981.
- 3. Patrick, W.P., Overview of Solid Particle LV Seeding Techniques Used at United Technologies Research Center, Proceedings of Workshop on Wind Tunnel Seeding Systems for Laser Velocimeters, NASA Conference Publication 2393, 1985.
- 4. Smart, A.E., and Mayo, W.T., Techniques for the Measurement of Turbulence and High Speed Flow with a Laser Transit Anemometer, AEDC Report TR-84-13, Volume 1, 1984.
- 5. Abbiss, J.B., Laser Transit Anemometer Software Development Program, NASA Contractor Report 4225, 1989.
- 6. Harris, J.E., and Blanchard, D.K., Computer Program for Solving Laminar, Transitional, or Turbulent Compressible Boundary Layer Equations for Two-Dimensional and Axisymmetric Flow, NASA TM-83207, 1982.
- 7. Fisher, D.F., and Dougherty, N.S., Flight and Wind Tunnel Correlation of Boundary Layer Transition on the AEDC Transition Cone, AGARD-CP-339, 1982.

TABLE I: LASER TRANSIT ANEMOMETER / VELOCITY PROFILE DATA


Es	(m/s)	1.186	1.245	1.133	0.624	0.944	0.684	0.664	0.577	0.758	0.594	0.511	0.485	0.816	1.278	0.716	0.602	0.534	0.643	0.571	0.063	0.241	0.357	0.438	0.409	0.608	0.380	0.332	0.327	0.536	0.498	0.362	0.381	0.366	0.375	0.482	0.464	0.318	0.418	0.671
Em	(m/s)	2.076	2.328	1.705	1.089	1.526	1.141	0.954	0.959	1.228	1.033	0.835	0.955	1.479	2.368	1.397	0.941	0.835	0.898	0.941	0.103	0.407	0.537	0.636	0.591	1.040	0.594	0.522	0.529	0.814	0.767	0.570	0.680	0.594	0.625	928.0	0.753	0.634	0.748	1.345
F	(%)	2.278	2.940	3.184	4.539	6.286	6.573	7.455	7.308	9.434	8.572	11.589	10.941	19.164	25.300	26.707	3.038	2.718	3.853	4.102	4.329	4.909	5.401	5.480	5.582	5.811	6.337	6.940	7.608	8.333	9.480	10.746	9.184	11.252	15.547	13.457	17.396	17.773	18.140	19.675
Λ	(s/m)	568.752	562.752	545.719	533.416	516.988	483.248	477.549	462.045	437,495	425.397	401.140	374.001	340.092	313.691	251.422	565.181	561.856	552.421	549.710	541.965	532.304	531.204	525.866	516.601	504.984	497.251	482,181	468.752	455.889	436.931	422.100	410.547	398.133	370.349	370.281	355.037	297.941	270.426	278.560
ST	(sec)	15.0	15.0	15.0	15.0	15.0	15.0	45.0	45.0	45.0	0.09	120.0	0.06	120.0	120.0	120.0	30.0	30.0	45.0	45.0	45.0	15.0	15.0	15.0	15.0	15.0	30.0	30.0	30.0	30.0	45.0	45.0	45.0	45.0	45.0	45.0	45.0	45.0	45.0	45.0
DT	(nSec)	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	40.0	40.0	80.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	30.0	30.0	30.0	50.0	50.0	50.0
ΔΖ	(mm)	3.810	3.302	2.794	2.286	1.778	1.524	1.270	1.016	0.762	0.508	0.381	0.254	0.203	0.152	0.101	3.810	3.556	3.302	3.048	2.794	2.540	2.540	2.286	2.032	1.778	1.524	1.270	1.016	0.762	0.508	0.381	0.254	0.203	0.152	0.101	0.076	0.051	0.025	0000
Station	(mm)	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0
Re	(x 10 ⁶)	3.281	3.281	3.281	3.281	3.281	3.281	3.281	3.281	3.281	3.281	3.281	3.281	3.281	3.281	3.281	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562
Mach #	100000	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5	2.5
File#		1	2	3	4	30 S 1 S 1 S 1 S	9	7	∞	6	10		12	13	14	1.5	16	11	18	19	82	21	2	23	24	25	26	27	28	23	æ	31	32	33	34	35	36	37	8 8	39

TABLE I: LASER TRANSIT ANEMOMETER / VELOCITY PROFILE DATA (continued)

44 (5.10 ⁴) (mm) (mSec) (sec) (ms)	File #	Mach #	Re	Station	∇Z	DT	ST	Λ	E	Em	Es
45 3.281 (35.0) 2.5197 15.0 30.0 11.25.1 10.995 0.605 4.5 3.281 (35.0) 1.249 15.0 30.0 116.397 1.030 4.5 3.281 (35.0) 1.679 1.50 30.0 17.106 1.39 0.301 4.5 3.281 (35.0) 6635 1.60 45.0 17.106 1.39 0.301 4.5 3.281 (35.0) 5635 1.50 45.0 0.893 2.41 0.302 4.5 3.281 (35.0) 45.0 669.977 2.899 0.50 4.5 3.281 (35.0) 45.0 669.977 2.899 0.70 4.5 3.281 (35.0) 4.50 45.0 669.977 2.899 0.70 4.5 3.281 (35.0) 4.50 45.0 669.977 2.890 0.70 4.5 3.281 (35.0) 4.50 45.0 669.977 2.890 0.80			$(x 10^6)$	(mm)	(mm)	(nSec)	(sec)	(m/s)	(%)	(m/s)	(m/s)
45 3281 6350 1247 150 300 716.339 1495 0.350 45 3281 6350 10.797 150 300 716.397 1295 0.331 45 3281 6350 1671 150 450 717.106 1399 0.284 45 3281 6350 5639 150 450 788.50 1887 0.284 45 3281 6350 511 150 450 788.50 1897 0.284 45 3281 6350 511 150 450 689533 2417 0.524 45 3281 6350 5131 150 450 689533 2417 150 450 68870 0.524 45 3281 6350 3435 150 450 68770 0.524 0.524 45 3281 6350 3435 150 450 66730 4488 0.524 45	40	4.5	3.281	635.0	25.197	15.0	45.0	717.521	0.995	0.626	0.397
45 3381 6350 10.719 15.0 45.0 1710 128 0.30 45 3.281 6350 16.75 15.0 45.0 717106 1.29 0.264 45 3.281 6350 2.653 15.0 45.0 713189 1.47 0.234 45 3.281 6350 5.631 15.0 45.0 689537 2.47 0.294 45 3.281 6350 4.115 15.0 45.0 689207 2.89 0.88 45 3.281 6350 4.115 15.0 45.0 660.90 0.75 0.89 45 3.281 6350 3.437 15.0 45.0 660.70 3.76 0.75 45 3.281 6350 2.897 15.0 45.0 660.70 3.76 0.75 45 3.281 6350 2.37 15.0 45.0 660.70 3.76 0.75 45 3.281 6350	41	4.5	3.281	635.0	12.497	15.0	30.0	716.339	1.409	0.350	0.285
45 3.281 655.0 8.687 15.0 45.0 17106 1339 0.264 45 3.281 655.0 5.639 15.0 45.0 713859 1497 0.224 45 3.281 655.0 5.639 15.0 45.0 605.973 2.847 0.524 45 3.281 655.0 4.115 15.0 45.0 605.973 2.847 0.524 45 3.281 655.0 4.115 15.0 45.0 605.973 2.847 0.657 45 3.281 655.0 3.415 15.0 45.0 668.20 2.977 0.657 45 3.281 653.0 2.845 15.0 30.0 662.146 3.02 0.657 45 3.281 653.0 2.895 15.0 30.0 662.146 0.50 45 3.281 635.0 2.891 15.0 30.0 662.146 2.97 0.60 45 3.281 635.0<	42	4.5	3.281	635.0	10.719	15.0	30.0	716.967	1.295	0.301	0.222
45 3.281 655.0 5.671 15.0 45.0 1.887 1.897 0.294 45 3.281 655.0 5.639 15.0 45.0 609.937 2.47 0.524 45 3.281 655.0 5.639 15.0 45.0 609.937 2.47 0.524 45 3.281 655.0 4.15 15.0 45.0 609.937 2.47 0.524 45 3.281 655.0 3.18 15.0 60.0 60.0 0.527 0.524 45 3.281 655.0 3.55 15.0 45.0 66.820 3.68 0.57 45 3.281 655.0 3.55 15.0 30.0 672.163 3.67 0.52 45 3.281 655.0 2.591 15.0 30.0 664.794 6.52 12.0 45 3.281 635.0 2.892 15.0 30.0 664.794 6.52 12.0 45 3.281	43	4.5	3.281	635.0	8.687	15.0	45.0	717.106	1.339	0.263	0.211
45 3281 6350 5.639 15.0 45.0 73845 1881 0.379 45 3.281 635.0 5.639 15.0 45.0 699.33 2417 0.524 45 3.281 635.0 4.13 15.0 45.0 688.230 2.917 0.659 45 3.281 635.0 4.13 15.0 45.0 688.230 2.917 0.659 45 3.281 635.0 3.059 15.0 45.0 688.230 2.917 0.659 45 3.281 635.0 3.059 15.0 45.0 66.05 3.73 0.609 45 3.281 635.0 2.845 15.0 30.0 672.13 3.68 0.77 45 3.281 635.0 2.845 15.0 45.0 66.20 45.1 10.77 45 3.281 635.0 1.829 15.0 45.0 66.2149 45.2 1.219 45 3.281	4	4.5	3.281	635.0	7.671	15.0	45.0	713.839	1.497	0.294	0.230
45 3.281 635.0 5.639 15.0 45.0 699.337 2.447 0.524 45 3.281 635.0 4.623 15.0 45.0 699.977 2.849 0.537 45 3.281 635.0 4.673 15.0 45.0 699.977 2.840 0.537 45 3.281 635.0 3.677 15.0 45.0 678.20 3.687 0.767 45 3.281 635.0 3.187 15.0 45.0 672.76 3.786 0.587 45 3.281 635.0 2.891 15.0 45.0 664.95 4.18 11.0 45 3.281 635.0 2.891 15.0 45.0 664.95 4.48 1.12 45 3.281 635.0 2.891 15.0 45.0 664.95 4.48 1.12 45 3.281 635.0 1.27 1.50 45.0 662.13 2.52 1.21 45 3.281	45	4.5	3.281	635.0	6.655	15.0	45.0	708.565	1.881	0.379	0.300
4.5 3.281 635.0 5.131 15.0 45.0 695.977 2.889 0.587 4.5 3.281 635.0 4.115 15.0 45.0 688.20 2.917 0.659 4.5 3.281 635.0 3.415 15.0 45.0 676.592 3.089 0.776 4.5 3.281 635.0 3.089 15.0 30.0 672.133 3.686 0.776 4.5 3.281 635.0 2.894 15.0 30.0 672.133 3.685 0.776 4.5 3.281 635.0 2.894 15.0 30.0 672.134 4.53 0.890 4.5 3.281 635.0 2.894 15.0 30.0 65.2134 6.221 0.207 4.5 3.281 635.0 1.32 1.50 30.0 65.2134 6.22 1.127 4.5 3.281 635.0 1.32 1.50 30.0 65.2134 6.22 1.127 4.5	9	4.5	3.281	635.0	5.639	15.0	45.0	699.933	2.417	0.524	0.350
45 3281 6350 4653 150 450 601445 3012 0.609 45 3281 6350 3407 150 450 666.822 2917 0.637 45 3281 6350 3.697 150 450 676.822 3.689 0.776 45 3281 6350 3.093 150 300 672.13 3.655 0.890 45 3281 6350 2.845 150 300 666.455 448 1.10 45 3281 6350 2.891 150 300 664.794 6521 0.890 45 3281 6350 1.875 150 300 624.794 6521 1090 45 3281 6350 1.875 150 300 624.794 6521 1080 45 3281 6350 1.875 150 300 624.794 6521 1080 45 3281 6350 1.87	47	4.5	3.281	635.0	5.131	15.0	45.0	695.977	2.869	0.587	0.493
4.5 3.281 635.0 4115 15.0 45.0 676.592 3.288 0.277 4.5 3.281 635.0 3.637 15.0 30.0 676.592 3.688 0.776 4.5 3.281 635.0 3.533 15.0 30.0 6672.763 3.788 0.870 4.5 3.281 635.0 2.845 15.0 30.0 660.700 4661 0.999 4.5 3.281 635.0 2.891 15.0 30.0 660.700 4661 0.999 4.5 3.281 635.0 2.837 15.0 30.0 662.134 5.52 1.129 4.5 3.281 635.0 1.829 15.0 30.0 662.00 5.81 0.587 4.5 3.281 635.0 1.829 15.0 30.0 662.00 5.81 0.588 0.588 4.5 3.281 635.0 1.829 1.50 30.0 662.04 44.418 1.175	48	4.5	3.281	635.0	4.623	15.0	45.0	691.445	3.012	609.0	0.494
45 3281 6350 3607 15.0 45.0 676.592 3.686 0.776 45 3281 635.0 3.353 15.0 30.0 672.730 3.736 0.880 45 3.281 635.0 2.845 15.0 30.0 666.495 4418 1.127 45 3.281 635.0 2.891 15.0 45.0 660.700 4461 0.590 45 3.281 635.0 2.891 15.0 45.0 662.134 5.22 1.219 45 3.281 635.0 2.891 15.0 30.0 672.134 6.51 0.507 45 3.281 635.0 1.829 15.0 30.0 672.134 6.51 0.508 45 3.281 635.0 1.871 15.0 30.0 672.05 8.986 0.508 45 3.281 635.0 1.872 1.80 30.0 672.05 1.173 45 3.281 635.0 </td <td>49</td> <td>4.5</td> <td>3.281</td> <td>635.0</td> <td>4.115</td> <td>15.0</td> <td>45.0</td> <td>688.220</td> <td>2.917</td> <td>0.637</td> <td>0.404</td>	49	4.5	3.281	635.0	4.115	15.0	45.0	688.220	2.917	0.637	0.404
4.5 3.281 635.0 3.333 15.0 30.0 672.760 3.756 0.890 4.5 3.281 635.0 2.892 15.0 30.0 662.133 3.655 0.890 4.5 3.281 635.0 2.891 15.0 45.0 660.700 46.1 10.90 4.5 3.281 635.0 2.837 15.0 45.0 662.134 5.25 1.219 4.5 3.281 635.0 1.829 15.0 30.0 662.134 6.52 1.219 4.5 3.281 635.0 1.673 30.0 662.134 6.52 1.219 4.5 3.281 635.0 1.677 3.0 662.106 6.50 0.50 4.5 3.281 635.0 1.677 1.50 30.0 662.134 6.58 0.50 4.5 3.281 635.0 1.677 1.50 30.0 50.048 0.50 4.5 3.281 635.0 0.305	50	4.5	3.281	635.0	3.607	15.0	45.0	676.592	3.686	0.776	0.493
45 3281 655.0 3099 15.0 30.0 672.153 3.655 0.969 45 3.281 635.0 2.845 15.0 45.0 660.00 44.8 1.127 45 3.281 635.0 2.337 15.0 45.0 660.10 5.22 1.129 45 3.281 635.0 2.033 15.0 30.0 652.134 5.22 1.219 45 3.281 635.0 1.829 15.0 30.0 652.134 5.22 1.219 45 3.281 635.0 1.829 15.0 30.0 652.05 7.886 0.508 45 3.281 635.0 1.067 15.0 30.0 50.0480 9.678 0.718 45 3.281 635.0 0.815 2.0 30.0 50.0480 0.678 45 3.281 635.0 0.133 2.0 30.0 50.0480 0.678 45 3.281 635.0 0.133 </th <th>51</th> <th>4.5</th> <th>3.281</th> <th>635.0</th> <th>3.353</th> <th>15.0</th> <th>30.0</th> <th>672.760</th> <th>3.736</th> <th>0.890</th> <th>0.650</th>	51	4.5	3.281	635.0	3.353	15.0	30.0	672.760	3.736	0.890	0.650
45 3.281 655.0 2.845 15.0 45.0 664.95 4418 1127 45 3.281 653.0 2.531 15.0 45.0 652.134 652.1 6.990 45 3.281 635.0 2.283 15.0 30.0 654.794 6.521 1.219 45 3.281 635.0 1.822 15.0 30.0 652.05 1.219 45 3.281 635.0 1.822 15.0 30.0 652.06 6.528 0.508 45 3.281 635.0 1.067 15.0 30.0 602.05 3.88 0.618 45 3.281 635.0 1.067 15.0 30.0 50.480 9.678 0.738 45 3.281 635.0 0.183 2.0 30.0 52.048 9.678 0.738 45 3.281 635.0 0.178 2.0 45.0 45.2 1.051 45 3.281 635.0 0.182	52	4.5	3.281	635.0	3.099	15.0	30.0	672.153	3.655	6960	0,621
4.5 3.281 635.0 2.591 15.0 45.0 660.700 4661 0.990 4.5 3.281 635.0 2.337 15.0 45.0 652.134 5.252 1.219 4.5 3.281 635.0 1.829 15.0 30.0 652.134 5.252 1.219 4.5 3.281 635.0 1.829 15.0 30.0 616.266 2488 0.506 4.5 3.281 635.0 1.071 15.0 30.0 60.207 8.986 0.618 4.5 3.281 635.0 1.067 15.0 30.0 570.035 11.010 0.823 4.5 3.281 635.0 0.813 15.0 30.0 570.035 11.010 0.823 4.5 3.281 635.0 0.178 2.0 30.0 455.26 11.544 1.053 4.5 3.281 635.0 0.178 2.0 45.0 45.22 1.249 1.053 4.5	53	4.5	3.281	635.0	2.845	15.0	30.0	666.495	4.418	1.127	0.728
4.5 3.281 655.0 2.337 15.0 45.0 652.134 5.252 1.219 4.5 3.281 635.0 2.083 15.0 30.0 654.794 6.521 0.507 4.5 3.281 635.0 1.575 15.0 30.0 662.66 7.836 0.507 4.5 3.281 635.0 1.321 15.0 30.0 662.66 7.836 0.678 4.5 3.281 635.0 1.821 15.0 30.0 662.67 8.986 0.678 4.5 3.281 635.0 0.835 20.0 30.0 570.035 1.1010 0.828 4.5 3.281 635.0 0.305 20.0 45.0 485.752 1.154 1.063 4.5 3.281 635.0 0.305 20.0 45.0 485.752 1.154 1.063 4.5 3.281 635.0 0.138 25.0 90.0 455.05 1.173 4.5 6.562	54	4.5	3.281	635.0	2.591	15.0	45.0	660.700	4.661	066.0	9990
45 3.281 635.0 2.083 15.0 30.0 654.794 6.521 0.507 45 3.281 635.0 1.829 15.0 30.0 6616.26 6.498 0.508 45 3.281 635.0 1.321 15.0 30.0 602.075 8.986 0.618 45 3.281 635.0 1.321 15.0 30.0 602.075 8.986 0.618 45 3.281 635.0 1.067 15.0 30.0 570.035 1.010 0.618 45 3.281 635.0 0.359 20.0 485.752 11.544 1.063 45 3.281 635.0 0.178 25.0 90.0 485.752 11.544 1.063 45 3.281 635.0 0.178 25.0 90.0 485.752 1.7544 1.063 45 6.562 635.0 10.1829 15.0 30.0 720.444 1.063 45 6.562 635.0	55	4.5	3.281	635.0	2.337	15.0	45.0	652,134	5.252	1.219	0.971
45 3.281 635.0 1.829 15.0 30.0 627.016 6.498 0.508 45 3.281 635.0 1.575 15.0 30.0 616.266 7.836 0.616 45 3.281 635.0 1.671 15.0 30.0 500.267 9.678 0.616 45 3.281 635.0 0.813 15.0 30.0 570.035 11.010 0.828 45 3.281 635.0 0.813 15.0 30.0 570.035 11.010 0.828 45 3.281 635.0 0.839 20.0 520.08 485.752 11.64 10.63 45 3.281 635.0 0.133 25.0 90.0 485.752 11.64 10.63 45 3.281 635.0 10.133 25.0 90.0 455.19 27.48 1.063 45 6.562 635.0 10.1829 15.0 30.0 72.644 1.163 0.248 45	56	4.5	3.281	635.0	2.083	15.0	30.0	634.794	6.521	0.507	0.374
45 3.281 635.0 1.575 15.0 30.0 616.266 7.836 0.616 45 3.281 635.0 1.321 15.0 30.0 550.480 9.678 0.678 45 3.281 635.0 0.813 15.0 30.0 550.480 9.678 103.2 45 3.281 635.0 0.835 20.0 45.0 45.0 17.54 1.03.2 45 3.281 635.0 0.305 20.0 45.0 45.0 17.54 1.063 45 3.281 635.0 0.138 25.0 90.0 45.0 45.1 27.48 1.063 45 3.281 635.0 10.182 25.0 90.0 75.256 21.879 1.735 45 6.562 635.0 10.1829 15.0 30.0 75.463 1.261 30.0 45 6.562 635.0 10.1829 15.0 30.0 724.081 1.281 0.204 <td< td=""><td>57</td><td>4.5</td><td>3.281</td><td>635.0</td><td>1.829</td><td>15.0</td><td>30.0</td><td>627.016</td><td>6.498</td><td>0.508</td><td>0.340</td></td<>	57	4.5	3.281	635.0	1.829	15.0	30.0	627.016	6.498	0.508	0.340
4.5 3.281 635.0 1.321 15.0 30.0 602.075 8.986 0.678 4.5 3.281 635.0 1.067 15.0 30.0 570.035 11.010 0.828 4.5 3.281 635.0 0.539 20.0 30.0 570.035 1.051 0.828 4.5 3.281 635.0 0.559 20.0 30.0 528.03 1.051 0.828 4.5 3.281 635.0 0.305 20.0 45.0 485.725 175.44 1.063 4.5 3.281 635.0 0.173 25.0 90.0 455.26 21879 1.735 4.5 6.562 635.0 127.229 15.0 30.0 724.64 1.168 0.204 4.5 6.562 635.0 76.429 15.0 30.0 724.081 1.281 0.309 4.5 6.562 635.0 51.029 15.0 30.0 724.081 1.26 0.039 4.5 </th <th>58</th> <th>4.5</th> <th>3.281</th> <th>635.0</th> <th>1.575</th> <th>15.0</th> <th>30.0</th> <th>616.266</th> <th>7.836</th> <th>0.616</th> <th>0.467</th>	58	4.5	3.281	635.0	1.575	15.0	30.0	616.266	7.836	0.616	0.467
4.5 3.281 655.0 1,067 15.0 30.0 590.480 9,678 0,735 4.5 3.281 635.0 0,813 15.0 30.0 570.035 11010 0,828 4.5 3.281 635.0 0,385 20.0 30.0 570.035 11010 0,828 4.5 3.281 635.0 0,385 20.0 45.0 485.752 17.544 1,063 4.5 3.281 635.0 0,178 25.0 90.0 455.256 21.849 1,735 4.5 6.562 635.0 127.229 15.0 30.0 72.444 1,168 0.204 4.5 6.562 635.0 15.0 30.0 72.448 1,168 0.730 4.5 6.562 635.0 15.0 30.0 724.681 1,359 0.730 4.5 6.562 635.0 15.0 30.0 724.081 1,364 0.488 4.5 6.562 635.0 25.0 <td>59</td> <td>4.5</td> <td>3.281</td> <td>635.0</td> <td>1.321</td> <td>15.0</td> <td>30.0</td> <td>602.075</td> <td>8.986</td> <td>0.678</td> <td>0.512</td>	59	4.5	3.281	635.0	1.321	15.0	30.0	602.075	8.986	0.678	0.512
4.5 3.281 635.0 0.813 15.0 30.0 570.035 11.010 0.828 4.5 3.281 635.0 0.559 20.0 30.0 528.058 15.686 1.051 4.5 3.281 635.0 0.305 20.0 45.0 485.722 17.544 1.063 4.5 3.281 635.0 0.178 25.0 90.0 455.26 21.879 1.735 4.5 6.562 635.0 127.229 15.0 30.0 728.44 1.063 0.204 4.5 6.562 635.0 15.0 30.0 728.44 1.168 0.204 4.5 6.562 635.0 15.0 30.0 728.44 1.168 0.204 4.5 6.562 635.0 51.029 15.0 30.0 720.61 1.261 0.669 4.5 6.562 635.0 38.329 15.0 30.0 720.64 1.164 0.669 4.5 6.562 635.0 </td <td>8</td> <td>4.5</td> <td>3.281</td> <td>635.0</td> <td>1.067</td> <td>15.0</td> <td>30.0</td> <td>590.480</td> <td>879.6</td> <td>0.735</td> <td>0.514</td>	8	4.5	3.281	635.0	1.067	15.0	30.0	590.480	879.6	0.735	0.514
45 3.281 635.0 0.559 20.0 30.0 528.058 15.686 1.051 45 3.281 635.0 0.305 20.0 45.0 485.752 17.544 1.063 45 3.281 635.0 0.178 25.0 90.0 455.226 21.879 1.735 45 6.562 635.0 127.229 15.0 30.0 728.464 1.168 0.204 45 6.562 635.0 101829 15.0 30.0 724.631 1.261 0.653 45 6.562 635.0 51.029 15.0 30.0 724.631 1.261 0.653 45 6.562 635.0 51.029 15.0 30.0 724.081 1.079 0.181 45 6.562 635.0 33.249 15.0 30.0 720.644 1.164 0.669 45 6.562 635.0 25.527 15.0 30.0 720.667 1.131 0.448 45 <td>61</td> <td>4.5</td> <td>3.281</td> <td>635.0</td> <td>0.813</td> <td>15.0</td> <td>30.0</td> <td>570.035</td> <td>11,010</td> <td>0.828</td> <td>0.522</td>	61	4.5	3.281	635.0	0.813	15.0	30.0	570.035	11,010	0.828	0.522
45 3281 6350 0.305 20.0 45.0 485.752 17544 1.063 45 3281 635.0 0.178 25.0 90.0 455.26 21.879 1.735 45 3281 635.0 0.173 25.0 90.0 455.26 21.879 1.735 45 6562 635.0 127.229 15.0 30.0 728.464 1.168 0.204 45 6562 635.0 101.829 15.0 30.0 728.464 1.168 0.204 45 6562 635.0 101.829 15.0 30.0 726.33 1.359 0.730 45 6562 635.0 38.329 15.0 30.0 726.44 1.164 0.669 45 6562 635.0 33.249 15.0 30.0 720.644 1.164 0.689 45 6562 635.0 25.527 15.0 30.0 720.757 1.135 0.1348 45 <t< td=""><td>62</td><td>4.5</td><td>3.281</td><td>635.0</td><td>0.559</td><td>20.0</td><td>30.0</td><td>528.058</td><td>15.686</td><td>1.051</td><td>0.654</td></t<>	62	4.5	3.281	635.0	0.559	20.0	30.0	528.058	15.686	1.051	0.654
45 3281 635.0 0.178 25.0 90.0 455.226 21.879 1.735 45 3.281 635.0 0.153 25.0 90.0 492.119 22.748 3.487 45 6.562 635.0 127.229 15.0 30.0 728.464 1.168 0.204 45 6.562 635.0 101.829 15.0 30.0 728.464 1.168 0.204 45 6.562 635.0 15.0 30.0 724.681 1.281 0.653 45 6.562 635.0 31.229 15.0 30.0 724.081 1.281 0.488 45 6.562 635.0 33.249 15.0 30.0 720.644 1.164 0.669 45 6.562 635.0 25.629 15.0 30.0 720.644 1.164 0.648 45 6.562 635.0 25.527 15.0 30.0 720.644 1.134 0.134 45 6.562	63	4.5	3.281	635.0	0.305	20.0	45.0	485.752	17.544	1.063	0.634
4.5 3.281 635.0 0.153 25.0 90.0 492.119 22.748 3.487 4.5 6.562 635.0 127.229 15.0 30.0 728.464 1.168 0.204 4.5 6.562 635.0 101.829 15.0 30.0 724.633 1.359 0.730 4.5 6.562 635.0 76.429 15.0 30.0 724.681 1.359 0.730 4.5 6.562 635.0 51029 15.0 30.0 724.081 1.281 0.399 4.5 6.562 635.0 33.249 15.0 30.0 724.081 1.164 0.669 4.5 6.562 635.0 33.249 15.0 30.0 720.644 1.164 0.669 4.5 6.562 635.0 25.629 15.0 30.0 720.647 1.164 0.669 4.5 6.562 635.0 25.27 15.0 30.0 720.647 1.164 0.154 <t< th=""><th>2</th><th>4.5</th><th>3.281</th><th>635.0</th><th>0.178</th><th>25.0</th><th>0.06</th><th>455.226</th><th>21.879</th><th>1.735</th><th>1.039</th></t<>	2	4.5	3.281	635.0	0.178	25.0	0.06	455.226	21.879	1.735	1.039
45 6.562 635.0 127.229 15.0 30.0 728.464 1.168 0.204 45 6.562 635.0 101.829 15.0 30.0 731.463 1.359 0.730 45 6.562 635.0 76.429 15.0 30.0 725.633 1.261 0.653 45 6.562 635.0 31.029 15.0 30.0 724.081 1.281 0.399 45 6.562 635.0 38.329 15.0 30.0 720.012 1.079 0.181 45 6.562 635.0 25.629 15.0 30.0 720.607 1.164 0.669 45 6.562 635.0 25.527 15.0 30.0 720.607 1.131 0.448 45 6.562 635.0 25.527 15.0 30.0 720.607 1.134 0.134 45 6.562 635.0 12.929 15.0 30.0 720.537 1.246 0.134 45 <th>65</th> <th>4.5</th> <th>3.281</th> <th>635.0</th> <th>0.153</th> <th>25.0</th> <th>0.06</th> <th>492.119</th> <th>22.748</th> <th>3.487</th> <th>1.756</th>	65	4.5	3.281	635.0	0.153	25.0	0.06	492.119	22.748	3.487	1.756
4.5 6.562 6350 101.829 15.0 30.0 731.463 1.359 0.730 4.5 6.562 635.0 76429 15.0 30.0 725.633 1.261 0.653 4.5 6.562 635.0 51.029 15.0 30.0 724.081 1.281 0.399 4.5 6.562 635.0 38.329 15.0 30.0 720.012 1.079 0.181 4.5 6.562 635.0 25.629 15.0 30.0 720.644 1.164 0.669 4.5 6.562 635.0 25.527 15.0 30.0 720.647 1.131 0.448 4.5 6.562 635.0 25.527 15.0 30.0 720.537 1.246 0.134 4.5 6.562 635.0 12.929 15.0 30.0 719.886 1.115 0.093 4.5 6.562 635.0 12.827 15.0 30.0 700.248 2.647 0.513 <td< th=""><th>99</th><th>4.5</th><th>6.562</th><th>635.0</th><th>127.229</th><th>15.0</th><th>30.0</th><th>728.464</th><th>1.168</th><th>0.204</th><th>0.178</th></td<>	99	4.5	6.562	635.0	127.229	15.0	30.0	728.464	1.168	0.204	0.178
4.5 6.562 6350 76.429 15.0 30.0 725.633 1.261 0.653 4.5 6.562 635.0 51.029 15.0 30.0 724.081 1.281 0.399 4.5 6.562 635.0 33.249 15.0 30.0 720.644 1.164 0.669 4.5 6.562 635.0 25.629 15.0 30.0 720.644 1.164 0.669 4.5 6.562 635.0 25.629 15.0 30.0 720.647 1.131 0.448 4.5 6.562 635.0 25.527 15.0 30.0 719.76 1.246 0.134 4.5 6.562 635.0 12.929 15.0 30.0 710.575 1.115 0.134 4.5 6.562 635.0 12.929 15.0 30.0 710.867 1.115 0.093 4.5 6.562 635.0 12.827 15.0 30.0 700.248 2.647 0.513	29	4.5	6.562	635.0	101.829	15.0	30.0	731.463	1.359	0.730	0.483
4.5 6.562 6350 51.029 15.0 30.0 724.081 1.281 0.399 4.5 6.562 635.0 38.329 15.0 30.0 720.044 1.164 0.669 4.5 6.562 635.0 25.629 15.0 30.0 720.644 1.164 0.669 4.5 6.562 635.0 25.629 15.0 30.0 720.644 1.131 0.448 4.5 6.562 635.0 25.527 15.0 30.0 719.76 1.260 0.155 4.5 6.562 635.0 15.0 30.0 719.76 1.246 0.134 4.5 6.562 635.0 12.929 15.0 30.0 719.86 1.115 0.166 4.5 6.562 635.0 12.827 15.0 30.0 719.886 1.115 0.093 4.5 6.562 635.0 12.827 15.0 30.0 691.828 2.647 0.533 4.5 6.562	89	4.5	6.562	635.0	76.429	15.0	30.0	725.633	1.261	0.653	0.474
4.5 6.562 6350 38.329 15.0 30.0 720.012 1.079 0.181 4.5 6.562 635.0 23.249 15.0 30.0 720.644 1.164 0.669 4.5 6.562 635.0 25.629 15.0 30.0 719.76 1.260 0.155 4.5 6.562 635.0 25.197 15.0 30.0 719.76 1.260 0.155 4.5 6.562 635.0 12.929 15.0 30.0 718.677 1.115 0.166 4.5 6.562 635.0 12.827 15.0 30.0 719.867 1.115 0.093 4.5 6.562 635.0 12.827 15.0 30.0 719.886 1.192 0.093 4.5 6.562 635.0 3.937 15.0 30.0 691.828 3.272 0.354	69	4.5	6.562	635.0	51.029	15.0	30.0	724.081	1.281	0.399	0.322
4.5 6.562 6350 33.249 15.0 30.0 720.644 1.164 0.669 4.5 6.562 635.0 25.629 15.0 15.0 720.607 1.131 0.448 4.5 6.562 635.0 25.527 15.0 30.0 719.776 1.260 0.155 4.5 6.562 635.0 25.197 15.0 30.0 718.677 1.115 0.134 4.5 6.562 635.0 12.929 15.0 30.0 718.677 1.115 0.093 4.5 6.562 635.0 12.827 15.0 30.0 719.886 1.192 0.093 4.5 6.562 635.0 3.937 15.0 30.0 691.828 3.272 0.354	70	4.5	6.562	635.0	38.329	15.0	30.0	720.012	1.079	0.181	0.154
4.5 6.562 6350 25.629 15.0 15.0 720.607 1.131 0.448 4.5 6.562 635.0 25.527 15.0 30.0 719.776 1.260 0.155 4.5 6.562 635.0 12.929 15.0 30.0 718.677 1.115 0.166 4.5 6.562 635.0 12.827 15.0 30.0 719.886 1.115 0.093 4.5 6.562 635.0 4.039 15.0 30.0 700.248 2.647 0.513 4.5 6.562 635.0 3.937 15.0 30.0 691.828 3.272 0.354	71	4.5	6.562	635.0	33.249	15.0	30.0	720.644	1.164	0.669	0.392
4.5 6.562 6350 25.527 15.0 30.0 719.776 1.260 0.155 4.5 6.562 635.0 25.197 15.0 30.0 720.537 1.246 0.134 4.5 6.562 635.0 12.929 15.0 30.0 718.677 1.115 0.166 4.5 6.562 635.0 12.827 15.0 30.0 719.886 1.192 0.093 4.5 6.562 635.0 4.039 15.0 30.0 691.828 2.647 0.513 4.5 6.562 635.0 3.937 15.0 30.0 691.828 3.272 0.354	72	4.5	6.562	635.0	25.629	15.0	15.0	720.607	1.131	0.448	0.354
4.5 6.562 6350 25.197 15.0 30.0 720.537 1.246 0.134 4.5 6.562 635.0 12.929 15.0 30.0 718.677 1.115 0.166 4.5 6.562 635.0 12.827 15.0 30.0 719.886 1.192 0.093 4.5 6.562 635.0 4.039 15.0 30.0 691.828 2.647 0.513 4.5 6.562 635.0 3.937 15.0 30.0 691.828 3.272 0.354	73	4.5	6.562	635.0	25.527	15.0	30.0	719.776	1.260	0.155	0.132
4.5 6.562 6350 12.929 15.0 30.0 718.677 1.115 0.166 4.5 6.562 635.0 12.827 15.0 30.0 719.886 1.192 0.093 4.5 6.562 635.0 4.039 15.0 30.0 700.248 2.647 0.513 4.5 6.562 635.0 3.937 15.0 30.0 691.828 3.272 0.354	74	4.5	6.562	635.0	25.197	15.0	30.0	720.537	1.246	0.134	0.119
4.5 6.562 635.0 12.827 15.0 30.0 719.886 1.192 0.093 4.5 6.562 635.0 4.039 15.0 30.0 700.248 2.647 0.513 4.5 6.562 635.0 3.937 15.0 30.0 691.828 3.272 0.354	75	4.5	6.562	635.0	12.929	15.0	30.0	718.677	1.115	0.166	0.140
4.5 6.562 635.0 4.039 15.0 30.0 700.248 2.647 0.513 4.5 6.562 635.0 3.937 15.0 30.0 691.828 3.272 0.354	76	4.5	6.562	635.0	12.827	15.0	30.0	719.886	1.192	0.093	0.080
4.5 6.562 635.0 3.937 15.0 30.0 691.828 3.272 0.354	77	4.5	6.562	635.0	4.039	15.0	30.0	700.248	2.647	0.513	0.459
	78	4.5	6.562	635.0	3.937	15.0	30.0	691,828	3.272	0.354	0.299

TABLE I: LASER TRANSIT ANEMOMETER / VELOCITY PROFILE DATA (concluded)

Es	(m/s)	0.345	0.377	0.529	0.487	0.113	0.619	0.553	0.767	0.821	0.865	0.925	1.061	0.721	0.831	0.912	0.923	0.979	1.022	1.065	1.254	1.161	1.253	1.260	1.199	1.148	1.124	1.167	1.144	1.418	1.405	1.395
Em	(m/s)	0.413	0.510	0.647	0.642	0.168	0.935	0.859	1.149	1.162	1.156	1.394	1.451	1.251	1.185	1.564	1.356	1.463	1.544	1.642	1.810	1.731	1.895	1.928	1.924	1.926	1.895	1.937	1.994	2.392	2.345	2,465
L	(%)	3.554	4.058	4.279	4.593	5.285	5.237	5.948	7.516	8.101	8.211	9.852	9.885	10.343	9.750	12.881	11.251	11.824	12.629	13.320	14.010	13.149	15.418	15.448	15.478	15,116	14.076	15.298	14.855	18.251	18.316	17.297
>	(s/m)	687.612	680.531	678.065	668.234	656.287	652.655	643.759	624.856	617.518	614.507	594.956	600.653	583.029	594.032	562.409	578.755	577.000	564.924	555.855	555.539	556.784	545.558	535.187	532.287	535.557	537.788	527.429	526.443	508.580	501.819	514.103
ST	(sec)	30.0	30.0	30.0	30.0	30.0	30.0	45.0	45.0	60.0	0.09	0.09	0.09	0.06	0.06	0.06	90.0	90.0	0.06	0.06	0.06	0.06	0.06	0.06	90.0	0.06	0.06	0.06	0.06	0.06	0.06	0.06
DT	(nSec)	15.0	15.0	15.0	15.0	15.0	15.0	15.0	15.0	15.0	15.0	15.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0	20.0
Z∇	(mm)	3.683	3.429	3.023	2.921	2.413	2.007	1.905	1.397	1.143	0.991	0.889	0.737	0.635	0.610	0.508	0.483	0.432	0.381	0.330	0.280	0.254	0.229	0.203	0.178	0.153	0.127	0.102	9200	0.051	0.026	0.000
Station	(mm)	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0	635.0
Re	(x 10 ⁶)	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6.562	6,562	6,562	6,562	6,562	6,562	6.562	6.562	6.562	6,562
Mach #		4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5	4.5
File #		62	8	81	82	83	\$	85	88	87	8	68	8	91	32	93	94	95	8	16	8	8	501	101	102	103	107	105	106	107	108	109



Cross Section

Figure 1: LTA Sample Volume Geometry




Figure 2: LTA Optical Head Schematics




Figure 3: Block Diagram of LTA Hardware

ORIGINAL PAGE BLACK AND WHITE PHOTOGRAPH




Figure 4: Laser Transit Anemometer System Mounted on Scan Rig




Figure 5: Geometric Description of Model (in inches)




Figure 6: Mounted Model in Test Section 2




Figure 7: Close-up Photo of Vibrated Bed Seeding System




Figure 8: Orifice Geometry (in inches)




Figure 9: Close-up Photo of Orifice Configuration on Model

ORIGINAL PAGE BLACK AND WHITE PHOTOGRAPH

DATA ACQUISITION PROCEDURE




Figure 10: Laser Transit Anemometer Data Acquisition Procedure




Figure 11: Parabolic Fit Procedure for Determining Mean Flow Angle

CORRELOGRAM PROCESSING




Figure 12: Laser Transit Anemometer Data Analysis Procedure




Figure 13: Example of Laser Transit Anemometer Data Processing Sequence




Figure 14: Correlogram Plots for Entries in Table I




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Continued




Figure 14: Concluded




Figure 15: Turbulence Intensity Profile I




Figure 16: Turbulence Intensity Profile II




Figure 17: Turbulence Intensity Profile III




Figure 18: Turbulence Intensity Profile IV




Figure 19: Boundary Layer Velocity Profile I




Figure 20: Boundary Layer Velocity Profile II




Figure 21: Boundary Layer Velocity Profile III




Figure 22: Boundary Layer Velocity Profile IV




Figure 23: Shadowgraph Photos of Model Mach # = 2.50 for all Photos

(a) Re = 3.281×10^6 / meter - Seeding On

(b) Re = 6.562×10^6 / meter - Seeding On

(c) Re = 3.281×10^6 / meter - Seeding Off

(d) Re = 6.562×10^6 / meter - Seeding Off

National Aeronaulic's and Space Agrinistration	Report Docume	entation Page		
1. Report No. NASA TM-102775	2. Government Accessio	n No.	3. Recipient's Catalog	g No.
4. Title and Subtitle			5. Report Date	
Laser Transit Anemometer Measurements on a Slender Cone in the Langley Unitary Plan Wind Tunnel			December 1990	
			6. Performing Organization Code	
7. Author(s) William M. Humphreys, Jr., William W. Hunter, Jr. Peter F. Covell, and Cecil E. Nichols, Jr.			Performing Organi 10. Work Unit No.	zation Report No.
			763-01-31-	26
 Performing Organization Name and NASA Langley Research 	Center		11. Contract or Grant	No.
Hampton, Virginia 236	665-5225			
12. Sponsoring Agency Name and Ad	dress	a division .	13. Type of Report an	d Period Covered
National Aeronautics a Washington, DC 20546-	ion	Technical Memorandum 14. Sponsoring Agency Code		
16. Abstract A Laser Translayer on a slender (5 Tunnel. The anemometed 40 micrometers spaced ensembles of seeding preasurements, boundary with CFD calculations. boundary layer velocities the model surface were different flow condities model's longitudinal cone. The results income measurements deep into flow field caused by ourstream seeding is matching the paper presents a 17. Key Words (Suggested by Author Laser Transit Anemometer)	er system utilized a particles using a cross layer profiles around the measured boundary normalized to the collected with the mons, and were collected enter line at a local licate an excellent allocated the boundary layer. Inboard seeding, premaindatory if model flow description and resurs.	one model in the pair of laser to measure to ss-correlation and the model we ary layer profedge velocity and the late of the l	ne Langley United the transit time technique. From technique, from the constructed and the construction of the construction of the construction of the construction of the construction occurred important occurred impo	tary Plan Wind iameter of mes of rom these in and compared ing the of height above for four bisected the the forebody make velocity bances in the olying that aintained.
Laser Velocimetry, LV Boundary Layer Profile Wind Tunnel Flow Measu	es S		ed - Unlimited tegory - 35	
19. Security Classif. (of this report)	20. Security Classif. (of t	his page)	21. No. of pages	22. Price
Unclassified	Unclassified		46	A03

1