



Direct-Solve Image-Based Wavefront Sensing

This method saves time and effort in testing of optical systems.

Goddard Space Flight Center, Greenbelt, Maryland

A method of wavefront sensing (more precisely characterized as a method of determining the deviation of a wavefront from a nominal figure) has been invented as an improved means of assessing the performance of an optical system as affected by such imperfections as misalignments, design errors, and fabrication errors. Unlike some prior methods, this method does not require the use of an expensive, complex interferometric instrument for testing the optical system of interest: indeed, if the system under test includes an image sensor at its focal plane, then this method does not require any optical instrumentation other than the optical system under test. Unlike some other prior methods, this method does not involve processing of multiple defocused images by a nonlinear iterative

phase-retrieval algorithm and interpretation of results by a human expert in phase retrieval. Instead, this method involves a single non-iterative algorithm that solves for the wavefront from a single in-focus image, without need for interpretation of results. Hence, the main advantages of this method over the prior methods are reduced computing time and reduced labor.

At the time of writing this article, only fragmentary information about the method is available. Beyond what has been stated above, what is known is the following:

- The method is implemented by software running on a single-processor computer that is connected, via a suitable interface, to the image sensor (typically, a charge-coupled device) in the system under test.

- The software collects a digitized single image from the image sensor.
- The image is displayed on a computer monitor.
- The software directly solves for the wavefront in a time interval of a fraction of a second.
- A picture of the wavefront is displayed.
- The solution process involves, among other things, fast Fourier transforms. It has been reported to the effect that some measure of the wavefront is decomposed into modes of the optical system under test, but it has not been reported whether this decomposition is postprocessing of the solution or part of the solution process.

This work was done by Richard G. Lyon of Goddard Space Flight Center. Further information is contained in a TSP (see page 1). GSC-15208-1

Use of UV Sources for Detection and Identification of Explosives

UV excitation is used to simultaneously detect Raman and fluorescence spectral information of explosive materials.

NASA's Jet Propulsion Laboratory, Pasadena, California

Measurement of Raman and native fluorescence emission using ultraviolet (UV) sources (<400 nm) on targeted materials is suitable for both sensitive detection and accurate identification of explosive materials. When the UV emission data are analyzed using a combination of Principal Component Analysis (PCA) and cluster analysis, chemicals and biological samples can be differentiated based on the geometric arrangement of molecules, the number of repeating aromatic rings, associated functional groups (nitrogen, sulfur, hydroxyl, and methyl), microbial life cycles (spores vs. vegetative cells), and the number of conjugated bonds. Explosive materials can be separated from one another as well as from a range of possible background materials, which includes microbes, car doors, motor oil, and fingerprints on car doors, etc. Many explosives are comprised of

similar atomic constituents found in potential background samples such as fingerprint oils/skin, motor oil, and soil. This technique is sensitive to chemical bonds between the elements that lead to the discriminating separability between backgrounds and explosive materials.

The unique combination of the wavelength, optics, mechanical configurations, and chemometrics enables standoff (1 to 5 m) identification of trace amounts of explosive materials with rapid spatial scanning capability. Each data point, which can include both the native fluorescence and Raman signals, is automatically identified in <100 μ s by the real-time analysis engine. The rapid acquisition and real-time analysis allows a user to scan the instrument over a large region such that the probability of false negatives resulting from a heterogeneous distribution of explosive material on a surface is

dramatically reduced.

The hand-held or robot-mounted instrument has been tested using a number of experimental conditions. In one example, a car panel doped with RDX (an explosive nitroamine) was placed 1 m away from the instrument. The car panel segment was rotated as the instrument collected data to mimic scanning from a fixed distance. The composite traces from the detectors are used by the analysis to show a relatedness index (high values indicate a high match) of each data point as a function of time and spatial position. This sample was part of a blind test to determine whether it was possible to identify RDX on the car panel in the presence of Arizona dust (a standardized interferant sample). As the sample is scanned, the RDX is found only in specific areas. In the other related experiments where the RDX samples were not so heterogeneous,