



Analysis of Clinical Records as a Means to Validate Non-Invasive Assessment of Intracranial Pressure Using the Cerebral and Cochlear Fluid Pressure (CCFP) Analyzer

D. Ebert, G. Gianoli, J. Soileau, R. Ploutz-Snyder, and R. Danielson



The EAR and
BALANCE INSTITUTE



Baylor
College of
Medicine

Background

- Visual impairment / intracranial pressure (VIIP) is a top risk for human spaceflight
- Currently no reliable means to monitor intracranial pressure (ICP) noninvasively
- The Cerebral and Cochlear Fluid Pressure (CCFP) Analyzer may be a viable option
- Additional validation required

Background

- The Ear and Balance Institute (Covington, LA) has used the CCFP as a screening tool for over 10 years
- Extensive clinical record database that includes lumbar puncture (LP) data
- vh3 retrospective clinical study was designed


Slide 3

vh3

Capitalize the "R"

vhurst, 2/4/2014

CCFP Physiology



Tympanic membrane displacement (TMD) concept: vh4 cochlear aqueduct connects the subarachnoid space to the perilymph and causes changes in ICP to be manifest as changes in the resting position of the stapes in the oval window.

Figure from: Marchbanks, R.J. and Reid, A. Cochlear and cerebrospinal fluid pressure: their interrelationships and control mechanisms. *British Journal of Audiology*, 1990; 24: 179-187.


Slide 4

vh4

Should this be in bold like the words before it?


vhurst, 2/4/2014

CCFP Evoked Mode




- Stapes resting position is affected by ICP
- **In response to a 1KHz tone**, the acoustic reflex changes stapes position and moves the tympanic membrane
- Volume change (tympanic membrane displacement) is measured in external ear canal (T_{tm} , measured in nL)
- Inward motion $\rightarrow -V_{tm}$ (panel a)
- Outward motion $\rightarrow +V_{tm}$ (panel b)
- Direction and magnitude indicate ICP changes

CCFP Evoked Mode



From: Philips, A.J. and Marchbanks R.J. Effects of posture and age on tympanic membrane displacement measurements. British Journal of Audiology, 1989, 23: 279-284.


CCFP Evoked Mode




*Figure 8 from: Samuel M, Burge DM, Marchbanks RJ. Tympanic membrane displacement testing in regular assessment of intracranial pressure in eight children with shunted hydrocephalus. J Neurosurg 1998;88:983-95.

CCFP Passive Mode

- Passive recording of tympanic membrane movement




LP = 20 cm



LP = 23 cm




LP = 35 cm



LP = 60 cm

**Lumbar Puncture
Pressure - cm saline**



Lumbar Puncture Pressure - cm saline

Data from a 15 year old idiopathic intracranial hypertension (IIH) patient; From: Measurement of inner ear fluid pressure and clinical applications. Marchbanks RJ (2003) Chapter 17, pp 289-307. In 'Textbook of Audiological Medicine: Clinical aspects of hearing and balance'. Ed L. Luxon. Martin Dunitz Ltd. 2003, ISBN 1-90186-534-7.

Inclusion/Exclusion Criteria

- CCFP and LP measures within several days, with CCFP occurring first
- No medical interventions between measures
 - Medication changes, surgery
 - Existing stable medication was not an exclusion
- LP-CCFP data pairs included:
 - 29 without superior semicircular canal dehiscence (SSCD)
 - 35 with SSCD

Caveats

- LP and CCFP data not collected at the same time
 - ICP variability
 - Patient medication compliance
- All subjects have some pathology, most are ear related
 - SSCD (~50%)
 - and perilymph fistula (~65%)

Superior Semicircular Canal Dehiscence (SSCD)





Image from <http://vestibular.org/superior-canal-dehiscence-scd>

Methods

- LP opening pressures taken in lateral decubitus



- CCFP taken from ear with better hearing and middle ear measures
- When multiple dB levels of test data were available, higher results were used
 - maximizes saturation of the acoustic reflex

Data




- 47 parameter fields
 - clinical data- not all fields had available data
 - many fields used for subject sorting and verification of valid LP-CCFP paired data
- Primary analysis focused on: LP, Vm (seated and supine), pulse amplitude (seated and supine), air conduction audiometric threshold at 1kHz, middle ear peak pressure and compliance, and age at time of test

Data


- Interactive graphs
 - courtesy of James Fiedler of the JSC Biostatistics Laboratory
 - *Note to reviewers: the following 2 slides are screenshots of the interactive graphs that will be demonstrated during this portion of the presentation*

Data: SSCD


Interactive Graphs: SSCD Subjects




filter




filter





filter



filter








filter




Data: non-SSCD


Interactive Graphs: non-SSCD Subjects




filter




filter



filter



filter



filter

Statistics

- Somers' D: LP vs Vm
 - Assumption-free
 - Allows for data clustering (multiple measures per person)
- SS CD Subjects
 - Seated: Coefficient= $-.109$, std err = $.147$, $p=.460$
 - Supine: Coefficient= $-.037$, std err = $.162$, $p=.819$
- Non-SS CD Subjects
 - Seated: Coefficient= $-.323$, std err = $.114$, **$p=.004$**
 - Supine: Coefficient= $-.339$, std err = $.118$, **$p=.004$**


Statistics

- Mixed-effects regression- LP compared to:
 - V_m
 - Days between LP and CCFP
 - Pulse amplitude
- Untransformed data
 - $\ln(LP)$ similar results

Statistics

- Non-SSCD Subjects
 - Vm Seated: Coefficient= $-.0084$, std err = $.0026$, **p=.001**
 - No significant effect of days between tests and pulse amplitude
 - Supine model did not fully converge
 - Vm Coefficient= $-.0053$, std err = $.0029$, p= $.068$
 - Pulse Amplitude Coefficient= $-.0074$, std err = $.0046$, p= $.104$
 - No significant effect of days between tests


Cochlear Aqueduct Patency



delta Vm/Vm > 0.1 indicates patent cochlear aqueduct

Cochlear Aqueduct Patency

Age and Patency Distribution



Age group	0-10	10-20	20-30	30-40	40-50	50-60	60-70	70-80	80-90
Patency rate (%)	-	100	100	86	77	86	100	100	-

Sleep – Wake LP measures

	LP (cm H ₂ O) ^{vh5}		
	<u>wake</u>	<u>sleep</u>	<u>delta</u>
	25	43	18
	18	27	9
	23	28	5
	16	19	3
	17	25	8
	15	32	17
Average	19	29	10

Patients are in lateral decubitus position for both measures

Patients are being closely monitored by anesthesiologist; not likely a hypoventilation effect

vh5

Captitalize

vhurst, 2/4/2014



Conclusions

- There is a significant correlation of LP and seated Vm in the non-SSCD subjects in this population
- There is trending of LP with seated Vm and seated pulse amplitude in the non-SSCD subjects in this population
- Clinical factors such as hearing loss likely affect repeatability and reliability, although specific contribution levels could not be identified within this data set
- In this population, cochlear aqueduct patency rates are high and do not appear to decline with age
- Considering the limitations associated with this clinical data set, results are highly encouraging as to the utility of the CCFP as an ICP screening tool

Acknowledgements

- Rachel Brady – spreadsheet work
- James Fiedler- interactive graphing software

Discussion



wyle

The EAR and
BALANCE INSTITUTE



Baylor
College of
Medicine