

# X-Ray Photoelectron Spectroscopy

Binayak Panda, Marshall Space Flight Center / NASA

## General Uses

- Elemental analysis of surfaces of all elements Lithium and higher
- Analyses the first few atomic layers on samples
- Thin surface layers that cannot be analyzed by other techniques
- Chemical state identification of surface elements
- In-depth composition profiles for elemental distribution as well as oxidation states in samples
- Surface analysis of samples when destructive effects of electron or ion beam techniques must be avoided

## Examples of Applications

- Determination of oxidation states of metal atoms in metal compounds
- Identification of surface contaminations
- Measurement of surfaces film thickness
- Identification and degradation of polymers

## Samples

- *Form* : Flat solids with low vapor pressure are preferred
- *Size* : Depends on machine – up to 4 inches
- *Preparation*: Must be free of fingerprints, oils, or surface contamination

## Limitations

- Data collection is slow compared to other surface analysis techniques, but analysis time can be decreased substantially when high resolution or chemical state identification is not needed
- Lower lateral resolution in imaging and analysis compared to electron microscopes
- Surface sensitive, comparable to other surface analysis techniques

- Charging effects may be a problem with insulating samples. Some instruments are equipped with charge-compensation devices

## Estimated Analysis Time

- Requires few hours of vacuum pump down before analysis
- Qualitative analysis can be performed in 5 to 10 minutes
- Quantitative analysis requires 1 h to several hours, depending on information desired

## Capabilities of Related Techniques

- *Auger electron spectroscopy*: Compositional analysis of surfaces. Faster, with better lateral resolution than XPS. Has depth-profiling capabilities. Electron beam can be very damaging; bonding and other chemical state information are not easily interpreted. Very thin surface layers cannot be analyzed.
- *Low-energy ion-scattering spectroscopy*: Sensitive to the top atomic layer of the surface and has profiling capabilities. Quantitative analysis requires use of standards; no chemical state information; poor mass resolution for high-Z elements
- *Ultraviolet Photoelectron spectroscopy (UPS)*: UPS operates with the same principles as the XPS, the exciting radiation in this case is the ultraviolet light with a narrow band of wavelength. In this technique the core level electrons are not accessible but does provide useful information at the surface level (valence band level). Analysis depth is estimated around 20 to 30 Angstroms. Used for information on surface species, reaction products, catalysis, and chemisorption.

## Introduction:

The interaction of photon and the electron goes back to the early part of 19<sup>th</sup> century emanating from the photo-electric effect depicted by none other than Albert Einstein (Ref 1) described in 1905, and the redistribution of kinetic energy resulting from the interaction of x-ray and solids reported during early part of the century (Ref.2). The spectrum resolutions obtained at that time was not sufficient to observe distinct peaks in spectra for materials. Thus, these phenomena hardly attracted any attention for many years following these discoveries.

The modern X-ray Photoelectron Spectroscopy (XPS) has been possible by the extensive and significant contribution from Kai Siegbahn and others (Ref.3, 4) of Uppsala University. Siegbahn developed and employed a high-resolution electron spectrometer that revealed electron peaks in a spectrum emerging from the interaction of x-rays and solids. Eventually, Kai Siegbahn received Nobel Prize in 1981 for his contributions to XPS. Around 1958, shifts in elemental peaks were realized in compounds when the same elements are bound to other but different elements. This discovery resulted in the chemical state identification in various chemicals as well as the oxidation states of atoms in compounds. Because of these useful physical effects, the Uppsala group named XPS with a synonymous name of ESCA (Electron Spectroscopy for Chemical Analysis) used widely today and will be used here alternatively. Therefore, XPS or ESCA not only identifies the element, but also the compound these elements form, from their chemical shifts.

Compared to other micro-analytical techniques such as Energy Dispersive (EDS) or Wavelength Dispersive (WDS) techniques, XPS analyzes only few atomic layers present on the surface. This was discovered early in 1966 (Ref. 5). While this has awarded a merit to the analytical technique to analyze very thin layers such as films and coatings, it often analyzes the adsorbed superficial gases and contaminations on a sample introduced to its analytical chamber. This necessitates the surface is cleaned and the underlying material, material of interest, is exposed in a clean environment such that the material of interest is analyzed. The cleaning is accomplished by a scanning ion gun within the analytical chamber of the instrument. Ion gun uses an argon gas and is commonly attached in most modern machines. Reliable and efficient vacuum systems employed in modern machines does not allow adsorbed layers to rebuild after the surface is cleaned.

Development of efficient and reliable vacuum pumps over these developmental years is yet another important step in the commercialization of XPS machines. Vacuum levels of better than  $10^{-7}$  torr are essential to increase the mean free path of electrons released from the sample surface. Thus, modern machines are equipped with high capacity ion, turbo or cryogenic pumps in their analytical chambers.

Today, XPS has advanced from an applied physics laboratory to industry for use in quality control as well as analysis of contaminants and has taken a dominant role in microanalysis. Its uniqueness arises from the fact that it is considered non-destructive compared to other common micro-analytical techniques using the electron and ion excitation sources. Polymers and plastics could be analyzed since the binding energies of saturated and unsaturated bonds in atoms could be separated. Extremely thin layers could be analyzed including materials with layered structures. The technique, though did not advance for many years, has now opened a new window for research as well as applications in industry due to its ability to separate and measure the chemical shifts in bound elements.

## Principles

Fig. 1(Ref. 6) illustrates the electronic transitions involved in an XPS process. It shows an energetic x-ray beam impinging on the surface. Due to their high energy, they eject one or more core electrons. The ejected electrons are collected by the spectrometer and eventually detected by a multi-channel

analyzer. The process of this interaction conserves total energy for each interaction and from the voltages needed for the retardation of the ejected electron, the energy of the electron is known. This associated energy leads to the identification of the element and the orbit from which the electron is knocked-off. Mathematically, if the kinetic energy  $KE$  is the core electron immediately after it was ejected, then, due to the conservation of energy,

$$KE = h\nu - BE + \alpha \quad (\text{Eq. 1})$$

Where,  $h\nu$  is the energy of the exciting radiation,  $BE$  is the binding energy of the emitted electron in the solid, and  $\alpha$  is the spectrometer work function. Value of  $\alpha$  depends on the machine design that would vary from model to model.

Often the vacancy created by the released electron from the photon interaction is filled by an electron from a higher energy level. The extra energy would be released by photon emission or by knocking of another electron with discrete energy known as Auger electron. The spectrometer also collects the Auger electron and displays it along with the other core level electrons. Auger emission is the more probable transition for low atomic number elements.

From Eq.1 it is clear that the position of the XPS lines (Binding energy of peaks) depend on the energy of the x-ray photon and hence the target material of the x-ray source. Most common target materials are Aluminum and Magnesium. Existing data bases include both line positions for Al and Mg x-ray targets since the positions would vary depending on the wavelengths of the exciting beams. The Auger line positions, however, do not change as the Auger electron energies liberate from the internal electron exchange process within an atom rather than on the impinging photon. When two or more elements combine to form a compound, their electrons change in their binding energies which is reflected in their corresponding line position shifts in the spectrum. Al and Mg x-ray targets are selected in most XPS machines due to strong x-ray beam intensity and the capability of Al  $K\alpha$  being filtered by quartz crystals.

Spectra generated by XPS instruments are mostly similar to the one shown in Fig.1 (Ref.6) for Cu. Fig.1 shows both photoelectron and Auger lines, there are, however, several other types of peaks found in XPS spectra largely due to aberrations and complexity of the interactions that the emitted photo-electron experiences.

X-ray Satellite peaks are found when the x-ray is not quite monochromatic. Small peaks that appear towards the lower binding energies of XPS peaks. See Fig.2 (Ref. 7, page18)) coming from small number of photons with higher energy than the main exciting x-rays. Ghost lines appear due to impurities in anode material. The x-rays generating from the impurity atoms cannot get filtered and show additional lines in the spectra. Shake-up lines appear by the retardation effects of the ions that are created by the ejection of a photoelectron. This retarding effect generates small humps toward the higher binding energy side of a peak. The  $\pi \rightarrow \pi^*$  shake-up for C 1s (see the next section on nomenclatures such as C 1s, Si 2p etc. used in this chapter) line is a common example. Fig.3 (Ref.7, page 19) shows shake-up lines in copper compounds.

Photoelectrons may react with the surface atoms and lose their energy. Such electrons, when detected by the spectrometer, appear as a hump at about 20 – 25 eV above the binding energy of a peak. They are called as energy loss lines or plasmon lines. Some cases they appear in a periodic fashion in 20-25 eV intervals of diminishing intensity. Fig.4 (Ref.7, page 21) shows plasmon lines in Al.

XPS spectrum very close to Fermi level (0 to 20 eV), known as Valence Band Spectrum, is often utilized to distinguish between conductors and insulators

## **Nomenclature:**

The Binding Energy of an electron is the energy required to free the electron from its orbit. This is also known as the ionization energy and is commonly expressed in electron-volt (eV). XPS data from a sample often plotted with eV on the abscissa and intensity of the peaks in the number of counts on the ordinate. Such presentations are presented from right to left with increasing electron volts. Sometimes, the ejected electron energies are presented in the form of Kinetic Energies and are presented with increasing electron volts from left to right. Most presentations here are shown in this Chapter are Binding Energy spectra.

The Binding Energy may be regarded as energy difference between the initial and final state of an atom after an electron is ejected. The energy levels of an atom involved in photo-emission are represented in terms of an energy-level diagram that provides the energy of the atom when one electron of the indicated energy level, s, p, d, or f is missing. Fig. 5 (Ref. 7, page 10) shows an energy level diagram for Uranium metal. XPS spectra obtained from a machine shows peaks of both Auger and photoelectron lines. Nomenclature used for the photoelectron and Auger lines are different and are shown in Fig.6 (Ref.7, page 11). Photoelectron lines are designated as s, p, d, and f levels (from the shell the electron came out of), whereas the Auger electrons are designated using the x-ray spectral designations, K, L, M etc. depending on the Auger process. Fig. 7 (Ref.7 page 194) shows a survey spectrum obtained from Uranium metal using Al K-alpha radiation. It can be compared with the energy levels of U shown in Fig. 5 (Ref.7, page 10).

## **Systems and Equipment:**

XPS instruments analyze the first few atomic layers of the sample. Usually, these layers also contain the contaminations from the atmospheres and the environment that the sample is exposed prior to the sample reaching the laboratory. While enough care is taken to not to contaminate any further in the laboratory, it becomes essential to remove the unwanted contamination from the surface of the sample to reach the base material inside. An ion gun is used for this purpose and is an essential part of a modern machine. There are several types of ion guns available but an Argon ion gun is most common.

The main part of the machine consists of an x-ray generator unit with a filtration system for the generation of monochromatic x-rays.

Vacuum systems are inherent to the design of the instrument. Better vacuum of the order of  $1.0 \times 10^{-10}$  can give excellent results, though a  $1.0 \times 10^{-7}$  is generally adequate. The vacuum system is an integral part of the instrument.

Charge neutralizers are equally important as the part of the XPS system when non-metallic materials are analyzed, and come as an integral part of the XPS system.

### **X-Ray Source:**

As mentioned earlier, XPS literature is clustered around sources that have either Al or Mg anodes for the generation of x-rays, are relatively inexpensive, and can generate intense beams. The sources start with fine electron beams (fineness of an electron beam can be controlled but is difficult to control the fineness of an x-ray beam) on to a cooled but pure Al or Mg targets. The x-ray that comes out of this electronic impact is an intense white radiation. As such this radiation can be used

for XPS but would create a large number of peak lines in the spectrum.

A narrow and often monochromatic intense beam is sought for XPS analysis which is accomplished by filtering the white radiation generated from the source. For Al K-alpha radiation, a bent quartz crystal is used. While the filtration reduces the intensity, to generate more intense beams, in some advanced machines, seven such crystals are employed, arranged with one at the center and six others around the center one. This configuration increases the intensity seven-fold compared to a single crystal. Each crystal subtending the same solid angle with the source reflects only a narrow wavelength (Al K $\alpha$ ) by the quartz crystal atomic layers. These crystals are not only bent to focus the x-ray beam, but also are maintained at a constant temperature for constant lattice spacings.

### **Analyzers:**

Electrons emerging from the sample are collected by field lenses, some designs may use a strong magnet below the sample to focus these electrons to the analyzer. Generally, the analyzers are either a hemispherical analyzer or a cylindrical mirror analyzer. Fig.8 (Ref.8, page 131) shows a schematic of a hemi-spherical analyzer. In a hemispherical or spherical sector analyzer, the two concentric charged hemi-spherical lens segments are charged separately such that the electrons coming from the same electronic shell are focused to a point on the detector generating the signal for a spectral line. A detector is generally a channel plate or an electron multiplier.

Modern machines employ a multi-channel plate with a computer for digital counting. Photo-electrons hitting the multi-channel plate create pulses which are eventually counted as a function of time. Sometimes the scans are repeated when the whole scanning process is also repeated. The number of photo-electrons gathered at the channel plate constitute the height of a peak of a particular binding energy and all the binding energies of all photo-electrons are covered by varying voltages of the spherical analyzers.

### **Vacuum Systems:**

The capacities of the vacuum systems are such that the vacuum levels at the sample chamber and the analyzer is maintained very low to have a long mean free path for the photo-electrons generated. Vacuum levels of the order of  $1.0 \times 10^{-8}$  or  $1.0 \times 10^{-9}$  torr are obtained using ion, cryogenic or turbo pumps of adequate capacity attached to the analytical chamber. These pumps are most efficient at high vacuum level and therefore are backed by a conventional mechanical rotary or dry pump generating a vacuum level of around  $1.0 \times 10^{-3}$  torr. In most instruments, there is also a sample preparation chamber prior to the analytical chamber to introduce the sample from air and to expel volatiles. The vacuum level in this chamber is maintained at a lower level (around  $1.0 \times 10^{-3}$  torr). Samples with holders are introduced to the analytical chamber after they spend enough time to remove most volatiles in this sample preparation chamber.

### **Charge neutralizers:**

When photo-electrons leave the surface of a sample a positive hole is created. In conductive samples the electron hole is neutralized by electrons coming from ground immediately (provided there is a good connection to ground). For non-conducting samples, the positive hole remains on the surface and attracts the ejecting photo-electron reducing its kinetic energy. Erroneously, due to the reduction in kinetic energy, these electrons appear at a higher binding energy level on the spectrum and broaden the peak. A charge-neutralizer attempts to compensate this effect by simply spraying low energy electrons onto the sample surface. Thus a charge neutralizer is nothing but an electron gun. It may appear simple, but the x-ray spot has varying intensity needing more electron at the

center of the spot for compensation which is not easily done. Modern XPS machines have complex charge neutralizers to counteract this effect as much as possible.

### **Ion Guns:**

Argon ion guns are extensively used for XPS analysis to remove material from or to clean the surface of the samples being analyzed. In recent days, cluster atom ion guns are gaining popularity for use in polymeric materials since the cluster ions prevent damage to the surface bonds often encountered when argon ion guns are used.

An argon ion gun generates ions when argon gas is passed over heated wires carrying electricity. While a small percentage of the flowing gas is ionized, these ions could be manipulated and focused using a series of electro-static lenses to generate a scanning ion beam. The ions due to their heavy mass compared to an electron in an electron microscope are difficult to manipulate and require strong electrostatic force to focus. In any case, these charged ions hit the sample surface like bowling balls and remove surface layers. The scanned area generally covers more area than the area being analyzed.

Samples with uneven surface would be cleaned unevenly by an ion gun by creating a shadowing effect as the guns are normally at an angle to the sample surface. To counteract this effect, the samples are rotated around the point being analyzed. This is accomplished by placing the sample at the center of the analytical stage. Some instruments have arrangements (called as eucentric stages) where the sample and the stage rotate around the point of interest rather than around the center of the stage.

### **Accessories:**

Since XPS has an electron spectrometer, other analytical excitation sources such as an electron source for Auger Electron Spectroscopy and an ultraviolet photon source for Ultraviolet Photoelectron Spectroscopy (UPS) are often come as attachments to the main XPS instrument. Kratos ESCA shown in Fig.9 has attachments such as Fracture Stage and Reaction cell where fresh surfaces in metals and alloys could be studied in vacuum after their fractured inside the instrument or after a chemical reaction is exercised. In the Reaction Cell chamber, studies in catalysis and chemical reactions can be made after chemical reactions take place between the candidate materials and gases and the reaction products are transferred within the instrument under high vacuum to perform chemical analysis avoiding reactions with atmosphere. Fracture Stage attachment is used when the grain boundaries in alloys are suspected to be responsible for low fracture test values and need investigation.

## **Specimen Preparation:**

As mentioned, XPS analyzes the first few layers on a specimen. It is therefore, important to not to contaminant these layers by touching or mis-handling the specimen. This is especially true for specimens targeted for failure analysis. Often times it has become necessary to use gloves that are manufactured with no mold releasing agents (generally silicones as the silicones migrate to the analysis sites and can give false indications for the presence of silicones).

Any sample exposed to air would show adsorbed molecules of H<sub>2</sub>O or CO<sub>2</sub>. These adsorbed molecules must be removed to analyze the material underneath. Ion guns are employed for this purpose. Ions guns are mounted at an angle to the sample surface, and therefore, non-uniform cleaning occurs on an uneven surface. A flat surface or sample rotation is required to avoid this.

The specimen should be placed flat (as opposed to a small angle) to the x-ray beam. If the sample plane interacts at a low angle with the x-ray beam surface areas are analyzed more than the inside materials. This behavior is taken advantage of when a film or thin coating on the specimen needs to be characterized. This is called as Angle Resolved Spectroscopy (AR-XPS) when the specimen is tilted at different angles and analysis is performed providing more information on the surface or the coating.

### **Powder specimens:**

Powder samples can be mounted and analyzed using adhesive tapes. Since most tapes use organic volatiles, tapes compatible with high vacuum in the analytical chamber should be used. Usually, the powder would not cover the tape surface completely, carbon or silicon peaks on the tapes would show up on the spectrum when the x-ray would fall on the uncovered areas.

Another technique effectively used to analyze powder specimens is by mounting the powder in soft metals such as Indium. In this case, the powder is placed on an Indium sheet and the sheet is folded over to retain the powder inside. Generally, the Indium piece that is used as a sheet is cut from a small Indium ingot with a sharp knife such that the clean knife creates two fresh surfaces where the powder would be placed (on one side only) on one of the fresh surfaces. The folded Indium sheet is then flattened further by rolling or hammering. This process not only embeds the powder into the metal sheet, but also, the particles get squeezed to make good electrical contact. The two folded sides are then separated to expose the powder and then the embedded powder is analyzed. In this case the In lines are seen in the spectrum obtained along with the spectrum of the powder.

The powder can also be briquetted to generate a flat surface for analysis using a die and punch. While this process generates a solid specimen for analysis, there is always a small amount of material transfer from the die and punch. Care also should be taken that the specimen does not transform to a compound under high pressure and adiabatic temperature rise when the die and the punch are squeezed.

Specimens required for the XPS analysis are small and may be cut using wire saw or a diamond wheel. However, the surface of specimens would be contaminated with cutting fluid or the cutting blade materials. The analyst needs to know such material/chemical information of the fluids and the saw materials.

### **Sample Charging:**

Sample charging is inherent with insulators. For semiconductors, it depends on the incoming x-ray intensity. Conductive metals can also accumulate charge if the electron holes are not filled quickly by the electron from the instrument body. Most instruments incorporate spring clips or flat springs in their stage designs to make the contact as good as possible. While the charging effect distorts the spectrum completely in insulators, they can shift the spectrum slightly in alloys. It is important, therefore, that the specimen be loaded to the analytical chamber such that enough conduction path for electrons exists.

Mounted metallographic specimens cannot be analyzed due to lack of conductivity and the outgassing of the embedding matrix. To improve the analytical situation the mounting material must be of very small amount and a gold (or any other known coating) may be applied to provide conductivity. The mounts should spend enough time in the pre-analytical (or sample preparation) chamber prior to their insertion to the analytical chamber to remove volatiles.

Charge compensation in modern instruments are done by supplying electrons to the location where the x-ray beam is falling spot. However, for some instruments, the intensity of the x-ray spot is not uniform across. This poses problems due to some areas of the spot not being able to compensate.

Some instruments have their patented charge neutralizing combination of low energy electron gun with a mesh-screen device shown in Figure 10. The device incorporates a very thin stretched nickel screen on top of the specimen holder and the specimen placed about 0.5 to 1.2 mm below the screen. The flood gun throws low energy electron on to the sample through the screen (Ref. 9).

## **Calibration and Accuracy:**

Most machines are calibrated using lines for noble metals such as Au, Ag, and Cu. Excepting gold, other pure elements react with the atmosphere to form oxides or sulfides on the surface that need to be cleaned by sputtering. Thus, the calibration material is common but must be clean.

Most metals exposed to atmosphere are covered with carbon known as ‘adventitious carbon’ from the atmosphere. This is saturated carbon and is taken advantage of as the reference for the entire XPS binding energy spectrum from 0 eV to 1200 eV. The position for C 1s line is 284.6 (some take it as 284.4 or even 285 eV). The span of binding energy is calibrated with respect to the BE for copper 3s (122.4 eV) and copper 2p<sub>3/2</sub> (932.5 eV).

Older systems such as those made by SSI ask for daily check of gold (Au 4f<sub>7/2</sub>) at 83.98 eV and the split of 4f lines for Au to be at 3.68 eV. For SSI X-Probe, sanity of the machine is checked from the Au 4f line separation (to be at 3.68 eV ±0.02eV) and the location of the Au 4f<sub>7/2</sub> line to be at 83.98±0.1 eV. The width of the peak at half height (FWHM) of the peak should be 0.95 eV. In addition, the integrated intensity of the peak should be at least 60,000 counts. Modern instruments which incorporate intense monochromatic x-rays such as the one shown in Fig.9 are expected to yield ten times more counts for the same lines. Kratos uses Ag 3d<sub>5/2</sub> peak at 368.3 eV for their intensity calibrations and system health check.

XPS machines have several adjustments to generate the desired spectra. Pass Energy regulation allows control of passing electrons of a narrow energy band to reach the channel plate. A low pass energy means fewer bandwidth of electrons being collected which indicates a higher resolution spectrum. To get a high resolution spectrum with sufficient counts, an intense x-ray beam is essential, however, machines with low intense beams accomplish high resolution by increasing the analysis time. Fig.11 shows the Si2p line split with low pass energy using Kratos AXIS Ultra spectrometer with high resolution.

### **Accuracy:**

The spectrum obtained from XPS has two types of information in two axes of the spectrum. They are the binding energy in eV and the intensity of the photoelectron line. The binding energy gives information on the binding state of an atom e.g., whether the O is present as an oxide, di-oxide, or sulfate. These states are estimated from the line shift. Such chemical shifts are often small and when accurate, provide useful information. It is estimated that a well performed analysis would have an accuracy of about +/- 0.2 eV of the line position depicted in the literature.

The intensity, or the area under the curve that constitute a peak, depends on a number of factors including sample charging, interference from other peaks, sample thickness, sample inclination to the beam, etc. This peak integration in a resolution that is estimated to be around 10%. In the next section, “Data analysis and Reliability” some aspects of this accuracy in quantitative analysis is elaborated.

## **Data Analysis and Reliability:**



Quantitative surface analysis by XPS requires several operations including the specimen preparation, adjustment to machine parameters, and data manipulation by the operator that effect the level of accuracy. The final result in terms of uncertainty in measurements is the total amount of uncertainty introduced by all these operations.

The most useful aspect of XPS is the shift of binding energy lines due to chemical interactions between atoms. These shifts and the presence of other reacting elements establishes the nature of compounds that are present on the surface of the sample. Shift in spectral lines are then compared with the shifts available in literature and other data bases and handbooks (Ref. 7 and 9). Fortunately, there is not a lot of variation in the measurements of line positions. Even if there is a small amount of charging effects seen on the sample (and as a result there is a small amount of shift in the line position) the difference between other line positions for elements that are suspected to be inactive (such as the adsorbed oxygen or carbon on the surface) can be used as reference and their position differences can be used for the actual shift determination. There are several data bases that are available (Ref. 7,9, and 10) but should be used with caution as data from older instruments, perhaps obtained using non-monochromatic radiation may have been included in them. It is always good to go to the original publication to know how the original experiments were carried out. In any case, there may be a variation of about 0.2 eV between two experiments so far as the line positioning is concerned.

The intensity of the photoelectron lines is an aspect which reduces confidence in the quantitative analytical results of XPS. The following illustrates why such is the case. In the XPS process it is assumed that the number of electrons collected and detected for a given transition are proportional to the number on the surface being analyzed. What is not correct in this assumption is that all the ejected electrons are accounted for. Electrons with higher or lower energies than the ones forming the main photoelectron peak due to acceleration or retardation by overcompensation or under compensation of specimen charge compensation system tend to shift away from the detection by the spectrometer. They broaden the peak and make the peak asymmetric. This phenomenon effects spectrum for insulating samples where the peaks are wider than those obtained for conducting specimens. In general, however, the peaks are broadened to the high binding side of peaks and may even create separate lines (Shake-up lines in polymers). To separate the background from the peaks, one of two popular methods are incorporated in commercial softwares. The background separation from the peak can be made by a straight-line by joining bottommost points of the peaks or they can be joined by a Shirley curve (Ref. 11). The Shirley curve shape is determined by an iterative process where the background intensity at a point is proportional to the total peak area above the back ground and to higher energy.

Fig.12 (Wikipedia) compares simple background subtractions such as linear or quadratic to a curved Shirley background. Background corrections by Tougaard (Ref.10,12) have been claimed by some to be better than Shirley corrections. The background subtraction is inherent to the software used and may be different for different XPS instrument manufacturers.

Once the background has been subtracted, the intensity of the peak remains and it is processed using the software that eventually leads to a composition table. When performing these calculations, Relative Sensitivity Factors (RSFs) are used to scale the measured peak areas so that they are related to the amount of material present on the sample. RSF calculations date back to Scofield's report (Ref.13) from early 70's and are modified to instrument design. For example, for Surface Science's SSX-100 software, an Adjusted Sensitivity Factor, SF is modified such that it is equal to  $SF_0$  multiplied by  $\{(Al K_{\alpha} - B.E.)/(Al K_{\alpha} - C 1s)\}^x$  where, B.E. is the Binding Energy of an electron from Scofield Table (Ref.13),  $Al K_{\alpha}=1486.6$  eV,  $C 1s$  is the Binding Energy for  $C 1s$ , and  $x$  has a default value of 0.7. Following Table lists some SF values of Fe taken with  $C 1s$  as 1.0:

Line	SF	Line position
C 1s	1.0	284.6 eV
Fe 2p <sub>3</sub>	10.82	710.0 eV
Fe 2p (doublets incl.)	16.42	710.0 eV
Fe 3p	1.669	56 eV

It is easy to see from the above table why the quantification for Fe analysis could be erroneous when one takes a single peak vs. a double peak. Software from Casa XPS (Ref. 10) list three RSF values for doublets depending which one of either of the doublets or both doublets together gain consideration for calculation. For SSX-100 software, the C 1s (SF taken as 1.00). Other machines by Physical Electronics or Kratos take F 1s as reference (value taken as 1.0). Figure 13 shows the Sensitivity Factors for elements (at top right-hand corners for each element) for an XPS machine by Physical Electronics where the x-ray source is at 90 degrees to the axis of the analyzer, along with other physical parameters (Ref. 7 page 261).

For metals and alloy analysis, polished surfaces of standard alloys can be used and peak areas evaluated for the known and unknown samples. With background appropriately subtracted, the peak area ratios between known and unknown specimens can yield reliable results.

### Data Analysis:

An XPS spectrum is displayed as a plot between the intensity (number of electrons of a definite binding energy) in the ordinate and the binding energy in the abscissa. Sometimes the binding energy is replaced by the electron kinetic energy. The binding energy plots increase from right to left whereas the kinetic energy plots increase from left to right. Data analysis starts with the identification of lines that are easily found in the XPS spectrum. The first step in line identification is to look for lines such as C 1s (known as adventitious carbon originating from the environment) and the O1s lines. C and O are invariably present on the surface which is unsputtered and unclean. They may be present even after light sputtering. After sputtering by Argon ions, one is expected to see Ar lines around 241 eV and may be used as reference too. It is important that both the C 1s and O1s appear as thin and sharp lines, otherwise, sample charging is suspected. The C 1s and O 1s line widths can be compared with the instrument standards and watch for deviations. In some instruments there is a possibility to adjust charge neutralizer parameters such that the width at half peak height (FWHM) is minimized and the position for these lines are noted. Even after such manipulations, the C 1s and O 1s lines may be found wider than the instrument specifies, indicates other peaks hiding inside. Fig.14 shows side-by-side scans by two machines for a common reference polymer, PET (Poly-ethylene Terephthalate) (Ref.9 page252 and Ref.14). It is interesting to see that even if all carbon peaks are resolved clearly, earlier instruments provide wider peaks and in some cases may not be able to resolve peaks to the satisfaction of the operator.

Once the C1s or O 1s line is identified and its position is determined accurately at 284.6 eV (for C 1s), other intense lines should be identified. Once done, smaller and perhaps wider lines such as the shake-up, satellite or Auger lines associated with the main intense lines could be identified. Following this, the low intensity lines are generally identified. It is also possible that the most intense lines of an unknown element present in small quantities can show up as a small peak. At the conclusion, spin doublets for p, d, and f lines are identified. They should have the right separation and height or intensity ratios. Multiple splitting and shake-up lines can also be utilized for chemical identification (Ref.7 page 20).

Important advantage of XPS analysis is its ability to measure small shifts due to chemical interactions in compounds. These shifts are small but measurable. These shifts are based on experimental data since theoretical calculations are not satisfactory. Handbooks and reference data (Ref. 7, 10) on shifts of main elemental lines have been compiled and provided for operators' references.

### **Curve Fitting:**

Often several peaks are superimposed resulting in skewed or wide envelopes for peaks. Most commercial XPS software provide a peak-fitting algorithm along with the peak identification software. Each of these programs have options for line positions for one or more peaks inside the large peak envelope. One can also adjust peak widths and different types of back ground subtractions so that the peaks best fit the large peak envelope. The fitting is done by the software and the extent of the curve-fitting is also calculated. The instrument operator has only to check the possibility of the existence of such peaks. See Fig. 15 (Ref. 14, page 196-197) for the fitting inside of C 1s line for Nylon 6, 6 and can be compared it with binding of carbon atoms on the molecular chain in page 196.

### **Valence Band Spectra:**

Valence band spectra, in the range of -5 eV to about 50 eV, has been found to be useful in several cases where the core level shift is subtle, valence band spectra have been employed. These spectra have also been used as fingerprints of materials rather than identifying a specific molecular orbital. Fig. 15 also shows valence band spectrum for Nylon 6,6 (ref. 14, page 197).

## **Applications and Interpretation:**

From the very onset of the development of commercial XPS systems with the associated monochromators and ion guns, XPS has emerged as a very useful tool in all kinds of analysis. Due to its ability to analyze monolayers of surface materials and that, can detect elements above the atomic number Lithium, and the presence of elements being registered in the form of sharp peaks, initial researchers used the technique extensively to understand catalysis. Since the surface contamination can be removed by Argon sputtering, and that the metals and alloys are conductive, metallurgical analyses can easily be performed and has taken center stage for the analysis of coatings, plating, study of diffusional processes, development of materials for adhesion and metallic bonding, as well as applications in microelectronics. The most attractive capability of XPS is perhaps in the analysis of polymers, their failure analysis, production issues and characterization capabilities as significant bonds in polymers could be resolved. With advancements in x-ray sources, cluster guns and imaging capabilities, XPS is useful for investigating almost all surface problems. Below are examples (brief summary of reports) of how XPS data can be used to solve problems with existing surface interactions, or to investigate new materials.

General applications of XPS analysis includes (1) identification of unknown substances or material characterization, (2) measurement of layer thickness in layered engineering materials, (3) quality control of products with thin coatings, (4) contamination analysis, and (5) general failure analysis where failure is investigated using XPS as a tool. What follows are examples of cases where XPS has been successfully employed in resolving or analyzing problems that cannot be done by other analytical instruments such as electron microscopes or electron microprobe analyzers.

**Example of #1:** Rhenium (Re) is an expensive but a ductile high temperature material. A rocket engine (thruster) with Re liner was tested and at the end of testing there was a clear liquid found all around inside the engine. The samples were brought to the NASA Marshall Space Flight Center laboratory using a cotton swab. To analyze the residue, the cotton swab was rubbed on an aluminum foil and the residue, though nearly invisible to the naked eye, was transferred to the aluminum foil. The sample spent one hour in the preparation chamber and then analyzed using a Surface Science SSX-100 XPS machine. Several Re peaks were identified indicating the residue to be  $\text{ReO}_3$ . There were also other Re oxides ( $\text{ReO}_2$ ,  $\text{ReO}$ ) present. Re oxides are hygroscopic that was the reason the sample looked clear after absorbing moisture.

**Example of #2:** Several optical materials such as mirrors, radiation deflectors, and lenses used by NASA are made from layered materials of different physical properties. The thickness of these layers can be measured by sputtering using the Argon ion gun and followed by chemical analysis to verify chemistry of the individual layers as they are removed. It is assumed that the sputtering rate of various substrates are nearly same and that the gun at its maximum capacity (commercially available) removes about 1.0 angstroms thickness per seconds for  $\text{SiO}_2$ . However, sputtering rates are not same for all matrices under the same ion gun settings. Sputtering rates of some of the substrates (relative rates at 4KV gun voltage) are as follows (Ref. 7, page 27).

<u>Target</u>	<u>Sputter Rate</u>
$\text{Ta}_2\text{O}_5$	1.00
Si	0.90
$\text{SiO}_2$	0.85
Pt	2.20
Cr	1.4
Al	0.95
Au	4.10

Therefore, care must be taken and sputtering rates for each material needs to be calibrated for accurate thickness determination.

**Example of #3:** Carbon fibers are coated with materials that aids in the adhesion of fibers to the matrix. Chemistry and amount of this coating material can be evaluated using ESCA since the fibers are very fine. To perform this analysis, a bunch of fibers are pressed and clamped to create a flat surface such that the surface created is covered by fibers. The surface with its attachment for the bundle is introduced into the analytical chamber and the coating can then be analyzed.

**Example of #4:** Contaminations on surfaces are a big problem when two surfaces are not able to bond leading to exfoliation, corrosion, or bulging of paint on a painted surface. This calls for a surface analysis and a possible remedy of the failure. Contamination on one of the inside layers on a multi-layered material becomes challenging since removal of layers by sputtering is very slow process and may not be uniform. There are several examples of contaminants in engineering components. They generally are body oils, lubricants, remnant chemicals from a cleaning process often originating from negligence during or after

manufacturing. To minimize such contaminations, production facilities use clean rooms and the operators use gloves and face masks during handling of critical hardware.

It is important, therefore, that the samples reaching the laboratory must be pristine so that the traces of the contaminants remain on the surface and are not further contaminated. For exfoliating paints or plating, the paint or plating could be peeled and the contaminated surface could be analyzed. Silicone contamination is a big culprit of non-adherence and can be easily identified using XPS. For silicones the Si 2p binding energy peak lies between that of SiO<sub>2</sub> and the Si element.

**Example of #5:** XPS can be used as a general tool for failure analysis. Examples would be identifying materials on a fracture surface. This can be accomplished despite the fracture surface could be rough and uneven. The example could be identification of chloride salts on a stainless-steel fracture surface. At NASA Marshall Space Flight Center fracture surface of a rocket engine component made from Niobium was analyzed. XPS analysis of the surface indicated to have reacted to fluoride containing lubricants resulting in compounds such as NbO<sub>2</sub> F. It is believed that the lubricant reacted with the surface to produce this compound and possibly helped propagate the crack.

Failure analysis of polymeric materials exposed to x-ray and other radiations can be accomplished by taking advantage of loss of peak ratios from data bases (Ref. 14 page 45). Peak ratio losses as function of time are shown in Fig. 16. This indicates that the polymers are degraded by exposure to x-ray during analysis and their ratios would indicate the length of exposure that could be related to failures.

## **Examples from Manufacturer's websites:**

Currently, there are three main manufacturers of XPS machines. They are: Physical Electronics Incorporated (PHI), ([www.phi.com](http://www.phi.com)); Thermo Fisher Scientific Inc., ([www.thermofisher.com](http://www.thermofisher.com)); and Kratos Analytical Limited ([www.kratos.com](http://www.kratos.com)). All the three have application sections in their websites. Following pages are excerpts from the websites and the websites may be visited for details of what follows. It is always worthwhile to visit their websites as they have examples of advanced applications, continuously developing capabilities, and XPS applications for new materials.

### **Physical Electronics (Ref.15)**

*Lithium Battery Electrode Analysis:* XPS is routinely used to study the surface composition of Li battery electrodes as a function of use, which is critical to understanding the mechanisms that may limit the life of a battery. Due to the reactivity of Li battery electrode surfaces it is important to have the ability to transport samples from a controlled environment, such as a glove box, to the surface analysis instrument under vacuum or with an inert cover gas. Shown in Fig.17 are spectra from a lithium anode surface with and without air exposure. The observed surface chemistries show how important it is to protect a lithium anode surface from air exposure prior to analysis.

When Li surface is exposed to air, it forms carbonate but when the sample is not exposed, it will show oxides and hydroxides on anode surfaces. XPS surface analysis can show these differences, see Fig.17.

*Fuel Cell Membrane Characterization:* Fuel cell membranes are multi-layered materials with a permeation membrane and two thin layers of noble metal in a polymer matrix acting as anode and the cathode, respectively, of the fuel cell. Cross-sectional chemical analysis of these membranes is useful for studying mechanisms that impact life and performance of a fuel cell. Fig.18 shows XPS chemical state distribution

images from an aged fuel cell membrane that indicates degradation by fluorocarbon products in the anode and cathode regions.

Contamination Identification on Polymer Surfaces: XPS has the capability to chemically characterize contaminants on a polymer surface. In this case, a scanning x-ray beam induced secondary electron image of the optically transparent PET sample revealed the presence of unexpected localized surface contaminants on the polymers surface. Micro-area XPS spectra obtained with a 20  $\mu\text{m}$  diameter x-ray beam quickly identified the contaminant as a fluorocarbon (See Fig. 19).

### **Termo Fisher Scientific (Ref.16)**

XPS Evaluation of Wear Resistant Coatings: Additives in lubricating oils play an important role in forming friction modifying layers on metal surfaces. Calcium sulfonate additives in lubricating oil, as well as zinc dialkyldiphosphonate (ZDDP), are used to deposit wear resistant layers on steel surfaces. It has been observed that ZDDP forms a protective, glassy phosphonate coating on surfaces under tribological loads. In this study various ratios of calcium sulfonate in oil and ZDDP were used.

Three tribological samples were tested in this evaluation. They are: GOOD<sub>NEW</sub>, BAD<sub>NEW</sub>, and GOOD<sub>OLD</sub>. Although, full details of all three samples were unavailable, the GOOD<sub>OLD</sub> sample was in the tribology test rig for a long time showing good tribological properties; GOOD<sub>NEW</sub> sample behaved well under the tribology test rig whereas BAD<sub>NEW</sub> did show to have unusual tribological parameters in the test. All three samples were analyzed using monochromatic XPS. Some of the results are shown in Fig.20.

Two of the samples, labelled GOOD<sub>OLD</sub> and GOOD<sub>NEW</sub>, were shown to have good friction stability properties and XPS showed that the presence of calcium carbonate tracks on these samples. A third sample, which was known to have poorer friction stability properties, labelled BAD<sub>NEW</sub>, had virtually no calcium carbonate on its surface. XPS analysis showed that the correct ratio of ZDDP to calcium detergent will result in the formation of calcium carbonate during tribological load, and it is this carbonate that confers good friction stability properties. Calcium carbonate, however, did not form on the BAD<sub>NEW</sub> sample, indicating an inappropriate ratio of ZDDP to detergent in the oil formulation.

XPS Analysis of a Surface Contamination on a Steel Sample: Ensuring a surface is free from contaminants is extremely important in guaranteeing that materials such as steel meet their desired performance specification. Surface contamination of steel can result in problems such as adhesion failure and contact bonding problems when components are used in manufacturing. Surface contaminants can also result in an “unsatisfactory” appearance for many steel finished products such as ovens and other domestic appliances. In addition, surface contaminants are often the source of cross-contamination, corrosion and electrical contact problems. Many of these surface contamination issues are difficult to detect during or after production of manufactured goods using steel parts. XPS was used to analyze surface chemistry to identify the contamination.

In this investigation, an area (3.7 mm  $\times$  4.8 mm) of a stainless steel surface was investigated and several elemental maps were acquired by scanning the sample stage under the X-ray spot and collecting the spectra. The atomic concentration maps of the analyzed area were taken and integrated. The maps show clearly the difference between clean stainless steel and the contaminated areas. The contamination was identified to be an organic residue.

### **Kratos Analytical (Ref.17)**

Applications in Ionic Liquids Characterization: In the past ten years or so, ionic liquids have become an area of increasing popularity in academic research which has resulted in the stellar rise in the number of

publications in this field. This rise can in part be attributed to the wide ranging applications these novel materials can be adapted to. Only recently has this new field attracted the attention of a growing number of surface scientists interested in exploring the interactions at the both the liquid/gas and liquid/solid interfaces. Of particular interest is the structure and composition of the liquid/gas interface as this is where the adsorption and desorption of gasses occur. These activities are known to play key roles in processes such as heterogeneous catalysis and gas distillation and separation. Aside from such obvious applications, the surface analysis of ionic liquid helps in the fundamental understanding of these unique materials. Website provides a list of publications where Kratos AXIS spectrometers have been used to generate valuable XPS data from ionic liquid characterizations.

Coatings and Thin Films: Surface coatings and thin films are of great commercial importance in many industries and are used to enhance or provide required properties to bulk materials specific to their applications. Thin films can range from tens of Angstroms to several microns in thickness and find application in areas as diverse as optical anti-reflective coatings, architectural glazing and drug eluting thin films in the pharmaceutical industry. XPS is ideally suited for characterizing the surface chemistry of these thin films and when combined with sputter depth profiling can be used to determine the elemental and chemical composition as a function of depth through the film using XPS spectrometers.

Polymers: Polymer materials are finding ever increasing application in numerous consumer products. Applications of polymer materials range from fields as diverse as food packaging to organic electronics and biomaterials to automotive body panels. The surface properties of these materials are often vital in determining performance of the polymer for the specific application. X-ray photoelectron spectroscopy is ideally suited to the surface characterization of these polymer materials as it can provide quantitative chemical state information from the upper 10nm of the material.

The majority of polymer materials are insulating and therefore effective charge compensation is paramount if generating high resolution spectra with monochromatic x-rays. The performance of modern spectrometers is demonstrated by the guaranteed performance on PET (Polyethylene terephthalate) where the FWHM of the component corresponding to the ester group is <0.68 eV (see Fig. 14) with a sensitivity defined by the maximum of the hydrocarbon peak counts greater than 12 kcps.

An interesting aspect of polymer characterization is the use of cluster ion guns. Kratos website literature demonstrates the advantages of polymer characterization by high resolution XPS combined with  $Ar_n^+$  cluster depth profiles.




Fig. 1: Shows the electronic transitions involved in an XPS (ESCA) process. It shows Al monochromatic x-rays coming from to left to a copper sample. Ejected electron from the shells are knocked off and collected by the XPS spectrometer to produce a spectrum (top right). The peak correspondences have been illustrated.






Fig.2 – Magnesium x-ray satellite peaks seen for around the C1s peak for graphite. Binding energy displacements are  $\alpha_3$  at 8.4 eV,  $\alpha_4$  at 10.1,  $\alpha_5$  at 17.6,  $\alpha_6$  at 20.6, and  $\beta$  at 48.7 (Ref.7 page 18).




Fig. 3: Shake-up lines (2p) as seen for some copper compounds (Ref. 7 page 19).




Fig.4 – Surface and bulk plasmon lines associated with Al 2s at normal and grazing take-off angles (Ref.7 page 21).




Fig.5 – Relative Binding Energies in eV (x-axis) for a Uranium atom (Ref.7 page 10)




Fig. 6 – XPS emission process for a model atom. An incoming photon causes the ejection of a photoelectron (top). The relaxation process results in an emission process of an Auger electron (shown in bottom) KL<sub>23</sub>L<sub>23</sub>. Final arrangement results in a two-electron vacancy at the L<sub>2,3</sub> or 2p level (Ref. 7 page 11).




Fig. 7 – Photoelectron spectra from Uranium using Al monochromatic radiation (Ref. 7 page 194).




Fig.8 – Schematic presentation of a spherical-sector analyzer with monochromatized x-ray source (Ref. 8, page 131).




Fig.9 – XPS Machine (Kratos AXIS Ultra) with fracture stage, UPS, Auger attachment, and a reaction cell (at the back – not shown).





Fig.10 – Charge neutralization cage for insulators goes on SSX-100 machine. The nickel screen is not shown but goes on top. The spring around grabs it and keeps it in place; sample goes inside the cage. The whole fixture is mounted at the bottom and goes to the analyzer stage.




Fig.11- Splitting of Si 2p line by a high resolution spectrometer.

- Simple background functions

- Linear
- Quadratic

$$B(E) = E_{\text{off}} + a \cdot E + b \cdot E^2$$

- Shirley (step-like)

- Background is proportional to integral over peak up to the point where background is determined

$$S(E) = E_{\text{off}} + a \cdot \int_{E_0}^E I(E') dE'$$

- Strictly,  $S(E)$  is found by iterative approach
- Can be approximated by analytical function.





Fig.12 – Various background corrections applied to XPS peaks.

		Atomic number	9	1.0	PHI sensitivity factor for designated photoelectron transition																														
		Element symbol	F																																
1	Most intense photoelectron transition		1s	685	Binding energy, most intense photoelectron transition		2																												
	Most intense Auger transition		KLL	647	Kinetic energy, most intense Auger transition		He																												
3	0.025	4	0.074					5	0.159	6	0.296	7	0.477	8	0.711	9	1.0	10	1.340																
Li	Be						B	C	N	O	F	Ne																							
1s 66	1s 112					1s 187	1s 285	1s 402	1s 531	1s 685	1s 883																								
KLL 43	KLL 103					KLL 177	KLL 264	KLL 380	KLL 508	KLL 655	KLL 816																								
11	1.685	12	0.252					13	0.193	14	0.283	15	0.412	16	0.570	17	0.770	18	1.011																
Na	Mg						Al	Si	P	S	Cl	Ar																							
1s 1072	2p 50					2p 73	2p 99	2p 130	2p 164	2p 198	2p 242																								
KLL 994	KLL 1185					LMM 68	LMM 93	LMM 120	LMM 161	LMM 183	LMM 215																								
19	1.30	20	1.634	21	1.678	22	1.798	23	1.912	24	2.201	25	2.42	26	2.656	27	3.255	28	3.653	29	4.768	30	3.354	31	3.341	32	3.100	33	3.570	34	0.722	35	0.895	36	1.086
K	Ca	Sc	Ti	V	Cr	Mn	Fe	Co	Ni	Cu	Zn	Ga	Ge	As	Se	Br	Kr																		
2p 284	2p 347	2p 389	2p 454	2p 512	2p 574	2p 638	2p 707	2p 778	2p 853	2p 933	2p <sub>1/2</sub> 1022	2p <sub>1/2</sub> 1117	2p <sub>1/2</sub> 1217	3d 42	3d 56	3d 69	3d 87																		
LMM 248	LMM 290	LMM 338	LMM 419	LMM 479	LMM 528	LMM 587	LMM 703	LMM 774	LMM 846	LMM 919	LMM 992	LMM1068	LMM1145	LMM1225	LMM1306	MNV 101																			
37	1.316	38	1.578	39	1.867	40	2.216	41	2.517	42	2.867	43	3.266	44	3.695	45	4.179	46	4.643	47	5.198	48	3.444	49	3.777	50	4.095	51	4.473	52	4.925	53	5.337	54	5.702
Rb	Sr	Y	Zr	Nb	Mo	Tc	Ru	Rh	Pd	Ag	Cd	In	Sn	Sb	Te	I	Xe																		
3d 111	3d 134	3d 156	3d 179	3d 202	3d 228	3d 253	3d 280	3d 307	3d 335	3d 368	3d <sub>5/2</sub> 405	3d <sub>5/2</sub> 444	3d <sub>5/2</sub> 485	3d <sub>5/2</sub> 528	3d <sub>5/2</sub> 573	3d <sub>5/2</sub> 619	3d <sub>5/2</sub> 670																		
MNN 102	MNV 131	MNV 156	MNV 188	MNV 188	MNN 248	MNN 275	MNN 302	MNN 328	MNN 358	MNN 384	MNN 411	MNN 438	MNN 465	MNN 492	MNN 519	MNN 545																			
55	6.032	56	6.361	57	7.706	72	2.221	73	2.588	74	2.858	75	3.327	76	3.747	77	4.217	78	4.674	79	5.240	80	5.787	81	6.447	82	6.968	83	7.602	84		85		86	
Cs	Ba	La	Hf	Ta	W	Re	Os	Ir	Pt	Au	Hg	Tl	Pb	Bi	Po	At	Rn																		
3p <sub>5/2</sub> 726	3d <sub>5/2</sub> 781	3d 836	4f 14	4f 22	4f 31	4f 40	4f 51	4f 61	4f 71	4f 84	4f 101	4f 118	4f 137	4f 157																					
MNN 668	MNN 691	MNN 633	NNN 181	NNN 181	NNN 180	NNN 178	NNN 176	NNN 163	NNN 170	NNN 163	NCO 81	NCO 88	NCO 86	NCO 104																					
87	88	89																																	
Fr	Ra	Ac																																	
58	7.399	59	6.356	60	4.897	61	3.754	62	2.967	63	2.210	64	2.207	65	2.201	66	2.198	67	2.189	68	2.184	69	2.172	70	2.169	71	2.156								
Ce	Pr	Nd	Pm	Sm	Eu	Gd	Tb	Dy	Ho	Er	Tm	Yb	Lu																						
3d 894	3d 932	3d 981	3d 1034	3d <sub>5/2</sub> 1081	4d 128	4d 140	4d 149	4d 152	4d 160	4d 167	4d 175	4d 182	4f 7																						
MNN 654	MNN 690	MNN 728	MNN 773	MNN 806	MNN 850	MNN 885	MNN 1076	MVV 1119	MVV 1173	MVV 1214																									
90	7.498	91		92	6.476	93		94		95		96		97		98		99		100		101		102		103									
Th	Pa	U	Np	Pu	Am	Cm	Bk	Cf	Es	Fm	Md	No	Lr																						
4f <sub>7/2</sub> 333		4f <sub>7/2</sub> 337																																	
NOV 88		NOV 75																																	

Fig. 13 – Sensitivity Factors along with other physical properties of elements.



Peak	Position BE (eV)	FWHM (eV)	Raw Area (CPS)	Atomic Conc %
C 1s CC,CH	285.077	0.896	29875.7	58.96
C 1s C-O	286.626	0.896	10062.6	19.86
C 1s ester	289.053	0.690	8364.6	16.50
C 1s sat	291.988	1.712	2369.7	4.68






Fig. 14 – Compares PET scans (C 1s area) in two XPS machines, top from SSX-100 and, bottom, Kratos Axis Ultra. Notice the peak widths.



**Nylon 6,6 (N66)**

	C 1s				O 1s	N 1s
	1	2	3	4		
BE (eV)	285.00	285.46	286.02	288.02	531.37	399.81
FWHM (eV)	0.95	0.99	1.03	0.95	1.18	1.17
Area (%)	51	17	17	15		
A	0.07	0.06	0.16	0.11	0.13	0.10
m	0.86	0.90	0.92	0.83	0.76	0.83



Source: Aldrich  
 Casting solution: 1% (w/w) in HFIP  
 Degradation index: 5, O/C  
 VB acquisition: 2  
 Charging reference: lowest BE component

Fig.15 – Spectrum from Nylon 6,6. Curve fitting of C 1s line is shown on top left. Bottom part shows the formula and the position of various carbon peaks fitted to C 1s at top left. Valence band spectra is seen at middle on right.




Figure 28. Plots of  $(X_t/X_0) \times 100$  vs  $t$  for some of the chlorine containing polymers in the database.  $X = \text{Cl}/\text{C}$  atom ratio.





Figure 27. Plots of  $(X_t/X_0) \times 100$  vs  $t$  for the aliphatic ether polymers in the database.  $X = \text{O}/\text{C}$  atom ratio.

Fig. 16 – Compares various polymers as they get damaged due to exposures to x-rays. Top is the loss of Cl intensity whereas the bottom is the loss of O peaks. (Ref. 14, page 45 Figs. 28 and 27 respectively).

### With Air Exposure

BE	Bond	%Area
52.88	Li	0
54.06	Li <sub>2</sub> O	3
55.26	Li <sub>2</sub> CO <sub>3</sub>	97



### Without Air Exposure

BE	Bond	%Area
52.88	Li	55
54.06	Li <sub>2</sub> O	25
55.28	LiOH	19




Fig. 17 - Lithium Battery Electrode Analysis using XPS shows details revealed, with and without any air exposure. Ref.15







Fig. 18 – Left shows F 1s spectrum region corresponding to the chemical state images on the right. Chemical state image showing the presence of fluorocarbon chemistry at the center (blue) and the degradation products, fluoride chemistry, in the anode and the cathode regions (green). Ref.15






Fig.19 – Contaminant analysis on a PET sample identified as a fluorocarbon. Top left is a Scanning Electron Microscope image, Top right is XPS scans from contaminant (red) and the PET (blue), and bottom left is detailed C 1s region. Ref.15.




Fig.20 - It shows spectrum obtained around C 1s on all three samples. It is seen that GOOD<sub>OLD</sub> shows significant carbonate compared to the other two; BAD<sub>NEW</sub> shows the least amount of carbonate. Ref. 16.

## References:

- (1) Albert Einstein, Ann. Physik, vol.17, 1905, p 132.
- (2) H. Robinson, Phil. Mag. Vol. 50, 1925, p 241.
- (3) K. Siegbahn et.al., ESCA: Atomic, Molecular and Solid State Structure Studied by Means of Electron Spectroscopy, Almqvist and Wiksells, Stockholm, 1967.
- (4) K. Siegbahn, D. Hammond, H. Fellner-Feldegg and E.F. Barnett, Science, vol.176, 1972, p 245.
- (5) K. Larsson, C. Nordling, K. Siegbahn, and E. Stenhagen, Acta. Chem. Scand. vol. 20, 1966, p.2880.
- (6) B. Vincent Crist, Ph. D., Wikipedia.
- (7) Handbook of X-Ray Photoelectron Spectroscopy, Physical Electronics, 1995 – various pages.
- (8) Methods and Phenomena 1, A. W. Czandrer Ed., Elsevier Scientific Publishing Co., 1975, p.131.
- (9) Handbook of Monochromatic XPS Spectra, B. Vincent Crist, John Wiley & Sons. Ltd. 2000, p xv.
- (10) Casa XPS data bases – Casa.XPS.com.
- (11) D.A. Shirley, Phy. Review, B5, 1972, p. 4709.
- (12) Background subtraction: 1. General behavior of Tougaard-style background in AES and XPS by M. P. Seah, vol.420, Surface Science Issues 2-3, January 1999, p. 285.
- (13) J. H. Scofield, Journal of Electron Spectroscopy, vol.8, 1976, p.129.
- (14) High Resolution XPS of Organic polymers, G. Beamson and D. Briggs, The Scienta ESCA300 Database, John Wiley & Sons, 1992, p.174.
- (15) Physical Electronics Inc. website: [www.phl.com](http://www.phl.com).
- (16) Thermo Fisher Scientific Inc. website: [www.thermofisher.com](http://www.thermofisher.com).
- (17) Kratos Analytical Limited. Website: [www.kratos.com](http://www.kratos.com).