


# NASA TECH BRIEF



NASA Tech Briefs are issued to summarize specific innovations derived from the U.S. space program, to encourage their commercial application. Copies are available to the public at 15 cents each from the Clearinghouse for Federal Scientific and Technical Information, Springfield, Virginia 22151.

## Instrument Sequentially Samples AC Signals from Several Accelerometers



### The problem:

In conducting random noise vibration tests using several accelerometers to obtain a representative measurement of the vibration levels over the entire test structure, it is necessary to provide a means for sampling each of the accelerometer signals in such a manner that all of the frequency information is preserved.

### The solution:

A scanner circuit that sequentially samples the ac signals from the accelerometers and provides an essentially time-averaged output signal.

### How it's done:

A main unit of the scanner is a modified ring counter employing silicon controlled rectifiers to actuate magnetically shielded, mercury wetted reed relays. A unijunction transistor is used as the active element of a sawtooth generator that provides a trigger pulse to switch the counter. The pulse duration

and repetition rate are adjustable to permit a sampling rate from less than 1 Hz to approximately 200 Hz. The number of accelerometer channels to be sequentially sampled can be selected by means of a selector switch on the front panel of the instrument. The direction of sampling (forward or reverse) can be controlled by another switch on the front panel. The sampling can also be stopped on any desired channel, so that only the signal from that channel will be gated through the instrument.

### Notes:

1. The output of the scanner has been applied to an automatic equalizer to enable precise control of the vibration levels of a vibration system used for random noise vibration tests on structural specimens. The scanner may also have application in conjunction with devices used for automatic control of vibration levels on motors or other equipment subject to mechanical vibrations.

(continued overleaf)

2. An instrument that will automatically select the peak ac signal from several accelerometers is described in NASA Tech Brief 66-10462. Inquiries may also be directed to:

Technology Utilization Officer  
Jet Propulsion Laboratory  
4800 Oak Grove Drive  
Pasadena, California 91103  
Reference: B67-10029

**Patent status:**

Inquiries about obtaining rights for the commercial use of this invention may be made to NASA, Code GP, Washington, D.C. 20546.

Source: Carl P. Chapman  
(JPL-884)