

N89 - 20061

**DEVELOPMENT OF A SOFTWARE INTERFACE FOR
OPTICAL DISK ARCHIVAL STORAGE FOR A NEW
LIFE SCIENCES FLIGHT EXPERIMENTS COMPUTER**

FINAL REPORT

NASA/ASEE Summer Faculty Fellowship Program -- 1988

Johnson Space Center

Prepared by:	Peter N. Bartram, Ph.D., P.E.
Academic Rank:	Associate Professor
University and Department:	Norwich University Computer Science Engineering Division of Engineering Northfield, Vermont 05663
 NASA/JSC	
Directorate:	Space and Life Sciences
Division:	Life Sciences Project Division
Branch:	Project Engineering Branch
JSC Colleague:	Donald Stilwell
Date Submitted:	3 August 1988
Contract Number:	NGT 44-005-803

ABSTRACT

The current Life Sciences Laboratory Equipment (LSLE) microcomputer for life sciences experiment data acquisition is now obsolete. Among the weaknesses of the current microcomputer are small memory size, relatively slow analog data sampling rates, and the lack of a bulk data storage device. While life science investigators normally prefer data to be transmitted to Earth as it is taken, this is not always possible. No down-link exists for experiments performed in the Shuttle middeck region. One important aspect of a replacement microcomputer is provision for in-flight storage of experimental data.

The Write Once, Read Many (WORM) optical disk was studied because of its high storage density, data integrity, and the availability of a space-qualified unit. In keeping with the goals for a replacement microcomputer based upon commercially available components and standard interfaces, the system studied includes a Small Computer System Interface (SCSI) for interfacing the WORM drive. The system itself is designed around the STD bus, using readily available boards. Configurations examined were (1) master processor board and slave processor board with the SCSI interface, (2) master processor with SCSI interface, (3) master processor with SCSI and Direct Memory Access (DMA), (4) master processor controlling a separate STD bus SCSI board, and (5) master processor controlling a separate STD bus SCSI board with DMA.

Storage times for 512-byte disk sectors ranged from 56 to 22.5 milliseconds without DMA and 53 to 8.9 milliseconds with DMA. The faster times resulted from using large blocks of data per disk write request. With the faster times, the storage rates significantly exceed anticipated data acquisition rates. While DMA is especially attractive, configuring workable systems from available boards without hardware modification is difficult.

INTRODUCTION

The success of manned space exploration depends in part on an understanding of the physiological effects of microgravity. A major concern of life sciences research in space is to guarantee the health and safety of astronauts for missions of both long and short duration. Additionally, astronauts have experienced many symptoms during exposure to microgravity. These latter symptoms have been subsumed under the rubric of "Space Adaptation Syndrome." It is part of the charter of the Life Sciences Flight Experiments Program (LSFEP) to perform experiments in space that will contribute to the understanding of the physiological effects of space travel on humans and, where possible, to develop strategies for reducing the debilitating effects.

Most flight investigations require the use of a microcomputer for high speed collection, storage, and downlinking of in-flight physiological data. The Life Sciences Laboratory Equipment (LSLE) Microcomputer has served these needs for life sciences investigations in the shuttle Spacelab module. This computer has provided the link between Spacelab experiments and the Science Monitoring Area (SMA) in the Life Sciences Project Division building at the Johnson Space Center (JSC). To perform this function the LSLE microcomputer sends formatted real-time data, in a digital serial form, to a High Rate Multiplexer (HRM) in the Spacelab module, which in turn transmits the data to the ground via a Tracking and Data Relay Satellite System (TDRSS) satellite.

While the old LSLE microcomputer has served its purposes well, it has become obsolescent and must be replaced by a newer design. Shortcomings of the old LSLE microcomputer include a lack of replacement units and parts, an absence of on-board mass storage or real-time data display, an inability to be programmed in a high level language, and poor general specifications when compared to today's standards. Because of these deficiencies, the development of a new LSLE microcomputer is now a high priority item. The requirements for a replacement LSLE microcomputer include all of the data acquisition, HRM downlink, and experiment control functions of the current machine. Additionally, the new microcomputer must have better and more extensive displays, faster data sampling rates, extensive archival storage capabilities for use in middeck experiments where no HRM downlink is available, greater processing speed, more memory, compatibility with popular commercially available systems such as the IBM PC, and the ability to be programmed by the principal investigator using common high-level languages. The design should be based to the maximum extent possible on commercially available boards and should avoid custom hardware as much as

possible. It is hoped that the HRM interface is the only custom design required.

A multiple processor system has several advantages over single processor systems. In the multiple processor system, one processor (called the "master") is reserved to the maximum extent possible for the principle investigator's use, while additional processors (called "slaves") are dedicated to each input/output operation. For example, separate slave processors are used for analog to digital conversion (A/D), digital to analog conversion (D/A), Parallel I/O, Serial I/O, the HRM interface, optical disk control, etc. In this way, work is shared among the many processors. Once this partitioning is defined, separate engineering efforts can attack each task and development work can proceed in parallel. Expansion would have minimal impact on the original parts of the system, as additional processors would handle the additional data acquisition load. This modularization facilitates software and hardware maintenance. It allows on-going replacement of obsolete components at the board level, without significantly disturbing other parts of the system. A chief advantage of this concept is that the modular nature of the system extends into the hardware itself, thereby making development and reconfiguration much simpler.

One of the most important elements of the new LSLE microcomputer is the optical disk. The Write Once Read Many (WORM) optical drive would allow the LSLE microcomputer to be used in the shuttle middeck where no downlink capabilities are available. This will allow the LSLE microcomputer to be used on a much larger number of flights (including those not flying the Spacelab), thereby greatly expanding opportunities for the Life Sciences Flight Experiments Program. In most cases, the investigator will prefer real-time data downlink, but as flight opportunities are limited, he may choose to fly his experiment in the middeck rather than to risk losing the opportunity altogether. In other cases, the investigator may be given a late opportunity to add his experiment to an existing fully developed payload. In this case, time may not permit the development of ground software to handle the HRM data stream. Again, the use of an optical disk will reduce data handling costs and development times. It has been estimated that the cost of ground software to handle the HRM data stream for life sciences experiments is \$250,000 per mission, regardless of the number of experiments served. The use of the optical disk instead of the HRM would remove this cost entirely. In summary, the use of an optical disk will allow single experiments to be inserted into missions on relatively short notice and will permit full use of the middeck by life science investigators. Other bulk storage systems, such as streaming tape and other magnetic devices, do not offer the robustness and high data density of the optical disk.

The objective of my work this summer was to develop prototype software needed to store collected data on an optical disk as well as to recover it. Also, timing studies were to be performed to determine if data storage rates would

be sufficient for the data collection sample frequency required of the new system.

EQUIPMENT USED

The commonly used STD bus was chosen for the system backplane. This bus is well supported, with several hundred manufacturers offering several thousand boards. Thus, prospects are excellent for future upgrades in keeping with advances in technology.

The STD bus processor boards selected for the system are based on the Intel 8088 and 80188 processors. Advantages of these boards include processor compatibility with the familiar IBM PC family and the availability of a wide range of software and hardware support products. In the system under study, the master processor board is the Ziatech Corporation ZT8815, containing an 8 MHz 80188. This board has computing power similar to the IBM PC/AT. Of particular interest, the 80188 processor chip contains Direct Memory Access (DMA), which provides fast memory -- peripheral data transfer in parallel with other processor functions. Additional processor boards, slave processors, are Ziatech ZT8830 boards. These contain 8 MHz 8088 processors, with computing power approximating that of the IBM PC. Both master and slave boards contain a standard SBX interface and connector for "piggyback" peripheral modules. Hundreds of modules are available from dozens of vendors, providing real-time input and output hardware. Thus self-contained data acquisition subsystems can be formed using, for example, an SBX analog to digital converter attached to a slave processor board.

The optical disk interface is the Small Computer System Interface (SCSI) standard. One configuration uses a slave processor with the Zendex Corporation model ZBX-280 SCSI SBX module. The ZBX-280 is based on the NCR 5380 SCSI controller chip. The same SBX module attached to the master processor is an alternative configuration, which allows the use of DMA, not available with the slave processor boards. Another alternative SCSI interface is the Ziatech STD bus peripheral board, the ZT8850, which provides a chip set for SCSI as well as other disk control devices. The ZT8850 includes its own DMA. Compared to the Zendex ZBX-280, the ZT8850 is simpler to program, with more SCSI signals generated by hardware rather than under program control. The master processor board controls the ZT8850 as a bus peripheral, unlike the ZBX-280 which is directly interfaced and does not reside on the STD bus. With the ZT8815 processor and the ZT8850 SCSI boards, DMA may be implemented using either the 80188 DMA of the ZT8815, or the ZT8850 on-board DMA circuits.

The optical disk drive is the Optotech model 5984. This drive was selected because Mountain Optech (Boulder, CO) makes a space qualified version of the drive (model 200SES) under contract to Goddard Space Flight Center. The model 5984 is identical in electrical and software characteristics to the model 200SES, but is less expensive. Both are 200 megabyte (per side) capacity optical disk drives, using 5 1/4 inch removable media cartridges.

A summary of the equipment used and location of principal vendors is given in Table 1.

SOFTWARE DEVELOPED

Test programs were written to assure that the optical disk could be both written and read in the various configurations considered, and to determine read and write times. Best case conditions were used in that no other competing activity was required of the processor chip or, where used, the STD bus. In all cases, a main program in C calls a collection of assembly language functions. For support of the Zendex SCSI piggyback (NCR 5380 chip) interface with either master or slave processor board, the assembly language routines are revisions of routines provided by Mountain Optech. Each principal SCSI phase is handled by a separate call: reset, selection (including arbitration), status, command, message in, data output, and data input. (The message out phase has not been implemented.) The data transfer is byte by byte, with polling and handshake signal generation by software. The calling program can thus check for proper progression of phases and handle error conditions. Additional functions were written for DMA use for the data input and data output phases with the Zendex SCSI interface used with the ZT8815 master processor board.

A separate set of assembly language functions, matching the names and organization of the above routines, has been prepared for use with the master ZT8815 board and ZT8850 peripheral SCSI board. These require no significant change in the C calling programs written for other configurations. While the example code available from Ziotech for the ZT8850 support was influential, these routines followed the organization of the functions for the Zendex interface, rather than the task oriented Ziotech routines. The routines for the ZT8850 are generally simpler than those for the NCR 5380 based Zendex interface. The ZT8850 does not support arbitration. The handshake signals required for data transfer are hardware

TABLE 1. -- EQUIPMENT USED.

HARDWARE

Ziatech 8862 Card Cage and Power Supply (STD bus)
Ziatech ZT8830 Intelligent I/O Control Processor (slave)
Ziatech ZT8815 80188 based CPU card (master)
Zendex ZBX-H280 SCSI controller multimodule (iSBX)
IBM PC compatible personal computer for downloading programs to STD bus boards
Digital storage oscilloscope for obtaining timing data
Mountain Optech Model 5984 Optical disk drive

SOFTWARE

Ziatech 8830 Debug software (and ROMs) to load and debug programs
Ziatech 8815 Debug software
Microsoft C Language, version 4.00
Microsoft Macro Assembler, version 4.00
Microsoft Linker (loader), version 3.51
Microsoft MS-DOS operating system, version 3.10 (for program development)

VENDORS

Ziatech Corporation
3433 Roberto Court
San Luis Obispo, CA
93401

(805) 541-0488

Zendex Corporation
6700 Sierra Lane
Dublin, CA
94568

(415) 828-3000

Mountain Optech
2830 Wilderness Place
Suite F
Boulder, CO
80301

(303) 444-2851

Microsoft Corporation
10700 Northrup Way
Bellevue, WA
98004

(206) 882-8089

generated. The data transfer is still byte by byte, with handshake signal polling to avoid data loss. Some DMA support routines for the ZT8850 have been developed.

Copies of software developed for this project may be obtained from Peter N. Bartram, Division of Engineering, Norwich University, Northfield, Vermont 05663 (telephone 802/485-2263), or Donald Stilwell, NASA/Johnson Space Center, Mail Code: SE3, Houston, Texas 77058 (telephone 713/483-7308).

RESULTS

Polled data transfer can be performed in all configurations studied (ZT8830 + ZBX280, ZT8815 + ZBX280, and ZT8815 + ZT8850). DMA writing to the optical disk may be performed using the ZT8815 with ZBX280 configuration, provided the SBX signal TDMA, not implemented by the ZT8815, is grounded (requiring the addition of a wire on either the ZT8815 or the ZBX280). The ZBX280 uses this signal for one alternative for ending DMA transfer. With it left floating, as on the ZT8815, the ZBX280 attempts to halt DMA prematurely. Even with this change, with which DMA writing to disk works well, DMA reading of the disk returns incorrect values. The source of difficulty has not been determined with confidence. The ZT8850 DMA controller functions correctly, provided the memory used is not on the ZT8815 processor card controlling it. (A separate memory board was used.) Also at the time of this writing, code for using the ZT8815 processor DMA with the ZT8850 is under preparation.

For polled data transfers, timing results were similar for both reading and writing. For the LSLE replacement microcomputer, disk writing in real-time is critical, whereas reading will be performed at a later time, when speed is of lesser importance. With the ZBX280 SCSI interface, writing a single 512 byte sector each call for data output, the average write time was less than 56 milliseconds per block, regardless of processor. With each write request specifying a ten-sector block (5120 bytes), the average write time for ten sectors was 250.4 milliseconds (25.04 milliseconds per sector). Some ten sector blocks required 225 milliseconds, others 276 milliseconds. These times were identical for both the ZT8830 and ZT8815 processor boards. For write request block sizes of 25 sectors (12800 bytes), the average time required was 562.5 milliseconds per block (22.5 milliseconds per 512 byte sector). This is over 22 kilobytes per second. Since each real-time measurement results in a two-byte value, data storage in excess of 11000 samples per second is possible.

For the ZT8850 SCSI with no DMA, write times for single sector requests averaged 52.7 milliseconds per sector. For ten sector blocks, the average time per block was 197 milliseconds (19.7 per 512 byte sector). With 25 sector blocks, the time averaged 369 milliseconds (14.8 milliseconds per 512 byte sector). With 25 sector (12800 byte) blocks, over 33.8 kilobytes per second can be stored, or over 16,900 measurement samples per second.

With DMA using the ZT8815 with ZBX280, single sector write times averaged 52.7 milliseconds per sector. For DMA using blocks of 10 sectors per write request, the write time per block fluctuated between 124 and 176 milliseconds, averaging less than 150 milliseconds per block (15 milliseconds per 512 byte sector). The time required to write blocks of 25 sectors fluctuated between 210 and 260 milliseconds, averaging 221 milliseconds per 12800 byte block, or 8.84 milliseconds per 512 byte sector. Thus with DMA, storage rates in excess of 56 kilobytes per second (28000 measurement samples per second) are possible.

Using the ZT8850 STD bus SCSI board with DMA (as provided on the ZT8850 board), single sector write times also averaged 52.7 milliseconds per sector. Average write times using blocks of ten sectors per write request also was less than 150 milliseconds per block, 15 milliseconds per sector, and the same fluctuations in time were observed. However, the times for writing blocks of 25 sectors were longer: fluctuating between 260 and 312 milliseconds, averaging about 269 milliseconds per block. With this, about 46.4 kilobytes per second average storage rate results.

CONCLUSIONS AND RECOMMENDATIONS

Without DMA, data storage rates are marginally adequate for anticipated data collection rates. With DMA, storage rates exceed any envisioned requirements. DMA is particularly attractive because during data transfer to the disk, the processor is freed for other tasks.

At this point, it appears better to use the master processor for control of the SCSI interface to the optical disk. At the time of this writing, no slave processor supporting DMA is available with an SBX piggyback connector (for an SCSI interface), though these are under development. In order to gain the benefits of DMA, the master must be used to control the SCSI interface. Even if the slower transfer rates of currently available slave processor boards supporting the SBX piggyback modules (such as the ZBX-280 SCSI) are accepted, significant master processor usage would still be involved. This is because one slave cannot access memory of others. Thus

the master would have to transfer data from data acquisition slaves to the SCSI slave memory. With the master controlling the SCSI interface, no memory to memory copy is required. The master can send data to the SCSI directly from data acquisition slave board memory.

Block size is the most important factor for write speed. This is probably because with large numbers of contiguous sectors written with one write command, fewer disk seeks are required to properly position the write heads. DMA only marginally improved single sector block write times, where seek time seems to be limiting. With larger block sizes, not only is speed improved with both methods, but DMA gives markedly improved performance. With large amounts of data being written with each output command, data transfer time becomes limiting, as the number of seeks is reduced. The uneven write times can probably be attributed to the occasional write head repositioning required for two contiguously addressed sectors located on adjacent disk tracks.

However, DMA is not without difficulty, especially in mixed-vendor configurations. The TDMA signal incompatibility has been mentioned earlier. It is suspected that other problems of non-standard standards may be involved in using DMA in other configurations. The system designer must have a good understanding of both hardware and systems programming.

Several recommendations now can be made. (1) Further work is needed to test alternative DMA configurations. (2) As the block sizes were here chosen arbitrarily, optimization of the number of sectors written per write request in light of memory requirements is required. Along with this, the need to interleave data from several data acquisition slaves needs to be considered in determining block size. (3) It is important that the timing studies be repeated in a prototype system with all parts functioning. With other activity on the STD bus and more demands on the master processor, it should be verified that data storage rates will be maintained at a level exceeding the requirements of the data acquisition slave processors. (4) When slave processor boards supporting DMA and the SBX piggyback modules become available, the wisdom of placing the SCSI under control of the master processor should be reconsidered.