

greater values in the sequence in which the liquids were required to be used. Then sequential filling of the staining chamber with the various liquids would be achieved by momentarily increasing the speed of the centrifuge to exceed the corresponding threshold accelerations at the required times. Once the sample had been exposed to each liquid for the required time, the valve would be opened to drain the liquid from the staining chamber into the waste reservoir.

The reservoirs, channels, and staining chamber could be created, either on a glass slide or on a uniformly thick coating material on the slide, by one or more of a variety of techniques that could involve

photolithography, laser writing, and/or etching. Once a slide had been thus prepared, a sample would be placed on the staining-chamber area of a thin glass cover, then cover would be adhesively or electrostatically bonded to the slide or to the coating material. The liquids would then be dispensed into their assigned reservoirs via access ports and the air displaced from the reservoirs would leave through vent ports. The slide would then be placed on a centrifuge in the form of a spinning disk, and the centrifuge would be operated as described above to expose the sample sequentially to the various liquids.

*This work was done by Daniel L. Feedback of Johnson Space Center and Mark S. F.*

*Clarke of University Space Research Association. For further information, contact the Johnson Commercial Technology Office at (281) 483-3809.*

*In accordance with Public Law 96-517, the contractor has elected to retain title to this invention. Inquiries concerning rights for its commercial use should be addressed to*

*University Space Research Association  
10227 Wincopin Circle, Suite 212  
Columbia, MD 21044-3459*

*Telephone No.: (410) 730-2656*

*Fax No.: (410) 730-3496*

*Internet: info@usra.edu*

*Refer to MSC-23179, volume and number of this NASA Tech Briefs issue, and the page number.*

## Two-Armed, Mobile, Sensate Research Robot

**This could be a prototype of robotic home health-care aides.**

*Lyndon B. Johnson Space Center, Houston, Texas*

The Anthropomorphic Robotic Test-bed (ART) is an experimental prototype of a partly anthropomorphic, humanoid-size, mobile robot. The basic ART design concept provides for a combination of two-armed coordination, tactility, stereoscopic vision, mobility with navigation and avoidance of obstacles, and natural-language communication, so that the ART could emulate humans in many activities. The ART could be developed into a variety of highly capable robotic assistants for general or specific applications. There is especially great potential for the development of ART-based robots as substitutes for live-in health-care aides for home-bound persons who are aged, infirm, or physically handicapped; these robots could greatly reduce the cost of home health care and extend the term of independent living.

The ART is a fully autonomous and untethered system. It includes a mobile base on which is mounted an extensible torso topped by a head, shoulders, and two arms. All subsystems of the ART are powered by a rechargeable, removable battery pack. The mobile base is a differentially-driven, nonholonomic vehicle capable of a speed >1 m/s and can handle a payload >100 kg. The base can be controlled manually, in forward/backward and/or simultaneous rotational motion, by use of a joystick. Alternatively, the motion of the base can be controlled autonomously by an onboard navigational computer.

By retraction or extension of the torso, the head height of the ART can be adjusted from 5 ft (1.5 m) to 6 1/2 ft (2 m), so that the arms can reach either the floor or high shelves, or some ceilings. The arms are symmetrical. Each arm (including the wrist) has a total of six rotary axes like those of the human shoulder, elbow, and wrist joints. The arms are actuated by electric motors in combination with brakes and gas-spring assists on the shoulder and elbow joints. The arms are operated under closed-loop digital control. A receptacle for an end effector is mounted on the tip of the wrist and contains a force-and-torque sensor that provides feedback for force (compliance) control of the arm. The end effector could be a tool or a robot hand, depending on the application.

The ART includes several built-in, ready-to-use sensory subsystems and has room for addition of other sensors. One of the built-in sensory subsystems is a bumper/shield subsystem that includes mechanical tape switches and photodetectors to detect actual or incipient collisions with objects on the floor, plus ultrasonic sensors to detect nearby overhanging objects or walls. Other built-in sensory subsystems include laser-based distance measuring equipment, dual cameras for vision (including stereoscopic vision), the aforementioned force-and-torque sensors, and shaft-angle encoders and switches that provide position feedback for control of

the arms. The two cameras can be panned and tilted independently of each other; they can be aimed in different directions or the same direction.

Each of the various sensory and motion-control subsystems operates in conjunction with a computer subsystem that processes the incoming sensory data and controls the affected component of the motion. The navigational computer communicates with the sensory and base-motion-control computers. Another computer processes the digitized outputs of the cameras and controls the aiming, focus, and zoom of the cameras. Still another computer processes the digitized outputs of the force-and-torque sensors and executes software for control of the motions of, and forces exerted by, the arms.

Overall control is exerted by a host computer, which is of a Pentium-based laptop class. This computer runs speech-recognition and speech-synthesis software for communication with a human user. For purposes of experimentation and development, the host computer is also capable of radio communication with an external computer or network of computers.

*This work was done by J. F. Engelberger, W. Nelson Roberts, David J. Ryan, and Andrew Silverthorne of HelpMate Robotics, Inc., for Johnson Space Center. For further information, contact:*

*HelpMate Robotics, Inc.*

*22 Shelter Rock Lane*

*Danbury, CT 06810*

*Refer to MSC-22985.*